

UC Berkeley

UC Berkeley Previously Published Works

Title

Noise limits for dc SQUID readout of high-Q resonators below 300 MHz

Permalink

<https://escholarship.org/uc/item/0569n02q>

Journal

Journal of Applied Physics, 138(9)

ISSN

0021-8979

Authors

Ankel, V

Bartram, C

Begin, J

et al.

Publication Date

2025-09-07

DOI

10.1063/5.0280831

Copyright Information

This work is made available under the terms of a Creative Commons Attribution License, available at <https://creativecommons.org/licenses/by/4.0/>

Peer reviewed

Noise limits for dc SQUID readout of high-Q resonators below 300 MHz

V. Ankel ; C. Bartram ; J. Begin ; C. Bell ; L. Brouwer; S. Chaudhuri ; John Clarke ; H.-M. Cho ; J. Corbin ; W. Craddock; S. Cuadra ; A. Droster ; M. Durkin ; J. Echevers ; J. T. Fry ; G. Hilton ; K. D. Irwin ; A. Keller; R. Kolevatos ; A. Kunder ; D. Li; N. Otto; K. M. W. Pappas ; N. M. Rapisdis ; C. P. Salemi  ; D. Schmidt ; M. Simanovskaia ; J. Singh ; P. Stark ; C. D. Tesche ; J. Ullom; L. Vale ; E. C. van Assendelft ; K. van Bibber ; M. Vissers ; K. Wells ; J. Wiedemann ; L. Winslow ; D. Wright ; A. K. Yi ; B. A. Young 



J. Appl. Phys. 138, 094505 (2025)

<https://doi.org/10.1063/5.0280831>

 CHORUS



Articles You May Be Interested In

Low noise niobium dc SQUID with a planar input coil

Appl. Phys. Lett. (February 1983)

Thin-film $\text{YBa}_2\text{Cu}_3\text{O}_{7-x}$ direct current SQUIDs with Josephson junctions made by direct electron beam writing

Appl. Phys. Lett. (October 1996)

Sensitivity enhancement of Quantum Design dc superconducting quantum interference devices in two-stage configuration

Rev. Sci. Instrum. (September 2001)



Nanotechnology & Materials Science



Optics & Photonics



Impedance Analysis



Scanning Probe Microscopy



Sensors



Failure Analysis & Semiconductors



Unlock the Full Spectrum.
From DC to 8.5 GHz.

Your Application. Measured.

[Find out more](#)

 Zurich Instruments

Noise limits for dc SQUID readout of high-Q resonators below 300 MHz

Cite as: J. Appl. Phys. 138, 094505 (2025); doi: 10.1063/5.0280831

Submitted: 14 May 2025 · Accepted: 26 July 2025 ·

Published Online: 4 September 2025



V. Ankel,^{1,2} C. Bartram,^{2,3} J. Begin,⁴ C. Bell,^{1,2} L. Brouwer,⁵ S. Chaudhuri,⁴ John Clarke,⁶ H.-M. Cho,^{2,3} J. Corbin,^{1,2} W. Craddock,³ S. Cuadra,⁷ A. Droster,⁸ M. Durkin,^{9,10} J. Echevers,⁸ J. T. Fry,⁷ G. Hilton,⁹ K. D. Irwin,^{1,2,3} A. Keller,⁸ R. Kolevator,⁴ A. Kunder,^{1,2} D. Li,³ N. Otto,⁴ K. M. W. Pappas,⁷ N. M. Rapidis,^{1,2} C. P. Salemi,^{1,2,3,a)} D. Schmidt,⁹ M. Simanovskaia,^{1,2} J. Singh,^{1,2} P. Stark,^{1,2,3} C. D. Tesche,¹¹ J. Ullom,⁹ L. Vale,⁹ E. C. van Assendelft,^{1,2} K. van Bibber,^{8,12} M. Vissers,⁹ K. Wells,¹ J. Wiedemann,⁴ L. Winslow,⁷ D. Wright,^{2,13} A. K. Yi,^{2,3} and B. A. Young¹⁴

AFFILIATIONS

¹Department of Physics, Stanford University, Stanford, CA 94305, USA

²Kavli Institute for Particle Astrophysics and Cosmology, Stanford University, Stanford, CA 94305, USA

³Stanford Linear Accelerator Center, Menlo Park, CA 94025, USA

⁴Department of Physics, Princeton University, Princeton, NJ 08544, USA

⁵Accelerator Technology and Applied Physics Division, Lawrence Berkeley National Laboratory, Berkeley, CA 94720, USA

⁶Department of Physics, University of California, Berkeley, CA 94720, USA

⁷Laboratory of Nuclear Science, Massachusetts Institute of Technology, Cambridge, MA 02139, USA

⁸Department of Nuclear Engineering, University of California, Berkeley, CA 94720, USA

⁹National Institute of Standards and Technology, Boulder, CO 80305, USA

¹⁰Department of Physics, University of Colorado Boulder, Boulder, CO 80305, USA

¹¹Department of Psychology, University of New Mexico, Albuquerque, NM 87131-1161, USA

¹²Physics Division, Lawrence Berkeley National Laboratory, Berkeley, CA 94720, USA

¹³Department of Electrical Engineering, Stanford University, Stanford, CA 94305, USA

¹⁴Department of Physics, Santa Clara University, Santa Clara, CA 95053, USA

^{a)}Author to whom correspondence should be addressed: salemi@berkeley.edu

ABSTRACT

We present the limits on noise for the readout of cryogenic high-Q resonators using dc Superconducting Quantum Interference Devices (SQUIDs) below 300 MHz. This analysis uses realized first-stage SQUIDs (previously published), whose performance is well described by Tesche–Clarke (TC) theory, coupled directly to the resonators. We also present data from a prototype second-stage dc SQUID array designed to couple to this first-stage SQUID as a follow-on amplifier with high system bandwidth. This analysis is the first full consideration of dc SQUID noise performance referred to a high-Q resonator over this frequency range and is presented relative to the standard quantum limit. We include imprecision, backaction, and backaction–imprecision noise correlations from TC theory, the noise contributed by the second-stage SQUIDs, wiring, and preamplifiers, and optimizations for both on-resonance measurements and off-resonance scan sensitivity. This architecture has modern relevance due to the increased interest in axion searches and the requirements of the DMRadio-m³ axion search, which uses dc SQUIDs in this frequency range.

© 2025 Author(s). All article content, except where otherwise noted, is licensed under a Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>). <https://doi.org/10.1063/5.0280831>

10 October 2025 18:36:09

ACRONYMS

ADM	=	axion dark matter
ALP	=	axion-like particle
BSM	=	beyond the standard model
CDM	=	cold dark matter
DFT	=	discrete Fourier transform
DM	=	dark matter
DP	=	dark photon
DR	=	dilution refrigerator
FFT	=	fast Fourier transform
FOM	=	figure of merit
GUT	=	grand unification theory
LSW	=	light shining through wall
MQS	=	magneto-quasistatic
OFHC	=	oxygen-free, high-conductivity
ppb	=	part-per-billion
ppm	=	part-per-million
PQWW	=	Peccei–Quinn–Wilczek–Weinberg
PQ	=	Peccei–Quinn
PSD	=	power spectral density
PT	=	pulse tube
QCD	=	quantum chromodynamics
QED	=	quantum electrodynamics
QND	=	quantum non-demolition
SHM	=	standard halo model
SM	=	standard model
SNR	=	signal-to-noise ratio
SQL	=	standard quantum limit
SQUID	=	superconducting quantum interference device
TEM	=	transverse electromagnetic
TE	=	transverse electric
TM	=	transverse magnetic
WIMP	=	weakly interacting massive particle

I. INTRODUCTION

SQUIDs have been used as highly sensitive magnetic flux and current amplifiers for 60 years;^{1,2} dc SQUIDs are well established in applications ranging from fundamental physics to prospecting and medical imaging. In this paper, we focus on their use to read out high quality factor (Q), low-temperature resonators at very high frequencies (VHF) and below ($f < 300$ MHz). Resonant applications include Weber bars for gravitational wave detection,³ low-field nuclear magnetic resonance (NMR),⁴ magnetic resonance imaging (MRI)⁵ measurements, and searches for light–dark–matter candidates including quantum chromodynamics (QCD) axions and hidden photons.^{6–11} The sensitivity of these applications and experiments directly depends on the noise of the SQUID readout chain.

Weak coupling between a high- Q , low-temperature resonator and the first-stage dc SQUID can be optimal. In this weak-coupling regime, optimized dc SQUID performance is consistent with the seminal work by Tesche and Clarke^{12–14} (hereafter, referred to as TC). This theoretical analysis established the conditions to optimize the single-stage dc SQUID noise performance and allows us to calculate the imprecision and backaction noise, as well as their correlations, at optimized conditions. The imprecision noise is defined as the noise when no input circuit is attached and the input terminals

are open; in which case, there can be no backaction-current response of the input circuit (and, hence, no backaction noise contribution on the output). When coupled to an input circuit, the backaction noise term presents as a voltage applied to the input circuit, driving an input current signal that is partially correlated with the imprecision current noise in the SQUID. We typically refer all noise contributions to the SQUID input.

This work lays out the noise limits for a SQUID-based amplifier chain coupled to a resonator (see Fig. 1) under the conditions of weak input coupling, high-frequency bandwidth, and low dynamic range (thus relaxing the need for fast flux feedback). The full system noise comprises contributions from the first-stage dc SQUID as well as from second-stage SQUID circuits and downstream electronics, including those at room temperature. We present gain, impedance, and noise data measured for a prototype second-stage SQUID circuit for this application. We also discuss several options for follow-on amplification, including commercial options at cryogenic and room temperatures as well as a custom, high-bandwidth amplifier based on a previously published work. We finally combine the previously published theoretical and experimental analyses of first-stage SQUIDs with the second-stage data and follow-on amplifier specifications to consider the full system noise that can be achieved in weakly coupled resonators.

The noise performance of this SQUID amplifier chain can be optimized for different applications. In this work, we initially compute the on-resonance noise match conditions to minimize the dc SQUID noise temperature and optimize the resulting on-resonance performance. We also analyze the optimal noise temperature that can be achieved by a dc SQUID coupled to an arbitrary complex input impedance. We show that this noise temperature is equivalent to that which can be achieved by a single




FIG. 1. The full SQUID amplifier chain coupled to an input resonator with complex impedance Z_{res} via an input inductor L_{in} . S_{VV} is the PSD of the backaction voltage noise acting on the first-stage input. S_{II} is the PSD of the first-stage dc SQUID imprecision noise referred to a current in the input, and $S_{II,\text{ref}}$ is the imprecision noise of the rest of the amplifier chain referred to the first-stage input. The first- and second-stage SQUIDs are coupled via a small interconnect inductor, L_{int} . The first-stage SQUID is voltage biased in series with the input coil of the second-stage series-array dc SQUID. The output of the second-stage array is fed into a warm or cryogenic preamplifier with a $50\ \Omega$ input inductance via a $50\ \Omega$ -impedance cable.

10 October 2025 18:36:09

dc SQUID coupled to an LC resonator that has been optimally detuned to minimize noise temperature, as computed by Clarke.¹⁴ This result shows, for the first time, that an LC resonator can achieve the optimum noise match to a dc SQUID, albeit in a narrow bandwidth. We further extend this analysis to include the noise contributions from follow-on amplifier stages.

Applications, including wavelike dark matter searches, benefit from optimizing the noise match to maximize the sensitivity of a wide frequency scan, rather than maximizing sensitivity on resonance or in a narrow bandwidth around a particular detuned frequency. The amplifier chain described in this paper, while applicable to a broad range of resonant applications, is well suited to resonant, low-mass axion dark matter searches such as the Dark Matter Radio experiments, specifically DMRadio-m³. Axion signals can be captured in resonant circuits, but to detect this tiny signal, the resonators must be optimally coupled to low-noise amplifiers. The expected signals have a small dynamic range; however, the signal frequency is unknown, so the amplifier must be able to operate over a wide bandwidth of resonator frequency tuning (5–200 MHz in the case of DMRadio-m³).^{6,7}

II. DC SQUID RESPONSE AND NOISE

In this section, we use Tesche and Clarke (TC) SQUID noise theory^{12–14} to determine the noise properties of a first-stage dc SQUID. Dc SQUIDS optimized according to TC have measured white-noise levels consistent with the TC predictions, as long as they are properly biased and shielded and are weakly coupled to their input circuits (see, for example, Wellstood^{15,16}). Such SQUIDS have also shown a low-frequency noise component that is not predicted by TC theory, which is not significant for the resonator frequencies considered in this work.¹⁷

TC theory predicts the three response functions for an optimized dc SQUID: the flux-to-voltage response $(\partial V_{\text{out}}/\partial \Phi)|_{I_{\text{out}}}$, the flux-to-current response $(\partial I_{\text{out}}/\partial \Phi)|_{V_{\text{out}}}$, and the current-to-voltage response (the dynamic resistance) $(\partial V_{\text{out}}/\partial I_{\text{out}})|_{\Phi}$. Here, we will use these response functions to refer the TC noise terms to the input circuit as imprecision noise current (S_{II}), backaction noise voltage (S_{VV}), and their correlations (S_{IV}).¹⁸ TC theory is well established, but these parameters will be needed in this form for the analysis of optimized dc SQUID matching to HF and VHF resonant circuits in Secs. III–VIII.

The noise of a full dc SQUID chain may include contributions from the first- and second-stage dc SQUIDS, the wiring, and the room-temperature preamplifier. In principle, if the first-stage dc SQUID is coupled sufficiently strongly to the second-stage dc SQUID (which may be achieved at low frequency), the first-stage SQUID noise will dominate. We first consider the case of negligible follow-on amplifier chain noise and calculate the noise of the first-stage dc SQUID alone. We proceed in Sec. V to include the noise from the second-stage SQUID and the preamplifier.

According to TC noise analysis, optimal operation of a single, symmetric, weakly coupled dc SQUID is under the following conditions:




FIG. 2. Circuit diagram of a TC dc SQUID (blue highlight) coupled to an LRC series resonator (pink highlight). The dc SQUID loop inductance L_{sq} is divided symmetrically on both sides of the SQUID loop, coupling to the input coil L_{in} with mutual inductance $M = \kappa \sqrt{L_{\text{in}} L_{\text{sq}}}$, where $\kappa \in [0, 1]$ is the coupling constant. $\kappa = 1$ corresponds to full flux sharing between the two inductors, whereas $\kappa = 0$ corresponds to zero flux sharing. A variable transformer is used to tune the value of M to optimize the signal-to-noise ratio. S_{II} is the PSD of the dc SQUID imprecision noise referred to a current in the input. S_{VV} is the PSD of the backaction voltage noise, which drives a physical noise current in the resonator.

$$\begin{aligned}\beta_L &\equiv L_{\text{sq}} I_0 / \Phi_0 \approx 1, \\ \beta_C &\equiv \pi I_0 R_J^2 C_J / \Phi_0 \approx 1, \\ \Gamma &\equiv 2\pi k_B T_j / I_0 \Phi_0 \lesssim 0.05.\end{aligned}\quad (1)$$

Here, β_L and β_C describe the inductive and capacitive properties of the dc SQUID and Γ is a unitless figure of merit for the temperature of the electrons in the dc SQUID shunt resistance (this value is achievable for dc SQUIDS cooled in dilution refrigerators). As shown in Fig. 2, L_{sq} is the SQUID inductance (split symmetrically on each side of the SQUID loop), C_J is the stray capacitance of a single Josephson junction, I_0 is the critical current of a single junction, and R_J is the shunt resistance across a single junction. k_B is the Boltzmann constant, Φ_0 is the flux quantum, and T_j is the physical temperature of the electrons in the dc SQUID shunts.

When the dc SQUID parameters are optimized to follow Eq. (1) and are operated at optimum flux bias ($\Phi = \Phi_0/4$) and current bias ($I_{\text{out}} = 1.8I_0$) and at frequencies well below the Josephson frequency $f_J = V/\Phi_0$ (V being the voltage across the junction), we can define the following transfer functions.¹⁹ The flux-to-voltage transfer function for fixed bias current I_{out} is^{14,20}

$$\left. \frac{\partial V_{\text{out}}}{\partial \Phi} \right|_{I_{\text{out}}} \approx \frac{R_J}{L_{\text{sq}}}. \quad (2)$$

The dynamic resistance of the SQUID $(\partial V_{\text{out}}/\partial I_{\text{out}})|_{\Phi}$ is plotted but not numerically stated in Ref. 12. In Appendix A of this

work, data from Refs. 12 and 20 are numerically fit to determine that

$$R_{\text{dyn}} = \left. \frac{\partial V_{\text{out}}}{\partial I_{\text{out}}} \right|_{\Phi} \approx R_j, \quad (3)$$

which is twice the resistance of the shunts across both junctions in parallel ($R_j/2$).²¹

While TC analysis was conducted with a current-biased dc SQUID, the dc SQUID readout chain that we present in Fig. 1 and Sec. IV relies on a voltage-biased first-stage dc SQUID. It should be emphasized that the transfer functions [Eqs. (2) and (3)] represent the low-frequency (< 300 MHz) average response. The ac Josephson junction dynamics occur near the much higher Josephson frequency (~ 10 GHz for typical first-stage dc SQUIDs) and its harmonics. TC analysis is conducted with a constant current bias, and this constant current bias must be maintained at the Josephson frequencies for TC theory to be valid. (In fact, a voltage bias on a Josephson junction at the Josephson frequencies would lead to the trivial solution of a linear temporal evolution of the phase across the junction.)

If a voltage bias is applied across an inductance in series with the dc SQUID, then a current bias can be maintained at the Josephson frequency while the first-stage dc SQUID is voltage-biased at VHF and HF frequencies.²² The impedance of the inductance must be large as compared to the dc SQUID shunt resistance at the Josephson frequency and its relevant harmonics but small enough for the necessary response bandwidth in the VHF and HF bands. The inductance of the input coil of the second-stage SQUID is typically shorted out by the coil capacitance at microwave frequencies, so the microwave inductance must be provided by a reactive element with high enough impedance at the Josephson frequency and its harmonics. Sometimes, a wirebond ($L \sim 1$ nH) is sufficient for this reactance. Sufficient inductance for well-behaved voltage-biased operation of the first-stage in a two-stage dc SQUID is routinely achieved in applications including astronomical bolometric instruments, which are routinely and robustly implemented with large dc SQUID channel counts.^{23–25} As will be discussed in Sec. V, we assume that an additional 1 nH inductance with good microwave properties is placed in series with the first-stage dc SQUID.

If the condition of current-bias at the Josephson frequency and voltage-bias at HF and VHF frequencies is obtained, then we can apply Euler's chain rule to Eqs. (2) and (3) to derive (see Appendix B) the final transfer function,

$$\left. \frac{\partial I_{\text{out}}}{\partial \Phi} \right|_{V_{\text{out}}} \approx -\frac{1}{L_{\text{sq}}}, \quad (4)$$

the flux-to-current response of a voltage-biased first-stage dc SQUID.

TC provides expressions for the noise of an optimized current-biased dc SQUID with no circuit attached to its input. The output voltage noise power spectral density (PSD) is¹³

$$S_{VV_{\text{out}}} \approx 16k_B T_j R_j. \quad (5)$$

This noise arises from the Johnson thermodynamic noise in the junction resistance modified by the Josephson oscillations in the active dc SQUID circuit. Since it is computed with an open input circuit, it can be interpreted as the imprecision noise of the first-stage dc SQUID (referred as a voltage on its output). The full system imprecision noise will also include uncorrelated imprecision noise contributions from later amplifier stages (Sec. ??).

TC also computes the physical circulating current noise in the SQUID loop when it is not coupled to an input circuit. From Ref. 13, this calculated physical circulating current noise has PSD,

$$S_{II_{\text{circ}}} \approx 11k_B T_j / R_j. \quad (6)$$

This circulating current exists even when the input circuit is open, so it sources a portion of the output imprecision noise that is in Eq. (5). $S_{VV_{\text{out}}}$ and $S_{II_{\text{circ}}}$ are, thus, partially correlated. In addition, when connected to an input circuit, $S_{II_{\text{circ}}}$ gives rise to a backaction voltage on the input circuit that will be calculated in Eq. (11).

The correlations between the output voltage noise and the circulating current noise can be described by their cross correlation¹³

$$S_{I_{\text{circ}}V_{\text{out}}} \approx 12k_B T_j, \quad (7)$$

and we note that this cross correlation is real.

The above Eqs. (5)–(7) do not take into account quantum noise, which is a small correction for the dc SQUIDs described here, as they operate above the standard quantum limit (SQL). However, we will compare these results to the SQL to confirm the validity of this assumption.

We now calculate the noise when the dc SQUID is inductively coupled to an input circuit, under the assumption that this coupling is weak enough that Eqs. (5)–(7) still apply, and that signals are small enough that the flux remains close to $\Phi_0/4$.

We make no assumptions yet about the form of the input circuit other than to specify how it is coupled, with mutual inductance, M , between the input circuit and the dc SQUID loop,

$$M = \kappa \sqrt{L_{\text{sq}} L_{\text{in}}}. \quad (8)$$

As shown in Fig. 2, here, L_{sq} is still the SQUID inductance, L_{in} is the inductance of SQUID's input coil, and $\kappa \in [0, 1]$ is a constant describing the efficiency of the input coupling.

We refer all the TC noise terms to an equivalent noise on the input circuit: the referred imprecision current noise S_{II} , the physical voltage backaction noise S_{VV} , and their correlations S_{IV} . First, we consider the imprecision noise of the first-stage dc SQUID, which is referred to an output voltage in Eq. (5). Combining Eq. (5) with Eqs. (2) and (8) gives the imprecision noise referred to a current PSD in the input circuit,

$$S_{II} \equiv \frac{S_{VV_{\text{out}}}}{\left(\left. \frac{\partial V_{\text{out}}}{\partial \Phi} \right|_{I_{\text{out}}} \right)^2 M^2} \approx \frac{16k_B T_j L_{\text{sq}}}{\kappa^2 R_j L_{\text{in}}}. \quad (9)$$

From Eqs. (6) and (8), the PSD of the physical flux fluctuations in the input circuit, caused by the circulating dc SQUID current, is

$$M^2 S_{II\text{circ}} \approx \frac{11k_B T_j \kappa^2 L_{\text{in}} L_{\text{sq}}}{R_j}. \quad (10)$$

The backaction voltage applied to the input circuit is the time derivative of the physical flux fluctuations in the input circuit so that

$$S_{VV} \equiv \omega^2 M^2 S_{II\text{circ}} \approx \frac{11k_B T_j \kappa^2 L_{\text{in}} L_{\text{sq}} \omega^2}{R_j}. \quad (11)$$

We can similarly use Eq. (7) to derive the cross correlation between the input-referred imprecision noise S_{II} and the backaction noise S_{VV} . The voltage backaction on the input is the derivative of the coupled flux, whereas the current imprecision on the input has no time derivative, so the real TC correlation in Eq. (7) produces a purely imaginary correlation. As we will later show, the square of this correlation is important for determining the minimum noise temperature when coupled to a resonator, so it is what we compute

$$(\text{Im}\{S_{IV}\})^2 \approx \left(\frac{12k_B T_j L_{\text{sq}} \omega}{R_j} \right)^2. \quad (12)$$

SQUID noise figures can be presented in terms of an “uncoupled energy sensitivity” (ϵ_{uc}), which refers all imprecision noise to the equivalent flux noise in the SQUID loop if it is completely uncoupled from its input (so that no backaction contributes to the output noise). For TC analysis and the approximations above,

$$\epsilon_{\text{uc}} = \frac{S_{\Phi\Phi}}{2L_{\text{sq}}} \approx \frac{8k_B T_j L_{\text{sq}}}{R_j}. \quad (13)$$

The prefactor is sometimes listed as 9 instead of 8 (e.g., Ref. 26). This slightly higher value for ϵ_{uc} corresponds to a slightly less steep flux response function, which is appropriate for $\Gamma = 0.05$, whereas we assume a somewhat lower Γ . The difference is not significant.

In practical dc SQUIDS, $\epsilon_{\text{uc}} > \hbar$. However, this is not a quantum limit (it does not even include backaction) but rather a practical limit from combined design constraints on junction and stray capacitance, inductance, critical current, resistance, and self-heating temperature of the resistive shunts.

The uncoupled energy sensitivity will be used in Sec. VII in computing the optimal matching of a dc SQUID for integrated sensitivity over frequencies detuned from the resonance, rather than just on-resonance sensitivity.

III. MATCHING TO A RESONATOR

When the inefficiently coupled input transformer of the dc SQUID is connected to a series RLC resonant circuit (See Fig. 2), the total impedance of the resonator circuit is

$$Z_{\text{res}}(\omega) = R + i \left(\omega L_{\text{tot}} - \frac{1}{\omega C_{\text{res}}} \right), \quad (14)$$

where $L_{\text{tot}} = L_{\text{res}} + L_{\text{in}}$ is the sum of the resonator inductance and the input inductance of the dc SQUID transformer. The resonant frequency and the quality factor are, thus,

$$\omega_0 = \frac{1}{\sqrt{L_{\text{tot}} C_{\text{res}}}}, \quad (15)$$

$$Q = \frac{\omega_0 L_{\text{tot}}}{R}. \quad (16)$$

A. SQUID noise temperature on resonance

The noise temperature T_n of the dc SQUID is determined by the imprecision, backaction, and their correlations. We first consider the special case of the noise that will be achieved by the dc SQUID perfectly on resonance ($\omega = \omega_0$). This is relevant in cases where the resonator is driven on-resonance or is measured when not being driven.

On resonance, the imaginary component of the input circuit impedance is zero, giving $Z_{\text{res}}(\omega_0) = R$. For a dc SQUID described by TC theory, $\text{Re}\{S_{IV}\} = 0$. Using the dual of Eqs. (5.73) and (5.74) in Refs. 27, with the extra factor of 2 from our two-sided PSD, we arrive at

$$4k_B T_n(\omega_0) R = S_{VV} + R^2 S_{II}. \quad (17)$$

The minimum noise temperature (“noise match”) is, therefore, achieved on resonance when the input circuit impedance is

$$R = \sqrt{\frac{S_{VV}}{S_{II}}}. \quad (18)$$

From Eqs. (17) and (18), the minimum noise temperature on resonance ($T_{\text{min}}(\omega_0)$) is, thus,

$$k_B T_{\text{min}}(\omega_0) = \frac{1}{2} \sqrt{S_{VV} S_{II}}. \quad (19)$$

This quantity is independent of correlations in the dc SQUID noise due to matching to a real impedance (as is the case on resonance). In Sec. III B, we extend this analysis to a complex matching impedance, and as will be seen in Sec. VII, the correlations are important for the off-resonance sensitivity of a dc SQUID in applications including axion searches.

We now consider the special case where the follow-on amplifiers introduce negligible additional noise. In principle, this condition can be achieved at low frequency with strong coupling to a second-stage amplifier. (We consider the imprecision noise contribution of the follow-on amplifiers in Sec. IV.) From Eqs. (9), (11), and (18),

$$R \approx \frac{\sqrt{11}}{4} \kappa^2 \omega_0 L_{\text{in}}. \quad (20)$$

We define the “global coupling efficiency” to the resonator:

$$\kappa_g^2 \equiv \kappa^2 \frac{L_{\text{in}}}{L_{\text{tot}}}. \quad (21)$$

Here, $\kappa_g \in [0, 1]$. Perfect global efficiency ($\kappa_g = 1$) corresponds to perfect SQUID coil coupling ($\kappa = 1$), and the condition that the input coil inductance $L_{\text{in}} = L_{\text{tot}}$ is the only inductance in the resonator.

From Eq. (20), the on-resonance condition for noise matching the dc SQUID to the resonator is then

$$\kappa_g^2 \approx \frac{4}{\sqrt{11}} \frac{R}{\omega_0 L_{\text{tot}}} = \frac{4}{\sqrt{11}} \frac{1}{Q} \approx \frac{1.2}{Q}. \quad (22)$$

This value of κ_g represents the optimal coupling efficiency. Any deviation from the optimal coupling constant—either higher or lower—worsens the noise temperature. The value of κ_g can be tuned with a variable input transformer, as is described in Sec. IV A.

It is evident that the optimal global coupling efficiency κ_g is very small for a TC-optimized dc SQUID coupled to a high- Q resonator. It is in this regime of inefficient dc SQUID coupling that TC theory applies. Note that due to this weak coupling, the extra damping of the resonator by the dc SQUID (due to lossy shunt resistors)^{28,29} is reduced compared to in a strong coupling configuration; thus, SQUID-induced loss can potentially be subdominant to the internal loss of the resonator. Increasing internal resonator Q to arbitrarily high values by reducing the internal loss (R) does not make it more difficult for SQUID loss to be made negligible, because the optimal coupling (κ_g) correspondingly decreases.

If the noise match condition is achieved, the minimum noise temperature for a TC-optimized dc SQUID on resonance is

$$k_B T_{\text{min}} \approx 2\sqrt{11} \frac{k_B T_j L_{\text{sq}} \omega_0}{R_j}. \quad (23)$$

We note that the SQL for the amplifier noise temperature is $k_B T_{\text{min}} = \hbar \omega_0 / 2$ (one half photon of added noise), and that TC theory does not explicitly include quantum noise and thus is not valid in the low-noise limit. The ratio of the noise temperature in Eq. (23) to the SQL for amplifier-added noise is

$$\frac{\sqrt{\mathcal{S}_{VV} \mathcal{S}_{II}}}{\hbar \omega_0} \approx 4\sqrt{11} \frac{k_B T_j L_{\text{sq}}}{\hbar R_j}. \quad (24)$$

For most practical designs, $k_B T_{\text{min}}$ is far enough above the SQL that quantum corrections can be assumed to be small. Although it is possible to design SQUIDS with noise close to the quantum limit, those designs use small junctions and loop areas, with inductance too low to couple to large objects like lumped-element MHz resonators. Thus, for the remainder of this work, we neglect quantum noise contributions.

B. SQUID noise temperature for complex input impedance

If we instead couple the input to a complex impedance (as is the case for an input resonator off resonance), correlations can either reduce or increase the noise temperature. From the dual of Eq. (5.76) in Ref. 27, the minimum achievable noise temperature for an optimal complex source impedance is

$$k_B T_{\text{min}} = \frac{1}{2} \sqrt{\mathcal{S}_{VV} \mathcal{S}_{II} - (\text{Im}\{\mathcal{S}_{IV}\})^2} + \frac{1}{2} \text{Re}\{\mathcal{S}_{IV}\}. \quad (25)$$

For a TC-optimized dc SQUID, $\text{Re}\{\mathcal{S}_{IV}\} = 0$, and therefore, the minimum noise temperature for a matched complex input impedance is

$$k_B T_{\text{min}}(\omega) \approx 2\sqrt{2} \frac{k_B T_j L_{\text{sq}} \omega}{R_j}. \quad (26)$$

To our knowledge, this is the first time that the minimum noise temperature of a TC-optimized dc SQUID has been calculated for an *arbitrary* complex input impedance. However, Clarke *et al.*¹⁴ have previously determined the minimum noise-temperature of a dc SQUID coupled to an input LCR resonator without a transformer. Clarke *et al.* optimized the minimum dc SQUID noise temperature by varying C and R of the resonator while holding the frequency and L fixed, finding¹⁴

$$k_B T_{\text{min}} = 2.8 \frac{k_B T_j L_{\text{sq}} \omega}{R_j}, \quad (27)$$

which is in close agreement with Eq. (26). From Clarke *et al.*'s result combined with the calculation above, we conclude that a dc SQUID can reach its global optimal minimum noise temperature when coupled to an LCR resonator of the optimal Q and detuned resonance frequency.

We define η as the ratio of the minimum noise temperature of an amplifier (with the optimal complex impedance coupled to its input) to the quantum limit,

$$\eta(\omega) = \frac{k_B T_{\text{min}}(\omega)}{\hbar \omega / 2}. \quad (28)$$

For a TC-optimized single-stage dc SQUID, then,

$$\eta_{\text{TC}} \approx 4\sqrt{2} \frac{k_B T_j L_{\text{sq}}}{\hbar R_j}, \quad (29)$$

which is independent of the frequency ω .

Equation (24) represents the ratio of the noise temperature of a TC-optimized dc SQUID on resonance to the SQL, whereas Eq. (29) is the ratio to the SQL for coupling to the optimal complex impedance. Thus, Eq. (29) defines the fundamental limit on this SQUID's noise performance. We will see in Sec. VII that η is a useful parameter for scan sensitivity, including off-resonance operation.

IV. AMPLIFIER CHAIN

Up to this point, we have presented an analysis of the noise of a single, weakly coupled dc SQUID optimized according to Tesche and Clarke. In this section, we describe the implementation of a two-stage dc SQUID chain and a room-temperature preamplifier. We cite previous experimental results for the first-stage dc SQUID and room-temperature preamplifier and present here data taken

with a prototype second-stage dc SQUID series array designed to inform this analysis. In Sec. V, we analyze the contribution of all stages to the noise and the bandwidth.

The dc SQUID amplifier chain consists of a tunable transformer to optimize the noise match of the dc SQUID to the input circuit, a voltage-biased first-stage single dc SQUID optimized according to TC theory, and a second-stage dc SQUID array with 50Ω output impedance that couples to a room-temperature preamplifier with 50Ω input impedance (see Fig. 2), allowing high bandwidth operation over a limited dynamic range.

The design drivers for such an amplifier chain (originally based on DMRadio-m³ requirements, which are described in Sec. VII) are the following:

- (a) The dynamic range of the input signal is very small.
- (b) The lowest signal frequency is above the $1/f$ knee of the SQUID amplifiers and preamplifier electronics.
- (c) The input signal has a narrow bandwidth (set by the resonator on the input).
- (d) Noise matching to the resonator on the input requires a weakly coupled input.

These drivers have important consequences in the design of the amplifier chain:

- (i) The small dynamic range of the input signal permits open-loop operation of the SQUID amplifier chain. We can use a simple, slow flux feedback to stabilize its gain and impedance match between amplifier stages.
- (ii) The small dynamic range also implies that the second-stage SQUID can be maintained at optimal (50Ω) impedance match to the room-temperature preamplifier through its full signal swing, helping to mitigate signal reflections and optimize the bandwidth.
- (iii) The relatively high lowest signal frequency is above the $1/f$ noise of the dc SQUID and avoids backgrounds that could prevent low dynamic range operations.
- (iv) The narrow bandwidth signal can help mitigate external interferences affecting the SQUID operation.
- (v) The weak input coupling allows the use of an input coil with low intracoil capacitance and low capacitive coupling to the first-stage SQUID, enabling a simple first-stage SQUID performance consistent with TC theory and limiting the degradation of Q from coupling to the loss in the dc SQUID.

Most modern dc SQUID applications require efficient coupling between the input circuit and the SQUID loop in order to maximize the signal-to-noise ratio (SNR). In many of these more efficiently coupled dc SQUIDs, the increased coupling leads to interactions between the input circuit and the SQUID itself both at signal frequencies and Josephson frequencies.^{30–33} When input coils are efficiently coupled, non-ideal effects including intracoil capacitance, capacitance between the input coils and the SQUID, internal resonances, and inductive screening of the dc SQUID lead to deviations from the predicted TC noise performance. The implications of the imprecision, backaction, and correlation terms have been previously considered when the SQUIDs are tightly coupled to the resonator.^{14,30–34} They have also been considered in the

weakly coupled limit, showing some of the same scaling that we find, but with SQUIDs that are not optimized according to TC theory.³⁵

To date, the full implications of these imprecision and backaction noise terms and their correlations over the frequency range up to 300 MHz have not been referred to noise performance in a weakly coupled high- Q resonator. This analysis is presented here.

The first- and second-stage dc SQUIDs as well as the room-temperature preamplifier are based on designs that have been realized in practice.

A. Tunable transformer

As seen in Sec. III, the noise performance of the amplifier depends on the backaction and imprecision noise (and their correlations), which can be traded by tuning the coupling, κ , of the first-stage SQUID to the high- Q resonator. In applications where the resonator frequency will be tuned, such as DMRadio-m³, the optimal balance of imprecision and backaction noise may be different across different frequencies, and thus, κ is tunable *in situ*. We achieve this by implementing a SQUID input transformer consisting of a mechanically tuned variable mutual inductance. We have demonstrated initial prototypes consisting of gradiometrically wound solenoidal coils on concentric sliding tubes, forming a variable mutual inductance, and actuated with a piezoelectric motor (similar to, e.g., Ref. 36).

B. First-stage SQUID

The first-stage dc SQUID considered here is based on SQUID C1 of Ref. 15, with additional details provided in Ref. 16. This dc SQUID has $I_0 = 6.3\mu\text{A}$, $R_j = 6\Omega$, and $L_{sq} = 200\text{ pH}$ and is weakly coupled to its input coil. In that work, it was shown that TC theory is consistent with the measured white noise of the dc SQUID when it is cooled to very low temperatures and the amplifier noise and low-frequency noise are subtracted. The resistive shunts in the dc SQUID self-heat to $T_1 \approx 150\text{ mK}$. Using Eq. (24) with these parameters, it is seen that SQUID C1 of Ref. 15 performs at approximately 9 times the SQL for optimal noise matching to a real impedance (e.g., on resonance), with no additional noise sources. From Eq. (29), it is seen that $\eta_{TC} \approx 3.7$, so SQUID C1 would perform at about 4 times the SQL for optimal noise matching to a complex impedance.

The construction of a C1-like dc SQUID can be updated for a modern fabrication process. Updates include using a niobium/aluminum trilayer for the Josephson junctions in place of the Nb-NbO_x-PbIn junction process originally used (described in Ref. 16).

SQUID C1 was demonstrated up to 200 MHz, and subsequently, similar SQUIDs were operated up to 300 MHz.³⁷ This paper considers frequency response to 300 MHz (the top of the VHF band).

C. Prototype second-stage SQUID

The second-stage dc SQUID deviates significantly from an optimal TC dc SQUID because it is designed to operate reliably after the first stage of amplification, with parameters chosen to

tolerate parasitic resonances with the input coil. Some internal microwave filters are implemented to damp these resonances and to block coupling between the SQUIDs at the Josephson frequency. The second-stage dc SQUID is based on a series array of dc SQUIDs very similar to those used in the NIST series-array dc SQUIDs that have been deployed broadly in CMB, millimeter wave, and x-ray experiments worldwide.^{23–25} These dc SQUIDs are asymmetric and wound as serial gradiometers, minimizing coupling to environmental signals. Each input coil is a single turn, which reduces the parasitic capacitance.

We conducted detailed measurements of the performance of a prototype second-stage dc SQUID array, consisting of a series configuration of 20 dc SQUIDs, and extracted the necessary parameters for our analysis. These measurements form the basis for our design of the second-stage dc SQUID arrays. This SQUID has an asymmetric design, which provides internal self-feedback by coupling the output SQUID current from the second-stage back into the second-stage SQUID loop. This self-feedback is negative on the shallow slope of the SQUID response (column “– feedback slope”) and positive on the steep slope (column “+ feedback slope”). In the SQUID chain under consideration, we use the steep, positive feedback slope since it optimizes both the responsivity and the dynamic resistance on the output for high-bandwidth performance.

The prototype was cooled in a liquid-helium bath to $T_2 = 4$ K. The output of the SQUID was connected through low-resistance Cu twisted triple (including ground) to a commercial Magnicon XXF-1 preamplifier. The output was further amplified and then digitized for analysis. All data were taken open-loop (some data were validated by comparison to flux-locked-loop measurements). The following parameters were measured and are tabulated in Table I: I_c min is the minimum critical current of the two SQUID junctions in parallel (at $\Phi_0/2$ flux bias), measured by determining

the bias current at which the first voltage response to input flux is observed. I_c max is the maximum critical current of the two junctions in parallel (at zero flux bias), determined by the bias current at which a voltage response was measured for all input flux values. $M_{FB,2}$ is the mutual inductance of the feedback coil to each SQUID in this second-stage SQUID array. We measure $M_{FB,2}$ by measuring the feedback current required for one full flux quantum of modulation. $M_{IN,2}$ is the mutual inductance to the input coil of each SQUID, measured in the same way as $M_{FB,2}$ except with the current applied to the input loop. Both $M_{IN,2}$ and $M_{FB,2}$ refer to the mutual inductances of individual SQUIDs in the array and are thus independent of the number of SQUIDs in the array.

The following data were taken at an optimal bias point ($\Phi \approx \Phi_0/2$) on both the positive and negative feedback slopes. P is the measured power dissipated in the full SQUID array. Preamp $V_{n,PA}$ and $I_{n,PA}$ are the voltage noise and the current noise, respectively, of the Magnicon room-temperature preamplifier used in the measurement. These values were specified by Magnicon and confirmed in the laboratory. $\partial\Phi/\partial I|_{20\times 1}$, $\partial V/\partial I|_{20\times 1}$, and $\partial V/\partial\Phi|_{20\times 1}$ are the three partial derivatives of the output voltage V , output current I , and input flux Φ of the 20×1 array, with the unlabeled parameter held fixed. The former is determined by Euler’s chain rule (see Appendix B) from the latter two measurements.

The noise at the optimal bias is white above 1 kHz with no obvious pickup lines. To accurately measure this level, the noise is averaged between 1 and 100 kHz with an SR760 100 kHz spectrum analyzer and referred to a flux noise in the input of the second-stage SQUID. Φ_{n2} total is the total measured effective flux noise referred to the input of the second-stage SQUID array at $T_2 = 4$ K. Preamp Φ_n is the Magnicon preamplifier input voltage and current noise (added in quadrature) referred to a flux noise at the input of the second-stage SQUID at the optimal bias point. SQUID $\Phi_{n,20\times 1}$ is the second-stage SQUID flux noise with the preamplifier noise subtracted in quadrature. Absolute noise levels are calibrated to flux by injecting different input fluxes at known amplitude and frequency.

TABLE I. Measured parameters of a prototype second-stage series-array SQUID. The SQUID array measured here consists of 20 SQUIDs in series. The measurements are taken at $T = 4$ K, at the bias point with optimum voltage-to-flux response (near $\Phi = \Phi_0/2$). Dynamic parameters of the 3 second-stage SQUID designs can be scaled from the parameters measured here.

20 × 1 SQUID	– feedback slope	+ feedback slope
I_c min	3.055 μ A	Same
I_c max	9.485 μ A	Same
$M_{FB,2}$	44 pH	Same
$M_{IN,2}$	105 pH	Same
P	0.834 nW	Same
Preamp $V_{n,PA}$	0.32 $\frac{nV}{\sqrt{Hz}}$	Same
Preamp $I_{n,PA}$	2.7 $\frac{pA}{\sqrt{Hz}}$	Same
$\partial\Phi/\partial I _{20\times 1}$	231 pH	95.5 pH
$\partial V/\partial I _{20\times 1}$	154 Ω	64.9 Ω
$\partial V/\partial\Phi _{20\times 1}$	6.66×10^{11} Hz	6.80×10^{11} Hz
Φ_{n2} total at $T_2 = 4$ K	0.453 $\frac{\mu\Phi_0}{\sqrt{Hz}}$	0.394 $\frac{\mu\Phi_0}{\sqrt{Hz}}$
Preamp Φ_n referred	0.380 $\frac{\mu\Phi_0}{\sqrt{Hz}}$	0.259 $\frac{\mu\Phi_0}{\sqrt{Hz}}$
SQUID $\Phi_{n,20\times 1}$ at $T_2 = 4$ K	0.247 $\frac{\mu\Phi_0}{\sqrt{Hz}}$	0.297 $\frac{\mu\Phi_0}{\sqrt{Hz}}$

D. Three second-stage SQUID versions

Three different second-stage dc SQUID designs are considered in this work. Each design is scaled from the measured parameters in Table I. All designs are based on the positive feedback slope (the right column). For our scaled design, in addition to using the measured data, we take the input coil inductance of each individual second-stage component SQUID of $L_{SQ2} \approx 120$ pH (calculated using WRSlice³⁸) and the estimated inductance from the wirebonds and interconnect to the first-stage SQUID, L_{int} , which depends on the details of the packaging. Here, we take $L_{int} \approx 2$ nH.

The three designs will have different numbers of SQUIDs connected in series (N_{ser}) and in parallel (N_{par}). The total number of dc SQUIDs in the array is, thus, $N_{SQ} = N_{ser} \times N_{par}$. The standard NIST SQUID array has 384 SQUIDs wired in different series and parallel configurations. Experience has shown that, when resonances are avoided, the dynamic parameters of these SQUIDs scale well in configurations combining both series and parallel arrays, up to 384 SQUIDs (more than the largest array presented here, which

has 144). We have found that the parameters in Table I scale as theoretically expected,

$$P \propto N_{\text{ser}} \times N_{\text{par}}, \quad (30)$$

$$\left. \frac{\partial \Phi}{\partial I} \right|_V \propto \frac{1}{N_{\text{par}}}, \quad (31)$$

$$\left. \frac{\partial V}{\partial I} \right|_{\Phi} \propto \frac{N_{\text{ser}}}{N_{\text{par}}}, \quad (32)$$

$$\left. \frac{\partial V}{\partial \Phi} \right|_I \propto N_{\text{ser}}, \quad (33)$$

$$\Phi_{\text{nsquid}}^2 \propto \frac{T_2}{N_{\text{ser}} N_{\text{par}}}. \quad (34)$$

Here, Φ_{nsquid} refers to the equivalent imprecision flux noise in the SQUID array itself with an open input and without any contributions from follow-on amplification. Equation (34) also shows scaling with T_2 , the temperature of the second-stage dc SQUID array. We find that this linear scaling is maintained to, at least, as low as 1 K. At well below 1 K, self-heating of the electrons in the dc SQUID shunt resistor can limit further cooling.

All three second-stage dc SQUID array options are designed to have 50 Ω dynamic output resistance at the optimal bias point. The arrays are scaled from the experimentally measured dynamic resistance of the prototype array that was tested. If the dynamic range of the expected input signals is $\Delta\Phi \ll \Phi_0$, then the dynamic resistance should remain in this range. The output of the dc SQUID is carried through a coaxial cable with 50 Ω characteristic impedance to a room-temperature SQUID preamplifier with a low-noise, active 50 Ω input impedance. No fast flux feedback is provided from room temperature to the SQUIDs—they are operated in open-loop mode. It is expected that the bandwidth will be limited either by coupling between first- and second-stage SQUIDs or by the bandwidth of the preamplifier circuit itself.

Using Table I (biased on the positive feedback slope) and the scaling in Eq. (32), the dynamic output resistance of a 16×1 series array of this design is $(16/20) \times 64.9 \Omega \approx 50 \Omega$, which is the optimal dynamic resistance for high-bandwidth coupling to the output circuit, which consists of a coaxial cable with 50 Ω characteristic impedance and a room-temperature preamplifier with 50 Ω active input impedance. The flux bias of the second-stage SQUID is periodically adjusted to maintain the dynamic output resistance at $\approx 50 \Omega$ to maximize the bandwidth of coupling to the preamplifier and to minimize signal reflections. Expected signals are small and should not appreciably change the dynamic output resistance. Second-stage dc SQUID arrays of dimension 32×2 and 48×3 will also have $\approx 50 \Omega$ output dynamic resistance so that all three of these designs are well matched to the same output circuit.

An example set of the three different second-stage SQUID designs to cover a range of bandwidths with minimal noise is

TABLE II. Three different two-stage SQUID designs considered in this work (low, medium, and high bandwidth). The table includes the number of SQUID unit cells in the array $N_{\text{ser}} \times N_{\text{par}}$, dynamic resistance R_{dyn} , power dissipation P , input-coil self inductance L_2 , and time constant τ for the coupling between first- and second-stage SQUIDs, using $R_{\text{dyn}1} = 6 \Omega$. All three designs use the same SQUID unit cell as the prototype 20×1 second-stage SQUID with measured parameters in Table I, and all SQUID parameters are scaled according to Eqs. (30)–(34).

	Low BW	Medium BW	High BW
# SQUIDs	144 (48×3)	64 (32×2)	16 (16×1)
R_{dyn} (Ω)	50	50	50
P (nW)	6.00	2.67	0.67
L_2 (nH)	20	10	4
$\tau = L_2/R_{\text{dyn}1}$ (ns)	3.3	1.7	0.7

determined from Table I using the scaling in Eqs. (30)–(34). The three second-stage SQUID arrays considered here are as follows:

1. High bandwidth: a 16×1 series array SQUID. This series-array SQUID is identical to the measured SQUID in Table I except that it has $N_{\text{ser}} = 16$ active SQUIDs in series instead of 20. All input coils are wound in series. The total input self-inductance of this design is $L_2 = N_{\text{ser}} N_{\text{par}} L_{\text{SQ2}} + L_{\text{dummy}} + L_{\text{int}}$. Here, $L_{\text{dummy}} = 2N_{\text{par}} L_{\text{SQ2}}$ represents the additional total inductance from the input coils of one “dummy” SQUID at each end of every parallel bank of SQUIDs. These “dummies” are SQUIDs made with open junctions in order to symmetrize the array; they are not connected on the output and, thus, do not contribute power dissipation or noise. The resulting total input inductance is $L_2 = 18L_{\text{SQ2}} + L_{\text{int}} \approx 3$ nH.
2. Medium bandwidth: a 32×2 series array SQUID. This second-stage consists of $N_{\text{par}} = 2$ banks of $N_{\text{ser}} = 32$ series array SQUIDs, with the outputs of the two banks wired in parallel. The inputs of all SQUIDs are wound in series. The total input self-inductance of this design is $L_2 = 68L_{\text{SQ2}} + L_{\text{int}}$. (68 instead of 64 because of the two dummy SQUIDs at the end of each series array.)
3. Low bandwidth: a 48×3 series array SQUID. This second-stage consists of $N_{\text{par}} = 3$ banks of $N_{\text{ser}} = 48$ series array SQUIDs, with the outputs of the two banks wired in parallel. The inputs of all SQUIDs are wound in series. The total input self-inductance of this design is $L_2 = 150L_{\text{SQ2}} + L_{\text{int}}$.

In Table II, we present the dynamic resistance, power dissipation, self-inductance L_2 , and also the time constant limiting the bandwidth of the connection between the first- and second-stage SQUIDs, $\tau = L_2/R_{\text{dyn}1}$, where $R_{\text{dyn}1} = 6 \Omega$ is the output dynamic resistance of the first-stage SQUID described in Sec. IV B.

E. Room-temperature op-amp mode SQUID preamplifier

At frequencies below 50 MHz, a commercial “op-amp mode” preamplifier, the Magnicon XXF-1,³⁹ is suitable for reading out the second-stage SQUID. These preamplifier electronics have an open loop bandwidth of 50 MHz, voltage noise of $0.33 \text{ nV Hz}^{1/2}$, input current noise of $2.6 \text{ pA Hz}^{1/2}$, and low-noise input impedance of

50 Ω , established by internal negative feedback from the output of the first transistor stage.

At frequencies above 50 MHz, we consider an “op-amp mode” preamplifier based on ultra-high-speed SQUID electronics described in Refs. 40 and 41, which was designed for a small-signal open-loop bandwidth of 300 MHz. These electronics utilize discrete bipolar rf transistors to achieve low-noise operation up to 300 MHz. It achieves voltage noise of 0.3 nV Hz^{1/2}, input current noise of 6 pA Hz^{1/2}, and low-noise input impedance of 50 Ω , established by internal negative feedback from the output of the first transistor stage.

F. Cryogenic RF amplifier

We also consider an alternative implementation in which a commercial cryogenic “scattering mode” RF amplifier is used for high-frequency measurements and is diplexed to a Magnicon XXF-1 for low-frequency SQUID characterization and biasing.

One example of such an RF amplifier is the CITLF2 produced by Cosmic Microwave Technology, Inc.⁴² It possesses an input impedance of 50 Ω , so it is impedance matched to the 50 Ω output impedance of all second-stage SQUID designs considered here. The CITLF2 has a gain of 32 dB over a 5–500 MHz frequency range. It achieves a noise temperature of less than $T_{n,RF} = 2$ K from 5 to 500 MHz. To measure the low-frequency response of the second-stage SQUID array, and to maintain the bias point with low-frequency flux feedback, the RF amplifier can be preceded by a diplexer that routes the information at and near DC to Magnicon XXF-1 electronics.

V. SQUID CHAIN NOISE AND BANDWIDTH

As discussed in Secs. II and III, the noise performance of the first-stage SQUID is determined by its imprecision noise current, backaction noise voltage, and their correlations. The backaction noise voltage is dominated by the contribution of the first-stage SQUID, but the imprecision noise includes contributions from the full dc SQUID chain—not only the first-stage dc SQUID but also the second-stage dc SQUID, the room-temperature preamplifier, and Johnson noise from normal resistance in the wires between the second-stage dc SQUID and the preamplifier. The noise contributions from the later stages, which are uncorrelated with the first-stage SQUID noise, are referred to the input of the first stage where they add in quadrature to the imprecision current noise S_{II} of the first-stage dc SQUID [as given by Eq. (9)].

Although the second-stage dc SQUID and preamp stages backact on the first-stage dc SQUID, their backaction does not significantly increase the physical fluctuation currents in the input dc SQUID loop at the impedances and frequencies under consideration. This is because the output resistance of the first-stage dc SQUID is much higher than the noise impedance of the second-stage SQUID. Therefore, we will assume that the later stages do not modify the first-stage backaction noise S_{VV} [Eq. (11)]. Since the referred contributions of the later stages to the imprecision noise of the first stage are uncorrelated with the first-stage backaction, they also do not significantly affect S_{IV} [Eq. (12)].

We begin by examining the added imprecision noise from the second-stage SQUID and subsequently include the added

imprecision noise from the leads to the room-temperature preamplifier and from the room-temperature preamplifier itself.

A. Second-stage SQUID bandwidth and noise

1. Response bandwidth

The output of the second-stage SQUID is 50 Ω , and it is impedance matched to the characteristic impedance of the coaxial transmission line and the input impedance of the room-temperature preamplifier, which is fast enough that the second-stage-to-preamp bandwidth is greater than the 300 MHz target for the full system. The bandwidth of the first-stage input coil can be limited by parasitic capacitance, but it is weakly coupled in this implementation ($\kappa \ll 1$) so that it can also achieve the full-system 300 MHz bandwidth.

The bandwidth of the full SQUID chain is limited by the first-stage to second-stage coupling, which is not impedance matched. The first-stage-to-second-stage SQUID coupling introduces a low-pass single-pole rolloff with time constant $\tau = L_2/R_{dyn}$. The time constants of the three SQUID chains considered here are listed in Table II, ranging from 0.5 to 3 ns.

This time constant can be used to refer the second-stage SQUID flux noise Φ_{n2} to a first-stage input imprecision current noise as a function of frequency ω ,

$$S_{II,ref}(\omega) = \frac{(\Phi_{n2})^2}{(M_{IN,2})^2} (1 + \omega^2 \tau^2) \left/ \left(M \frac{\partial I_{out}}{\partial \Phi} \right|_{V_{out}} \right)^2 = \frac{(\Phi_{n2})^2}{(M_{IN,2})^2} \frac{L_{sq}}{\kappa^2 L_{in}} (1 + \omega^2 \tau^2). \quad (35)$$

Here, we have used Eq. (8) for M and Eq. (4) for the current responsivity of the first-stage dc SQUID.

The noise of the second-stage SQUID itself and the noise of the follow-on preamplifier can be referred to an equivalent flux noise Φ_{n2} in the second-stage SQUID. The noise components add in quadrature,

$$\Phi_{n2}^2 \approx \Phi_{n20 \times 1}^2 \frac{20}{N_{ser} N_{par}} \frac{T_2}{4K} + \Phi_{n2_AMP}^2. \quad (36)$$

The first term on the RHS is the flux noise of the second-stage SQUID itself, using Eq. (34) to scale from the 20-dc SQUID array measurements in Table I measured at 4 K. We expect to operate the second-stage SQUID at temperature $T_2 = 1$ K, which has proven to be a stable operation temperature for dc SQUIDs of this design (see, e.g.,⁴³).

The noise of the follow-on amplifier is referred to a flux noise in the second-stage SQUID, Φ_{n2_AMP} in Eq. (36). This follow-on amplifier can be either a room-temperature op-amp mode SQUID preamplifier (Sec. V B) or a cryogenic RF amplifier (Sec. V C).

B. Noise of op-amp mode SQUID preamplifier

The referred second-stage flux noise of the op-amp mode room-temperature SQUID preamplifier has contributions from the

noise of the room-temperature preamplifier plus Johnson noise of the leads to room temperature $\Phi_{n2_VPAL}^2$ [Eq. (38)], and the current noise of the room-temperature preamplifier $\Phi_{n2_IPA}^2$ [Eq. (39)], which are presumed to be uncorrelated,

$$\Phi_{n2_AMP}^2 = \Phi_{n2_VPAL}^2 + \Phi_{n2_IPA}^2. \quad (37)$$

The voltage noise of the room-temperature preamplifier and the Johnson voltage noise of the leads to room temperature are referred to a flux noise in the second-stage SQUID array,

$$\Phi_{n2_VPAL}^2 \approx \frac{(V_{n_PA})^2 + 4k_B T_{lead} R_{lead}(\omega)}{(\partial V / \partial \Phi|_{20 \times 1})^2} \left(\frac{20}{N_{ser}} \right)^2. \quad (38)$$

Here, V_{n_PA} is the preamp voltage noise. The voltage transfer function is increased proportional to the number of dc SQUIDs in series, N_{ser} , according to Eq. (32). R_{lead} and T_{lead} are the effective real resistance and temperature of the leads from the second-stage SQUID to the room-temperature op-amp mode preamplifier. R_{lead} represents the cable loss at the operational frequency (which is generally higher than dc resistance). This noise is conservative, as the increase in the critical current of niobium films as they are cooled below 4 K causes the flux-to-voltage transfer function to steepen slightly on cooling to 1 K. For this analysis, we assume superconducting coaxial cable from the second-stage SQUID at 1 K to where it is heat sunk at 4 K. Copper coaxial cable from 4 K to a room-temperature preamplifier, with $R_{lead} = 1 \Omega$ (assumed frequency independent to simplify the analysis) and $T_{lead} = 200$ K, is assumed.

The current noise of the room-temperature preamplifier is referred to a flux noise in the second-stage SQUID array,

$$\Phi_{n2_IPA}^2 \approx (I_{n_PA})^2 \left(\frac{\partial \Phi}{\partial I} \Big|_{20 \times 1} \frac{1}{N_{par}} \right)^2. \quad (39)$$

Here, I_{n_PA} is the preamp current noise and $\partial \Phi / \partial I|_{20 \times 1}$ is the current-to-flux transfer function for the 20-SQUID array. We assume that the resistance of the leads is small enough that a stiff current bias can be maintained from room temperature at HF and VHF signal frequencies.

C. Noise of cryogenic RF preamplifier

We also consider an implementation with a commercial cryogenic “scattering mode” RF amplifier as an alternative to the “op-amp mode” room-temperature preamplifier. Since the RF amplifier is operated at 4 K, it is connected to the second-stage dc SQUID through a superconducting coaxial cable, so we do not include any Johnson noise between the second-stage SQUID and the cryogenic RF amplifier. The noise temperature $T_{n_RF} \approx 2$ K of the cryogenic RF amplifier with 50Ω input impedance can be referred to an equivalent voltage noise at the output of the second-stage SQUID. We may refer this voltage noise to a flux noise in the second-stage SQUID by matching the dynamic resistance of the second stage, $\partial V / \partial I|_V = 50 \Omega$, with a matched 50Ω

RF amplifier input impedance, and using Euler’s chain rule (see Appendix B),

$$\Phi_{n2_AMP}^2 \approx -4k_B T_{n_RF} \frac{\partial \Phi}{\partial I_{out}} \Big|_{20 \times 1} \times \frac{\partial \Phi}{\partial V_{out}} \Big|_{20 \times 1} \left(\frac{20}{N_{ser} N_{par}} \right). \quad (40)$$

By substituting Eq. (40) for a cryogenic follow-on amplifier, rather than Eq. (37), into Eq. (36), one can follow the same noise analysis as outlined above to determine the resulting total noise performance of the system. The cryogenic RF amplifier can similarly operate with the low-, medium-, or high-bandwidth versions of the second-stage SQUID (Table II).

D. Full system noise

We can now compute the full-system imprecision current noise referred to the input of the resonator circuit as

$$S_{II_sys}(\omega) = S_{II} + S_{II_ref(\omega)}, \quad (41)$$

where S_{II} has the TC value of imprecision current noise from Eq. (9) and $S_{II_ref(\omega)}$ is the referred noise from the follow-on amplifier chain from Eqs. (35) and (36).

As discussed in Sec. III B, the minimum noise temperature that can be achieved requires an optimal complex input impedance connected to the SQUID input [see Eq. (26) for a computation of the minimum noise temperature for a TC-optimized first-stage SQUID]. This noise temperature will not be achieved with a real input impedance, which cannot take advantage of the fact that the backaction and imprecision noise are correlated, but it is still a useful parameter. The minimum full-system noise temperature of the full amplifier chain connected to an optimal complex input impedance is calculated from Eq. (25). The value of η can then be computed as the ratio of this noise temperature to the SQL for added noise, $\hbar\omega/2$. Once again, we take $\text{Re}\{S_{IV}\} = 0$ so that

$$\eta(\omega) = \frac{k_B T_{min}(\omega)}{\hbar\omega/2} = \frac{1}{\hbar\omega} \sqrt{S_{VV} S_{II_sys}(\omega) - (\text{Im}\{S_{IV}\})^2}, \quad (42)$$

where the TC value of S_{VV} is used from Eq. (11), the TC value of S_{IV} is used from Eq. (12), and the full-system imprecision current noise $S_{II_sys}(\omega)$ is determined by Eq. (41). For our assumed SQUID chain parameters, we plot the value of $\eta(\omega)$ for the three second-stage SQUID versions up to 300 MHz in Fig. 3. In the three solid lines, we use a Magnicon XXF-1 preamplifier below 50 MHz. Above 50 MHz, we assume the use of the room-temperature op-amp mode preamplifier described in Sec. IV E. Different second-stage SQUIDs are optimal from a noise perspective at different frequency ranges. The dashed lines were attained using the cryogenic RF preamplifier and the low- and medium-bandwidth second-stage SQUID variety, which seems to be optimal (or close to optimal) at all relevant frequencies for the cryogenic RF amplifier.




FIG. 3. Noise parameter η as a function of frequency for the full dc SQUID amplifier chain [from Eq. (42)]. Blue lines are the high-bandwidth second-stage SQUID, purple are the medium-bandwidth, and yellow are the low bandwidth (see Table II). Solid lines use room-temperature (RT) op-amp mode amplifiers: Magnicon XXF-1 below 50 MHz and the high bandwidth room temperature preamp above 50 MHz (Sec. IV E). Dashed lines use the CITLF2 cryogenic RF preamp (Sec. IV F) above 10 MHz. The dashed line with the high BW second-stage SQUID with the cryogenic RF preamp is omitted for clarity; this second-stage SQUID option might be useful only at frequencies above 300 MHz. As discussed in Ref. 7, the low- and medium-bandwidth configurations have sufficiently low η to serve DMRadio-m³ to reach its axion sensitivity goals. We have also included the SQL line ($\eta = 1$) for reference.

The full-system uncoupled imprecision energy sensitivity can be computed from the full system current noise,

$$\epsilon_{uc}(\omega) = \frac{M^2 S_{II_{sys}}(\omega)}{2L_{sq}} = \frac{\kappa^2 L_{in} S_{II_{sys}}(\omega)}{2}, \quad (43)$$

where $S_{II_{sys}}$ is from Eq. (41), which contains both the first-stage TC contribution calculated in Eq. (13), and the referred noise from the follow-on amplifiers. We plot ϵ_{uc} for the three second-stage SQUID versions and the room-temperature amplifier, as well as the medium-bandwidth second stage with the cryogenic RF amplifier, in Fig. 4.

As will be seen in Sec. VII, the values of $\eta(\omega)$ and $\epsilon_{uc}(\omega)$ computed in Eqs. (42) and (43), with results in Figs. 3 and 4, determine the optimal matching conditions and the frequency-integrated scan sensitivity for resonant axion searches.

VI. SQUID CHAIN NOISE MATCH ON RESONANCE

Equations (22) and (23) describe the noise-match condition and the optimal on-resonance noise temperature for an ideal TC dc SQUID with no follow-on amplifier noise. It is straightforward to extend this analysis to the full amplifier chain. On resonance, the input circuit presents a real impedance. Then, from Eq. (18), the result is a resonance-frequency-dependent noise resistance,

$$R(\omega_0) = \sqrt{\frac{S_{VV}}{S_{II} + S_{II_{ref}}(\omega_0)}}, \quad (44)$$




FIG. 4. The uncoupled energy sensitivity ϵ_{uc} (in units of $\hbar$) as a function of frequency for the full dc SQUID amplifier chain [from Eq. (43)]. Details of all lines are the same as in Fig. 3 caption. ϵ_{uc} for the low-bandwidth SQUID with the cryogenic RF preamp has values less than $7.6\hbar$ below 50 MHz.

where the backaction S_{VV} arises from the TC-optimized first-stage SQUID backaction voltage [Eq. (11)], S_{II} is the TC-optimized first-stage SQUID imprecision noise current [Eq. (9)], and the follow-on amplifier referred noise $S_{II_{ref}}(\omega)$ is from Eqs. (35) and (36). As was previously done for the TC first-stage SQUID in Sec. III A, Eq. (44) can be solved for the value of κ_g that is noise matched, achieving the minimum noise temperature. The value of κ_g can be tuned with a variable input transformer.

Similarly, from Eq. (19), we use the full system imprecision noise to arrive at the minimum on-resonance noise-temperature that can be achieved with the full amplifier chain:

$$k_B T_{\min}(\omega_0) = \frac{1}{2} \sqrt{S_{VV}(S_{II} + S_{II_{ref}}(\omega_0))}. \quad (45)$$

These parameters can be computed using the derivations in Sec. V.

It is evident from Eq. (44) that the additional imprecision noise from the follow-on amplifiers decreases the noise impedance of the amplifier, which can be compensated for by tuning the transformer to a stronger coupling (higher κ_g and mutual inductance M). From Eq. (45), the addition of the follow-on amplifier noise increases the minimum noise temperature at this new optimal match.

VII. SCAN SENSITIVITY NOISE MATCH

In Sec. VI, we described the matching condition for optimal noise temperature on resonance, where the resonator impedance is real. In on-resonance applications, when the coupling is optimized, the imprecision noise current is equal to the backaction-driven noise current. On resonance, imaginary correlations do not affect the total noise, which is a quadrature sum of the backaction and imprecision. The presence of a thermal population of photons in the resonator does not affect the on-resonance noise-match condition.

In some applications, however, information is used that is detuned from the resonance frequency. At detuned frequencies, the

backaction-driven current is rolled off by the Lorentzian line shape, while imprecision noise is not. The presence of a thermal photon population, which is also rolled off by the Lorentzian, can significantly change the optimal match condition. Furthermore, the resonance has a complex impedance, so imaginary dc SQUID correlations have a significant effect on the noise, either increasing (on one side of the Lorentzian) or decreasing it (on the other side).⁴⁴

One example of such an application is axion searches. The mass/frequency of the axion is unknown, so the primary figure of merit is not the signal-to-noise ratio at resonance but the integrated scan sensitivity, i.e., the square of SNR integrated over all frequencies and summed over all tuning steps. The calculation of the integrated scan sensitivity using an arbitrary flux-to-voltage amplifier with correlations is worked out in Ref. 44 [Eq. (F58)], and the full analysis of DMRadio-m³ sensitivity has been calculated using this formalism in Ref. 7. After a dc SQUID amplifier chain is matched through a tunable transformer to an input resonator to optimize scan sensitivity, only the noise parameter $\eta(\omega)$ is needed to determine the SNR of the axion search. We leave the full scan sensitivity calculation to Ref. 44 and just state the conclusions here.

The optimal noise-match condition for scan sensitivity, defined by the total system imprecision noise, is [Eq. (F55) in Ref. 44],

$$S_{I\text{sys}} = \frac{\hbar\omega}{R} \frac{2\eta^2}{2n+1 + \sqrt{(2n+1)^2 + 8\eta^2}}, \quad (46)$$

where n is the resonator's thermal occupation number,

$$n = \frac{1}{\exp(\hbar\omega/k_B T_{\text{res}}) - 1}. \quad (47)$$

Using Eq. (43) to express $S_{I\text{sys}}$ in terms of ϵ_{uc} , and also using the definition of κ_g from Eq. (21), we solve for the optimal matching condition on κ_g ,

$$\kappa_g^2 = \frac{1}{Q} \frac{\epsilon_{\text{uc}}}{\hbar} \frac{2n+1 + \sqrt{(2n+1)^2 + 8\eta^2}}{\eta^2}. \quad (48)$$

The matching condition in Eq. (48) optimizes the scan sensitivity for a Lorentzian resonator, including the effect of the correlations. As the thermal occupation of the resonator n increases in Eq. (48), the coupling κ_g must also increase to optimize the scan sensitivity. We further see that $\kappa_g^2 \propto 1/Q$ so that high- Q resonators require low global coupling efficiency. In the case of DMRadio-m³, the optimal coupling efficiency is low enough for the use of the weakly coupled SQUID design C1 of Ref 15, whose noise was accurately described by TC theory.

The noise properties of a dc SQUID amplifier chain for a scanning experiment such as an axion search are completely specified by the two parameters η and ϵ_{uc} , which we have presented in this paper for both TC-optimized single SQUIDs [Eqs. (13) and (29)] and the full, high-bandwidth dc SQUID amplifier chain [Figs. 3 and 4].

Now that we have optimized the noise match for scan sensitivity, we explicitly evaluate whether our analysis safely remains within the weak-coupling regime assumed throughout this paper. The weak-coupling assumption will hold for the on-resonance noise match if it holds for the scan-sensitivity noise match, since the latter requires stronger coupling.

We begin by considering a representative resonator frequency of 100 MHz, centrally located in the DMRadio-m³ science band (see Ref. 7 for detailed parameters). Assuming a temperature $T = 20$ mK, Eq. (47) yields a thermal occupancy $n \approx 3.7$ at 100 MHz. Using the experimentally relevant values of $Q = 134\,000$, noise parameter $\eta = 9.8$, and normalized energy sensitivity $\epsilon_{\text{uc}}/\hbar = 10.9$, we compute the global coupling efficiency via Eq. (48) as $\kappa_g \approx 0.006$. This result places our system in the weak-coupling regime. At lower frequencies, higher temperatures, or lower Q , the optimal coupling would increase; thus, similar analyses must be performed to verify the applicability of the weak-coupling approximation for specific experimental conditions.

Additionally, we assess whether the rms flux fluctuations in the SQUID induced by thermal fluctuations of the resonator remain within the linear response limit, a necessary condition since we do not implement fast feedback. Using the equipartition theorem, the resonator current fluctuations are estimated by

$$\frac{k_B T}{2} = \frac{1}{2} L I_{\text{rms}}^2. \quad (49)$$

For the considered case⁷ of 100 MHz, with inductance $L = 63$ nH and temperature $T = 20$ mK, this yields $I_{\text{rms}} \approx 2$ nA. The corresponding rms flux fluctuations coupled into the SQUID are computed using the mutual inductance $M = \kappa_g \sqrt{LL_{\text{sq}}}$. With the previously determined SQUID inductance $L_{\text{sq}} = 200$ pH (Sec. IV B), we find the resulting rms flux fluctuations to be approximately $\Phi_{\text{rms}} \approx 20 \mu\Phi_0$, placing these fluctuations comfortably within the SQUID's linear response regime.

VIII. CONCLUSIONS

In this paper, we show the limits on noise for the readout of high- Q resonators below 300 MHz using dc SQUIDs optimized according to Tesche Clarke theory, including practical implementations with follow-on amplifier chains. These analyses take into account SQUID backaction noise, imprecision noise, and their correlations. In Eq. (23), we show the minimum noise temperature achievable with a TC-optimized first-stage SQUID on resonance. We also find the minimum noise temperature that can be achieved with a matched complex impedance in Eq. (26), show that this noise temperature can be achieved with a detuned LCR resonator, and compare to the quantum limit in Eq. (29).

We present data from a realized prototype second-stage SQUID and show how it scales to three different second-stage SQUID designs with different bandwidths. We analyze the properties of several full amplifier chain designs constructed of realized components: a demonstrated TC-optimized first-stage SQUID, three second-stage SQUID designs scaled from the realized prototype presented in this paper, the commercial Magnicon XXF-1 room-temperature preamplifier and CITLF2 cryogenic RF preamp,

and the 300 MHz room-temperature preamplifier described in Sec. IV E. We analyze bandwidth, backaction noise, imprecision noise, and correlations.

We calculate how to optimally match these SQUID amplifier chains to a resonator to maximize scan sensitivity for axion searches, and we provide the two parameters, η and ϵ_{uc} , that specify the achieved noise and matching conditions. The values of these two parameters are shown as a function of frequency for several versions of the SQUID amplifier chain in Figs. 3 and 4.

This analysis is of importance to any application that requires optimal matching of weakly coupled dc SQUIDs to low-temperature, high-Q resonators below 300 MHz. The design is the basis for the SQUID amplifier chain for the DMRadio-m³ axion search and is used for its full electromagnetic design and science reach analysis in.⁷

ACKNOWLEDGMENTS

The authors acknowledge support for DMRadio-m³ as part of the DOE Dark Matter New Initiatives program under SLAC FWP 100559. C. Bartram and A. Yi were supported by the Department of Energy, Laboratory Directed Research and Development program at SLAC National Accelerator Laboratory, under Contract No. DE-AC02-76SF00515 and as part of the Panofsky Fellowship awarded to C. Bartram. S. Chaudhuri acknowledges support from the R. H. Dicke Postdoctoral Fellowship and Dave Wilkinson Fund at Princeton University. C. P. Salemi was supported by the Kavli Institute for Particle Astrophysics and Cosmology Porat Fellowship. J. Clarke acknowledges support from the U.S. Department of Energy, Office of Science, National Quantum Information Science Research Centers. Certain equipment, instruments, software, or materials are identified in this paper in order to specify the experimental procedure adequately. Such identification is not intended to imply recommendation or endorsement of any product or service by NIST, nor is it intended to imply that the materials or equipment identified are necessarily the best available for the purpose.

AUTHOR DECLARATIONS

Conflict of Interest

The authors have no conflicts to disclose.

Author Contributions

V. Ankel: Conceptualization (equal); Writing – review & editing (equal). **C. Bartram:** Conceptualization (equal); Funding acquisition (equal); Project administration (equal); Writing – review & editing (equal). **J. Begin:** Conceptualization (equal); Writing – original draft (equal). **C. Bell:** Conceptualization (equal); Writing – review & editing (equal). **L. Brouwer:** Conceptualization (equal); Writing – review & editing (equal). **S. Chaudhuri:** Conceptualization (equal); Funding acquisition (equal); Project administration (equal); Writing – review & editing (equal). **John Clarke:** Conceptualization (equal); Methodology (equal); Writing – review & editing (equal).

H.-M. Cho: Conceptualization (equal); Funding acquisition (equal); Writing – review & editing (equal). **J. Corbin:** Conceptualization (equal); Writing – review & editing (equal). **W. Craddock:** Conceptualization (equal); Writing – review & editing (equal). **S. Cuadra:** Conceptualization (equal); Writing – review & editing (equal). **A. Droster:** Conceptualization (equal); Writing – review & editing (equal). **M. Durkin:** Conceptualization (equal); Writing – review & editing (equal). **J. Echevers:** Conceptualization (equal); Writing – review & editing (equal). **J. T. Fry:** Conceptualization (equal); Writing – review & editing (equal). **G. Hilton:** Conceptualization (equal); Writing – review & editing (equal). **K. D. Irwin:** Conceptualization (equal); Data curation (equal); Funding acquisition (equal); Investigation (equal); Methodology (equal); Project administration (equal); Writing – original draft (equal); Writing – review & editing (equal). **A. Keller:** Conceptualization (equal); Writing – review & editing (equal). **R. Kolevaton:** Conceptualization (equal); Writing – review & editing (equal). **A. Kunder:** Conceptualization (equal); Writing – review & editing (equal). **D. Li:** Conceptualization (equal); Funding acquisition (equal); Writing – review & editing (equal). **N. Otto:** Conceptualization (equal); Writing – review & editing (equal). **K. M. W. Pappas:** Conceptualization (equal); Writing – review & editing (equal). **N. M. Rapidis:** Conceptualization (equal); Visualization (equal); Writing – original draft (equal); Writing – review & editing (equal). **C. P. Salemi:** Conceptualization (equal); Data curation (equal); Funding acquisition (equal); Investigation (equal); Methodology (equal); Project administration (equal); Visualization (equal); Writing – original draft (equal); Writing – review & editing (equal). **D. Schmidt:** Conceptualization (equal); Writing – review & editing (equal). **M. Simanovskaia:** Conceptualization (equal); Writing – review & editing (equal). **J. Singh:** Conceptualization (equal); Writing – review & editing (equal). **P. Stark:** Conceptualization (equal); Writing – review & editing (equal). **C. D. Tesche:** Conceptualization (equal); Writing – review & editing (equal). **J. Ullom:** Conceptualization (equal); Writing – review & editing (equal). **L. Vale:** Conceptualization (equal); Writing – review & editing (equal). **E. C. van Assendelft:** Conceptualization (equal); Writing – review & editing (equal). **K. van Bibber:** Conceptualization (equal); Funding acquisition (equal); Project administration (equal); Writing – review & editing (equal). **M. Visser:** Conceptualization (equal); Writing – review & editing (equal). **K. Wells:** Conceptualization (equal); Writing – review & editing (equal). **J. Wiedemann:** Conceptualization (equal); Writing – review & editing (equal). **L. Winslow:** Conceptualization (equal); Funding acquisition (equal); Project administration (equal); Writing – review & editing (equal). **D. Wright:** Conceptualization (equal); Writing – review & editing (equal). **A. K. Yi:** Conceptualization (equal); Writing – review & editing (equal). **B. A. Young:** Conceptualization (equal); Writing – review & editing (equal).

DATA AVAILABILITY

The data that support the findings of this study are available from the corresponding author upon reasonable request.

10 October 2025 18:36:09

APPENDIX A: RESPONSE FUNCTIONS

From Fig. 12(b) in Ref. 12 (all image processing done using Ref. 45), at optimally biased flux ($\Phi = \Phi_0/4$) and current ($I = 1.8I_0$), and temperature figure of merit (FOM) $\Gamma = 0.050$,

$$\frac{\partial V}{\partial \Phi} = 1.6 \frac{I_0 R_j}{\Phi_0}. \quad (\text{A1})$$

Assuming β_L as in Eq. (1), we can rewrite the flux-to-voltage transfer function as

$$\frac{\partial V}{\partial \Phi} = 0.8 \frac{R_j}{L_{sq}}. \quad (\text{A2})$$

It should be noted that the original Tesche and Clarke paper¹² had an incorrect voltage transfer function that was updated in Ref. 20. By reducing the temperature (to get lower values of Γ), the transfer function scales as shown in Fig. 1(c) of Ref. 20. At current bias of $1.8I_0$, from $\Gamma = 0.050 \rightarrow 0.025$, the transfer function scales by a factor of 1.15, giving

$$\frac{\partial V}{\partial \Phi} = 0.9 \frac{R_j}{L_{sq}} \quad (\text{A3})$$

for the lower temperature. Although temperatures lower than ~ 4.2 K ($\Gamma \approx 0.025$) are not included in Ref. 12 or 20, modern cryogenics and SQUID package design can achieve considerably cooler operating temperatures leading to a prefactor closer to 1 in Eq. (A3). We take

$$\frac{\partial V}{\partial \Phi} \approx \frac{R_j}{L_{sq}} \quad (\text{A4})$$

throughout this work.

Using similar bias current and flux, we can extract the dynamic resistance from Fig. 12(a) in Ref. 12,

$$\frac{\partial V}{\partial I} \approx 0.97 R_j \approx R_j. \quad (\text{A5})$$

We round the coefficient to unity for our calculations in the body of the paper, which is also the value used by Drung.⁴⁰ Note that this indicates an average voltage output of $0.38I_0 R_j$.

APPENDIX B: EULER'S CHAIN RULE FOR dc SQUIDS

If the conditions of current bias at the Josephson frequency and voltage bias at HF and VHF frequencies are obtained, then the relationship among the flux Φ , voltage V_{out} , and current I_{out} at VHF and HF frequencies can be expressed as a surface in Φ, V, I space: $f(\Phi, V_{\text{out}}, I_{\text{out}}) = 0$. At the TC optimal bias point, all partial derivatives of this function are nonzero so that, on this surface, locally, each variable is given as a differentiable function of the remaining two. Then, the triple product rule (Euler's chain rule) can be invoked,

$$\left. \frac{\partial I_{\text{out}}}{\partial \Phi} \right|_{V_{\text{out}}} \left. \frac{\partial \Phi}{\partial V_{\text{out}}} \right|_{I_{\text{out}}} \left. \frac{\partial V_{\text{out}}}{\partial I_{\text{out}}} \right|_{\Phi} = -1. \quad (\text{B1})$$

Additionally, using the relationship

$$\left. \frac{\partial V_{\text{out}}}{\partial \Phi} \right|_{\Phi} \left. \frac{\partial \Phi}{\partial V_{\text{out}}} \right|_{\Phi} = 1 \quad (\text{B2})$$

and Eqs. (2) and (3), we arrive at

$$\left. \frac{\partial I_{\text{out}}}{\partial \Phi} \right|_{V_{\text{out}}} \approx -\frac{1}{L_{sq}}. \quad (\text{B3})$$

APPENDIX C: CRYOGENIC RF PREAMPLIFIER

As an alternative to the ultra-high-speed room-temperature preamp described in Sec. IV E, we consider the use of a cryogenic RF amplifier to read out signal frequencies above 50 MHz. One example of such an RF amplifier is the CITLF2 produced by Cosmic Microwave Technology, Inc. It possesses an input impedance of 50Ω , so it is impedance matched to the 50Ω output impedance of the second-stage SQUID. The CITLF2 has a gain of 32 dB with a gain flatness of ± 1 dB and a noise temperature of better than $T_{n,\text{RF}} = 2$ K from 5 to 500 MHz. To current bias the second-stage SQUID array and to maintain the bias point with low-frequency flux feedback, the RF amplifier may be preceded by a diplexer that routes the information at and near DC to the Magnicon XXF-1 electronics.

We now calculate the values of ϵ_{uc} and η for the full SQUID chain previously shown in Figs. 3 and 4. Since the RF amplifier is operated at 4 K, it is connected to the second-stage dc SQUID through a superconducting coaxial cable, and we take $R_{\text{lead}} = 0$. We can then refer the temperature noise of the cryogenic RF amplifier $T_{n,\text{RF}}$ to a flux noise in the second-stage SQUID,

$$\Phi_{n2,\text{RF}}^2 \approx 4k_B T_{n,\text{RF}} \left(\frac{20}{N_{\text{ser}} N_{\text{par}}} \right) \left. \frac{\partial \Phi}{\partial I_{\text{out}}} \right|_{20 \times 1} \left. \frac{\partial \Phi}{\partial V_{\text{out}}} \right|_{20 \times 1}. \quad (\text{C1})$$

Then, the RF amp and second-stage SQUID noise referred to a flux in the input in the second-stage SQUID becomes

$$\Phi_{n2}^2 = \Phi_{n2,\text{SQ}}^2 + \Phi_{n2,\text{RF}}^2, \quad (\text{C2})$$

where $\Phi_{n2,\text{SQ}}$ is still from Eq. (36). Equation 35 is then used to compute $S_{II,\text{ref}(\omega)}$, and the rest of the analysis proceeds as already described in Secs. IV–VIII. In Fig. 3, we plot $\eta(\omega)$ for the full SQUID chain with the CITLF2 RF amplifier, and in Fig. 4, we plot $\epsilon_{uc}(\omega)$.

REFERENCES

- ¹R. Jaklevic, J. Lambe, A. Silver, and J. Mercereau, "Quantum interference effects in Josephson tunneling," *Phys. Rev. Lett.* **12**, 159 (1964).
- ²J. Clarke and A. I. Braginski, *The SQUID Handbook* (Wiley Online Library, 2004), Vol. 1.

- ³P. Falferi *et al.*, “ $10h$ superconducting quantum interference device amplifier for acoustic gravitational wave detectors,” *Appl. Phys. Lett.* **93**, 172506 (2008).
- ⁴T. Sleanor, E. L. Hahn, C. Hilbert, and J. Clarke, “Nuclear-spin noise and spontaneous emission,” *Phys. Rev. B* **36**, 1969 (1987).
- ⁵C. Hilbert and J. Clarke, “DC SQUIDS as radiofrequency amplifiers,” *J. Low Temp. Phys.* **61**, 263–280 (1985).
- ⁶L. Brouwer *et al.*, (DMRadio Collaboration), “Projected sensitivity of DMRadio-m³: A search for the QCD axion below $1\mu\text{eV}$,” *Phys. Rev. D* **106**, 103008 (2022).
- ⁷A. AlShirawi, C. Bartram, J. Benabou, L. Brouwer, S. Chaudhuri, H.-M. Cho, J. Corbin, W. Craddock, A. Droster, J. Foster *et al.*, “Electromagnetic modeling and science reach of DMRadio-m³,” [arXiv:2302.14084](https://arxiv.org/abs/2302.14084) (2023).
- ⁸N. Crisosto, P. Sikivie, N. S. Sullivan, D. B. Tanner, J. Yang, and G. Rybka, “ADMX SLIC: Results from a superconducting LC circuit investigating cold axions,” *Phys. Rev. Lett.* **124**, 241101 (2020).
- ⁹A. V. Gramolin *et al.*, “Search for axion-like dark matter with ferromagnets,” *Nat. Phys.* **17**, 79–84 (2021).
- ¹⁰J. A. Devlin *et al.*, “Constraints on the coupling between axionlike dark matter and photons using an antiproton superconducting tuned detection circuit in a cryogenic penning trap,” *Phys. Rev. Lett.* **126**, 041301 (2021).
- ¹¹S. Chakrabarty, J. R. Gleason, Y. Han, A. T. Hipp, M. Solano, P. Sikivie, N. S. Sullivan, D. B. Tanner, M. Goryachev, E. Hartman, B. T. McAllister, A. Quiskamp, C. Thomson, M. E. Tobar, M. H. Awida, A. S. Chou, M. Hollister, S. Knirck, A. Sonnenschein, W. Wester, T. Braine, M. Guzzetti, C. Hanretty, G. Leum, L. J. Rosenberg, G. Rybka, J. Sinnis, J. Clarke, I. Siddiqi, R. Khatiwada, G. Carosi, N. Du, N. Robertson, L. D. Duffy, C. Boutan, N. S. Oblath, M. S. Taubman, J. Yang, E. Lentz, E. J. Daw, M. G. Perry, C. Bartram, J. H. Buckley, C. Gaikwad, J. Hoffman, K. W. Murch, and T. Nitta, “Low frequency (100–600 MHz) searches with axion cavity haloscopes,” [arXiv:2303.07116](https://arxiv.org/abs/2303.07116) [hep-ph] (2023).
- ¹²C. D. Tesche and J. Clarke, “dc SQUID: Noise and optimization,” *J. Low Temp. Phys.* **29**, 301–331 (1977).
- ¹³C. D. Tesche and J. Clarke, “dc SQUID: Current noise,” *J. Low Temp. Phys.* **37**, 397–403 (1979).
- ¹⁴J. Clarke, C. D. Tesche, and R. Giffard, “Optimization of dc SQUID voltmeter and magnetometer circuits,” *J. Low Temp. Phys.* **37**, 405–420 (1979).
- ¹⁵F. C. Wellstood, C. Urbina, and J. Clarke, “Low-frequency noise in dc superconducting quantum interference devices below 1 K,” *Appl. Phys. Lett.* **50**, 772–774 (1987).
- ¹⁶F. Wellstood, “Excess noise in the dc SQUID (Superconducting QUantum Interference Device); 4. 2 K to 20 mK,” Tech. Rep. (Lawrence Berkeley Lab., CA (USA), 1988).
- ¹⁷S. Anton, J. Birenbaum, S. O’Kelley, V. Bolkhovsky, D. A. Braje, G. Fitch, M. Neeley, G. Hilton, H.-M. Cho, K. Irwin *et al.*, “Magnetic flux noise in dc SQUIDS: Temperature and geometry dependence,” *Phys. Rev. Lett.* **110**, 147002 (2013).
- ¹⁸This work considers the dc SQUID used as an “op-amp mode” amplifier, measuring current. This is in contrast to a “scattering mode” or RF amplifier, which measures a transmitted wave (see Ref. 27 for the theory or, as an example of a scattering-mode amplifier, the microstrip SQUID amplifiers used in ADMX⁴⁶). For an op-amp mode current amplifier such as the dc SQUID, all noise sources can be referred to imprecision input current noise and backaction input voltage noise terms as well as their correlations²⁷.
- ¹⁹Here, we actually assume $\Gamma < 0.025$ instead of the nominal 0.05 from Eq. (1). If the nominal value is used, a slightly lower value of $(\partial V_{\text{out}}/\partial \Phi)|_{I_{\text{out}}}$ is attained, as in, e.g., Ref. 20. Reference 14 uses the same value as Eq. (2). See Appendix A for further discussion.
- ²⁰J. J. P. Bruines, V. J. de Waal, and J. E. Mooij, “Comment on: “dc SQUID: Noise and optimization” by tesche and clarke,” *J. Low Temp. Phys.* **46**, 383–386 (1982).
- ²¹For very high bias currents, $I \gg I_0$, the dynamic resistance is simply the parallel resistance of the two shunt resistors, $R_{\text{dyn}} = R_j/2$. Some sources erroneously assume that this is also the dynamic resistance at optimal TC bias. However, at optimal TC bias, the dynamic resistance is approximately twice this value. See Appendix A for details.
- ²²M. Kiviranta and H. Seppä, “Effect of voltage bias on the dc SQUID characteristics,” *J. Low Temp. Phys.* **123**, 127–136 (2001).
- ²³Z. Ahmed, M. Amiri, S. Benton, J. Bock, R. Bowens-Rubin, I. Buder, E. Bullock, J. Connors, J. Filippini, J. Grayson *et al.*, “BICEP3: A 95 GHz refracting telescope for degree-scale CMB polarization,” in *Millimeter, Submillimeter, and Far-Infrared Detectors and Instrumentation for Astronomy VII* (SPIE, 2014), Vol. 9153, pp. 540–551.
- ²⁴W. B. Doriese, K. M. Morgan, D. A. Bennett, E. V. Denison, C. P. Fitzgerald, J. W. Fowler, J. D. Gard, J. P. Hays-Wehle, G. C. Hilton, K. D. Irwin *et al.*, “Developments in time-division multiplexing of X-ray transition-edge sensors,” *J. Low Temp. Phys.* **184**, 389–395 (2016).
- ²⁵W. S. Holland, W. Duncan, D. B. Kelly, K. D. Irwin, A. J. Walton, P. A. R. Ade, and E. I. Robson, “SCUBA-2: A large format submillimetre camera on the James Clerk Maxwell Telescope,” in *Millimeter and Submillimeter Detectors for Astronomy* edited by T. G. Phillips and J. Zmuidzinas (International Society for Optics and Photonics, 2003), Vol. 4855.
- ²⁶J. Clarke, “Principles and applications of SQUIDS,” *Proc. IEEE* **77**, 1208–1223 (1989).
- ²⁷A. A. Clerk, M. H. Devoret, S. M. Girvin, F. Marquardt, and R. J. Schoelkopf, “Introduction to quantum noise, measurement, and amplification,” *Rev. Mod. Phys.* **82**, 1155–1208 (2010).
- ²⁸C. Hilbert and J. Clarke, “Measurements of the dynamic input impedance of a dc SQUID,” *J. Low Temp. Phys.* **61**, 237–262 (1985).
- ²⁹E. van Assendelft, H.-M. Cho, J. Corbin, S. Kuentsner, D. Li, A. Phipps, N. Rapis, J. Singh, K. Wells, and K. Irwin, “Measurements of dc SQUID damping effects on superconducting resonant circuits,” *IEEE Trans. Appl. Supercond.* **33**, 1–6 (2023).
- ³⁰C. Tesche, “Optimization of dc SQUID linear amplifiers and the quantum noise limit,” *Appl. Phys. Lett.* **41**, 490–492 (1982).
- ³¹C. Tesche, “Analysis of strong inductive coupling on SQUID systems,” *IEEE Trans. Magn.* **19**, 458–460 (1983).
- ³²J. M. Martinis and J. Clarke, “Signal and noise theory for a dc SQUID amplifier,” *J. Low Temp. Phys.* **61**, 227–236 (1985).
- ³³K. R. Carroll, “Small-signal analysis for dc SQUID amplifiers,” *J. Low Temp. Phys.* **82**, 257–277 (1991).
- ³⁴R. H. Koch, D. Van Harlingen, and J. Clarke, “Quantum noise theory for the dc SQUID,” *Appl. Phys. Lett.* **38**, 380–382 (1981).
- ³⁵L. Levitin, R. Bennett, A. Casey, B. Cowan, C. Lusher, J. Saunders, D. Drung, and T. Schurig, “A nuclear magnetic resonance spectrometer for operation around 1 MHz with a sub-10-mK noise temperature, based on a two-stage dc superconducting quantum interference device sensor,” *Appl. Phys. Lett.* **91**, 262507 (2007).
- ³⁶C. Boutan, M. Jones, B. H. LaRoque, N. S. Oblath, R. Cervantes, N. Du, N. Force, S. Kimes, R. Ottens, L. J. Rosenberg, G. Rybka, J. Yang, G. Carosi, N. Woollett, D. Bowring, A. S. Chou, R. Khatiwada, A. Sonnenschein, W. Wester, R. Bradley, E. J. Daw, A. Agrawal, A. V. Dixit, J. Clarke, S. R. O’Kelley, N. Crisosto, J. R. Gleason, S. Jois, P. Sikivie, I. Stern, N. S. Sullivan, D. B. Tanner, P. M. Harrington, and E. Lentz, (ADMX Collaboration), “Piezoelectrically tuned multimode cavity search for axion dark matter,” *Phys. Rev. Lett.* **121**, 261302 (2018).
- ³⁷C. Hilbert and J. Clarke, “DC SQUIDS as radiofrequency amplifiers,” *J. Low Temp. Phys.* **61**, 263–280 (1985).
- ³⁸W. R. Inc., see URL: <http://www.wrcad.com/wrspice.html> for “Wrspace.”
- ³⁹Magnicon GbR, see URL: <http://www.magnicon.com/squid-electronics/xxf-1> for “Manual of XXF-1 electronics.”
- ⁴⁰D. Drung, “High- T_c and low- T_c dc SQUID electronics,” *Supercond. Sci. Technol.* **16**, 1320–1336 (2003).
- ⁴¹D. Drung, C. Assmann, J. Beyer, M. Peters, F. Ruede, and T. Schurig, “dc SQUID readout electronics with up to 100 MHz closed-loop bandwidth,” *IEEE Trans. Appl. Supercond.* **15**, 777–780 (2005).
- ⁴²Cosmic Microwave Technology, Inc., see URL: <https://www.cosmicmicrowavetechnology.com/products/citlf2> for “CITLF2 cryogenic SiGe low noise amplifier.”
- ⁴³S. W. Henderson, J. R. Stevens, M. Amiri, J. Austermann, J. A. Beall, S. Chaudhuri, H.-M. Cho, S. K. Choi, N. F. Cothard, K. T. Crowley, S. M. Duff,

C. P. Fitzgerald, P. A. Gallardo, M. Halpern, M. Hasselfield, G. Hilton, S.-P. P. Ho, J. Hubmayr, K. D. Irwin, B. J. Koopman, D. Li, Y. Li, J. McMahon, F. Nati, M. Niemack, C. D. Reintsema, M. Salatino, A. Schillaci, B. L. Schmitt, S. M. Simon, S. T. Staggs, E. M. Vavagiakis, and J. T. Ward, “Readout of two-kilopixel transition-edge sensor arrays for advanced ACTPol,” in *Millimeter, Submillimeter, and Far-Infrared Detectors and Instrumentation for Astronomy VIII*, International Society for Optics and Photonics Vol. 9914, edited by W. S. Holland and J. Zmuidzinas (SPIE, 2016), p. 99141G.

⁴⁴S. Chaudhuri, K. Irwin, P. W. Graham, and J. Mardon, “Optimal impedance matching and quantum limits of electromagnetic axion and hidden-photon dark matter searches,” [arXiv:1803.01627](https://arxiv.org/abs/1803.01627) (2018).

⁴⁵A. Rohatgi, see URL: <https://automeris.io/WebPlotDigitizer> for “Webplot digitizer: Version 4.6” (2022).

⁴⁶S. R. O’Kelley, G. Hilton, and J. Clarke, “The microstrip SQUID amplifier in ADMX,” in *Microwave Cavities and Detectors for Axion Research: Proceedings of the 3rd International Workshop* (Springer, 2020), pp. 23–36.