



Darren Hart

Sandia National Laboratories



**Infrasound Sensor Evaluations Performed
at the
Facility for Acceptance, Calibration and Testing
(FACT) site**

June 15, 2012

CEA - France



- Infrasound Minimum Station Requirements
- Sensors
- Analysis and Evaluation Summary
- New sensor designs
- Discussion



1.2. Minimum Requirements for Infrasound Station Specifications

Characteristics	Minimum Requirements
Sensor type	Microbarometer
Number of sensors	Four element array ^a
Geometry	Triangle with a component at the centre
Spacing	Triangle basis: 1 to 3 km ^b
Station location accuracy	≤100 m
Relative sensor location	≤1 m
Measured parameter	Absolute ^c or differential pressure
Passband	0.02 to 4 Hz
Sensor response	Flat to pressure over the passband
Sensor noise	≤18 dB below minimum acoustic noise ^d
Calibration	≤5% in absolute amplitude ^e
State of health	Status data transmitted to the International Data Centre
Sampling rate	≥10 samples per second
Resolution	≥1 count per 1 mPa
Dynamic range	≥108 dB
Timing accuracy	≤1 ms ^f
Standard temperature range	-10°C to +45°C ^g
Buffer at the station or National Data Centre	≥7 days
Data format	Group of Scientific Experts format
Data frame length	≤30 s
Data transmission	Continuous
Data availability	≥98%
Timely data availability	≥97%
Mission capable array	≥3 elements operational
Acoustic filtering	Noise reduction pipes (site dependent)
Auxiliary data	Meteorological data ^h

^a In the case of noisy sites or when increased capability is required, the number of components could be increased.

^b 3 km is the recommended spacing.

^c Used for daily state of health.

^d Minimum noise level at 1 Hz: ~5 mPa.

^e Periodicity: once per year (minimum).

^f Better than or equal to 1 ms.

^g Temperature range to be adapted for some specific sites.

^h Once per minute.



Summary of IMS requirements:

- 1. Passband 0.02 – 4 Hz**
 - (flat over passband +/- ?)
- 2. Sensor (System) Noise \leq 18 dB below minimum noise model**
 - d Minimum sensor noise at 1 Hz is 5 mPa (-46 dB)
 - critical review is needed for any noise model produced
- 3. Calibration 5% absolute amplitude**
- 4. Resolution \geq 1 count per mPa**
 - Review current definition of CALIB ()
- 5. Dynamic Range \geq 108 dB**
 - RMS of Full-Scale Output Pa for tonal signal
 - Noise is estimated using RMS of passband
 - $20 \times \log_{10} (\text{FS/Noise})$



Martec RM2000/MB2000/MB2005

Power: MB2000 =4 watts and MB2005 =1.6 watts @ 12V

Sensitivity: MB2005 96.3 mV/Pa

Noise: -64 dB rel 1 Pa²/Hz ~ 0.7 mPa rms (0.5-2 Hz)

Full-scale Pressure: 107 Pa (zero to peak)

Dynamic Range: 104 dB

Passband:0.02 – 10 Hz

Seismic coupling: Yes



Observations for MB2005:

- TP9 output to re-center LVDT
- Reduced power consumption
- Differential outputs
- Increased input voltage range (9-36V)

General Comments


- Is the seismic sensitivity too large?




Sandia National Laboratories

Martec MB2000 1380 and MB2005 NV7009 Isolation Self-Noise (static)

DC Removal: **BLOCK** Window: **HANN** FFT Length: **16K** FFT Overlap: **5/8** 90% Confidence: **0.94532 dB** Unit: **Pressure**



2-Channel Coherence – Distributed model Self-Noise (dynamic)



Sandia National Laboratories

Martec MB2000 1380 and MB2005 NV7009 Isolation Self-Noise (static)

DC Removal: Window: FFT Length: FFT Overlap: 90% Confidence: 0.94532 dB Unit: Pressure







Figure taken from: Infrasound Monitoring for Atmospheric Studies, p.130

Fig. 4.10 Comparison between MB2007 MB2005 self-noise (Ponceau et al. 2008) and Infrasound Low Noise Model from Bowman (Bowman et al. 2007)



Sandia National Laboratories


Piston-Phone Linearity / Amplitude Response Verification




6 amplitudes at single frequency (1Hz); Sine fit in green; misfit in red



Linear voltage output with increasing amplitude




Reduced SRN with increasing amplitude




Sandia National Laboratories

CEA Visit

Closer look at 3.56 Pa tone...



Station: s1004
Channel: MB2000 1380
Rate: 50 Hz
Bitweight: 3.27 uV/count
Response: MB2000-SN1380...
Unit: Pressure



Station: s1004
Channel: MB2005 NV7009
Rate: 50 Hz
Bitweight: 3.27 uV/count



The residual has no apparent coherent structure.



Sandia National Laboratories

Chaparral 50A (102220 and 102221)

Power: CP50A = 0.35 watts @ 12V

Sensitivity: 0.420 mV/Pa

Noise: -81 dB rel 1 Pa²/Hz ~ 80 uPa rms (0.5-2 Hz)

Full-scale Pressure: 44.7 Pa (zero to peak)

Dynamic Range: 112 dB

Passband: 0.01 – 50 Hz

Observations:


- DC offset increased with signal amplitude, a check of the balance of differential outputs is warranted
- Testing was not conducted on temperature sensitivity
- Datasheets imply traceability to LANL infrasound calibration capability
- Decreased tonal SNR for amplitude linearity Test
- Self-Noise varies by analysis technique used




Sandia National Laboratories

Chaparral 50A (102220 and 102221) Isolation Self-Noise (static)

DC Removal: Window: FFT Length: FFT Overlap: 90% Confidence: Unit:




2-Channel Coherence – Distributed model Self-Noise (dynamic)




Sandia National Laboratories

CEA Visit


Piston-Phone Linearity / Amplitude Response Verification



Station: S1036
Channel: MB2000 1380
Rate: 100 Hz
Bitweight: 3.2591 uV/count
Response: MB2000-SN1380...
Unit: Pressure



Station: S1036
Channel: CP 50A 20
Rate: 100 Hz
Bitweight: 3.25492 uV/count




Station: S1036
Channel: CP 50A 21
Rate: 100 Hz
Bitweight: 3.25482 uV/count

15 amplitudes at single frequency (1Hz); Sine fit in green; misfit in red




Sandia National Laboratories


Linear voltage output with increasing amplitude



Reduced SRN with increasing amplitude



Closer look at 3.34 Pa tone...




The residual has twice the frequency as the test tone.
Harmonic distortion...




Sandia National Laboratories

CEA Visit



Sensor:
Low Gain: 0.02 V/Pa
High Gain: 0.1 V/Pa
Output(V): +/- 10

13 dB difference in Voltage in noise.



When converted by response to Pa, the Pa noise is the same.

Waveform	Noise (mPa rms)	Full-Scale (Pa)	DR 0.02-4 Hz	Noise (mPa rms)	Full-Scale (Pa)	DR 0.5-2 Hz
MB2005 HG	2.60	100	89.3 dB	1.16	100	96.0 dB
MB2000 LG	2.90	500	102.9 dB	1.13	500	110.0 dB





Sandia National Laboratories

CEA Visit

Sensor:
Low Gain: 0.4 V/Pa
High Gain: 2.0 V/Pa
Output(V): +/- 18

20 dB difference in Voltage in noise.

When converted by response to Pa, the Pa noise is 5 dB higher for HG.



Waveform	Noise (mPa rms)	Full-Scale (Pa)	DR 0.02-4 Hz	Noise (mPa rms)	Full-Scale (Pa)	DR 0.5-2 Hz
IS HG	1.37	9	73.3 dB	0.72	9	78.9 dB
IS LG	0.63	45	94.1 dB	0.29	45	100.1 dB




Sandia National Laboratories

CEA Visit

Sensors

- Chaparral 2.2, 2.5, 25
- IML SS and ST
- Miltec
- NCPA (IRIS-PASSCAL)
- infraNMT
- PCB Piezotronics



Sandia National Laboratories

National Center for Physical Acoustics (NCPA-IRIS)

Power: 12 mwatts @ +/- 4.5V (9V)

Sensitivity: 23 mV/Pa

Noise: -86 dB rel 1 Pa^2/Hz ~ 60 uPa rms (0.5-2 Hz)

Full-scale Pressure: 160 Pa (zero to peak)

Dynamic Range: 125 dB

Passband: 0.007 – 40 Hz (acoustic inlet controlled)



Observations:

- Testing was not conducted temperature sensitivity
- Sensor design requires differential power (-4.5, gnd, +4.5)
- Very small seismic sensitivity (only observed during isolation noise test)
- Possible issue with sensitivity change with shipping method (sealed back volume)



Sandia National Laboratories

National Center for Physical Acoustics (NCPA) Noise and Tonal SNR



Sandia National Laboratories

Discussion:

- Temperature Testing specifications
- Multiple Facilities for sensor evaluation tasks
- Stability of sensor specifications
- Multiple sensors at stations
- Incorporate MET data into analysis



Sandia National Laboratories