

Thermal Battery Tester Uncertainty and Mistakeproofing Project – An Update

Joe Garni
Senior Member of Technical Staff
Sandia National Laboratories
Power Sources Production Department

44th Power Sources Conference
Las Vegas, NV
June 14, 2010

Sandia National Laboratories is a multi-program laboratory operated by Sandia Corporation, a wholly owned subsidiary of Lockheed Martin Company, for the U. S. Department of Energy's National Nuclear Security Administration under contract DE-AC04-94AL85000.




Major Project Objectives

- Reduce the number of bad tests caused by equipment failure, setup errors, etc. on destructive thermal battery tester
 - Thermal batteries one-time use only; once test is started, battery is destroyed, costs > \$5000
- Develop an accurate standard for tester verification
 - Check out tester to verify that all equipment/cabling is properly configured
 - Collect, analyze data over time, determine whether tester is drifting. May lengthen intervals between calibrations, cut down on travel time/costs.
 - Right now, calibration interval is three months.



Major Project Objectives

- **Quantify an uncertainty number for the tester**
 - Recent battery developments have required very tight tolerance bands
 - Establish an error band around each reading





Need for Mistakeproofing

- Each battery requires re-routing of cables and/or special cable/box setup
- Improperly routed cables
- Sometimes cables have invisible loose or broken connections
- A test not properly performed destroys a battery that costs thousands of dollars to build with no good data to show for it.
- Reduce the risk of these “no-tests”



Check Standard

- Mimics performance of a thermal battery in terms of voltage rise and response to varied load levels.
- Will be used in a simulated test once each day during lot acceptance testing. Serves several purposes:
 - Validates tester and cabling/load box setup
 - Tester alerts operator if problem is found
 - Tester gathers data set for PMAP analysis
 - Data used to detect drift trend
 - Drift can indicate need for calibration; lack of drift may indicate no need to calibrate
 - Obviously out-of-tolerance data (0.04V when should be 10V) could indicate improper equipment configuration
 - Gathers its own data set for comparison to tester data set



Check Standard – Final Design

- About the length, width of a briefcase, but greater height
- Integrated Reference Standard (10V, 5V, 3V, 2.048V Outputs)
- Storage for battery and charger





Check Standard Enclosure

- Weighs 10.26kg (22.5 lbs)





Check Standard – Final Design


- Voltage Output Ports
- Voltage Reference Standard Outputs
- Data Drive Port (Flash Drive)
- Test Parameter Input Ports (Future)





Check Standard – Final Design

- Shown with custom cable for attachment to tester equipment
- One-step attachment - easier than running ten banana leads!






Check Standard Temperature Testing

- Measure how temperature affects performance
 - Test area at production facility is known to experience seasonal temperature variations
- Placed check standard in temperature chamber
- Connected four calibrated meters and one voltage calibrator (for activation signal)
- Set chamber to temperature (10C first day, 22C second day, 35C third day)
 - Simulate max and min in environmental test area
- Four test runs approx. 1 hour apart, then one run ten minutes after the fourth run.
- Manually recorded voltage on each channel 5 seconds after activation
- Compared against check standard's data reading at 5 sec. after activation

Temperature Test Results – Channel 1

Standard vs. Meter, 19.8 V Tap, (3/3 = 10C, 3/4 = 22C, 3/5 = 35C)





“Cooldown” Testing

- During temperature testing, noticed that readings were elevated when a test was run 10 minutes after the previous one (or if standard was on for several minutes before activation)
- Problematic because we want the standard to show as little variability as possible from run to run (within 2 mV)
- Needed to find out shortest interval between activations that would consistently show readings within 2 mV at 5 sec. after activation




“Cooldown” Testing

- Used the same meter/calibrator setup as with temperature testing
- Took readings at 5 sec. after activation
- Started with 10-minute cooldown
 - 5 minutes from activation to end of data file flash drive transfer plus 10 minutes until next activation
- Widened cooldown interval to 15, 20, then 30 minutes


10-minute Cooldown Interval


Check Standard Cooldown Test Results, Channel 1, 15 minutes
between test starts (10-minute cooldown)




30-minute Cooldown Interval

Check Standard Cooldown Test Results, Channel 1, 35 minutes
between test starts (30-minute cooldown)






How the Check Standard Works with the Tester



Sample Tester/Standard Overlay Graph


- Shows load profile for each channel
 - 0.5 Amp and 3.0 Amp pulse on each of 4 channels, 250ms wide
 - Each channel has a different constant-resistive background load
 - Accounts for differences in background loads

Tester vs. Standard, June 3, 2010



Tester/Standard Data Overlay – Single Channel Close-up

Tester vs. Standard, June 3, 2010





Data Analysis

- Looking for drift trends
 - Trend line may indicate need for calibration
- Looking for out-of-tolerance conditions
 - May indicate tester malfunction or improper equipment configuration
- User sets up analysis criteria sets
 - Can analyze minimum, maximum, standard deviation, or average in voltage or current data between two time points of interest
 - Can also analyze voltage or current at one time point
- Analysis criteria sets and results are stored in Access database
 - Code within Access performs data analysis

Example of Data Analysis Criteria Set

EDIT DATA COLLECTION PARAMETERS

Data Collection Parameter Set Name: 4-Channel Reference Standard

Hardware/Cal Record: PT3660 004

SAVE

RETURN


RETURN TO MAIN MENU

Channel #	Volts or Current	Statistic	Start Time (s)	Stop Time (s)	Lower Spec Limit	Upper Spec Lim
1	V	Average	0.00	1.00	9.90	10.1
1	V	Maximum	0.00	1.00		10.1
1	V	Minimum	0.00	1.00	9.90	
1	V	Standard Deviation	0.00	1.00		
1	V	Value	1.00	1.00	9.90	10.1
2	V	Average	0.00	1.00	4.90	5.1
2	V	Maximum	0.00	1.00		5.1
2	V	Minimum	0.00	1.00	4.90	
2	V	Standard Deviation	0.00	1.00		

Record: 1 of 20


Drift Trend Analysis (Tester Data)

- Taken recently at Sandia test facility



Drift Trend Analysis (Check Standard)

- Same set of test runs as shown in tester data on previous slides
- Much less scatter than tester data





Remaining Tasks

- **Uncertainty Determination**
 - **Automated Operational Uncertainty Software is under development**
 - Takes voltage inputs from reference standards on-board the check standard
 - Also takes measurements on highly-accurate meter
 - Gather a set of data from tester and meter every calibration
 - Calculation methodology has yet to be determined
 - RSS calculation with standard deviations of tester and meter data
 - Possibly bias factor (account for difference between tester and meter)
 - Combine each data set for tester and meter into one long-term set for each, then run RSS?
 - Take average of each data set's standard deviation over time, then run RSS?



Remaining Tasks

- Collect more data at production facility where most acceptance testing takes place
 - Use for development of expanded data analysis parameter sets
 - May help to extend interval between tester calibrations
 - Right now is every 3 months
 - No MTBF analysis, as cost of rejected lot would be too great
 - If we could extend calibration intervals, we could cut travel costs and time away from home
- Develop check standard calibration procedure and tester operating procedure
- Update software document
- Qualify tester software before next production lot (2011)