

LA-UR-16-22546

Approved for public release; distribution is unlimited.

Title: Uranium Detection - Technique Validation Report

Author(s): May, Iain
Lujan, Elmer J. W
Reilly, Sean Douglas
Colletti, Lisa Michelle
Mechler-Hickson, Alexandra Marie
Garduno, Katherine

Intended for: Open Source Milestone Report and Presentation for NNSA

Issued: 2016-04-14

Disclaimer:

Los Alamos National Laboratory, an affirmative action/equal opportunity employer, is operated by the Los Alamos National Security, LLC for the National Nuclear Security Administration of the U.S. Department of Energy under contract DE-AC52-06NA25396. By approving this article, the publisher recognizes that the U.S. Government retains nonexclusive, royalty-free license to publish or reproduce the published form of this contribution, or to allow others to do so, for U.S. Government purposes. Los Alamos National Laboratory requests that the publisher identify this article as work performed under the auspices of the U.S. Department of Energy. Los Alamos National Laboratory strongly supports academic freedom and a researcher's right to publish; as an institution, however, the Laboratory does not endorse the viewpoint of a publication or guarantee its technical correctness.

Uranium Detection - Technique Validation Report

L.M. Colletti,¹ K. Garduno,¹ E.J.W. Lujan,¹ A.M. Mechler-Hickson,^{1,2} I. May,¹ & S.D. Reilly¹

¹Chemistry Division, Los Alamos National Laboratory

²University of Wisconsin, Madison, Wisconsin

As a LANL activity for DOE/NNSA in support of SHINE Medical Technologies™ ‘Accelerator Technology’ we have been investigating the application of UV-vis spectroscopy for uranium analysis in solution. While the technique has been developed specifically for sulfate solutions, the proposed SHINE target solutions, it can be adapted to a range of different solution matrixes. The FY15 work scope incorporated technical development that would improve accuracy, specificity, linearity & range, precision & ruggedness, and comparative analysis. Significant progress was achieved throughout FY 15 addressing these technical challenges, as will be summarized in this report. In addition, comparative analysis of unknown samples using the Davies-Gray titration technique highlighted the importance of controlling temperature during analysis (impacting both technique accuracy and linearity/range). To fully understand the impact of temperature additional experimentation and data analyses were performed during FY16. The results from this FY15/FY16 work were presented in a detailed presentation, LA-UR-16-21310, and an update of this presentation is included with this short report summarizing the key findings.

The technique is based on analysis of the most intense U(VI) absorbance band in the visible region of the uranium spectra in 1 M H₂SO₄, at $\lambda_{\text{max}} = 419.5$ nm. During sample analysis aliquots of uranium solutions of unknown concentration would be diluted into 1 M H₂SO₄ prior to analysis (an accurately known *ca.* 40× dilution), both to allow ‘in range’ measurement and to maintain specific solution chemistry environment. As the intensity of this transition varies linearly with uranium concentration then the Beer-Lambert law allows concentration to be measured through the following equation: -

$$A = \varepsilon c l$$

where A = absorbance, ε = molar absorptivity (M⁻¹ cm⁻¹), c = concentration (M) & l = cell path length (cm)

In an effort to improve technique accuracy almost all sample concentration analysis was performed through mass measurements (*e.g.* mg/g, wt %), with accurate density measurements used to convert to, and from, molar concentrations at a measured temperature. Thus, “1 M” H₂SO₄ is actually 9.26 wt % H₂SO₄ in H₂O. In addition, a new “reads” data collection method (*vs.* a traditional scan) was developed to reduce uncertainties associated with absorbance values only at wavelengths of most interest; λ_{max} at 419.5 nm, the two adjacent peaks at 409 & 430 nm, baseline wavelengths values at 550 & 600 nm and an additional measurement at 500 nm.

Determination of what we predicted would be more accurate ε values was undertaken by dissolving uranium metal standard material (CRM 112-A) in nitric acid and converting to uranium standard solutions of various concentrations in “1 M” H₂SO₄. In addition, uranium samples were prepared in “0.9

M" (8.348 wt %) and "1.1 M" (10.133 wt %) H_2SO_4 , and analyzed by UV-vis to confirm technique range ("1.0 \pm 0.1 M" H_2SO_4). Initially these samples were measured using triplicate 10 mm path length cells from 3 different companies (Starna cells, Hellma Analytics & NSG precision) and triplicate 2 mm & 100 mm path length cells (Starna cells). The 10 mm Starna cells were found to be the most 'user friendly,' although all three manufacturers 10 mm path length cells worked well, and provided evidence of the impact of temperature on calculated molar absorptivity (see subsequent discussion). Increased uncertainties associated with the shortened path length precluded further studies with the 2mm path length cells. Technical difficulties, and standard solution concentrations largely incompatible with the path length, prohibited an effective assessment of the value of the longest path length/lowest associated uncertainties 100 mm cell. For all subsequent analysis only the three 10 mm Starna Cells were used.

The next stage of technique development was analysis of 5 actual uranium sulfate samples, 2 of which were very well characterized. Initial comparison of the results obtained from uranium analysis of these samples, both by the UV-vis technique and Davies-Gray titration, indicated that the UV-vis technique systematically underestimated uranium concentration. Temperature fluctuations in the laboratory that correlated with observed ϵ (increased with temp.) were determined to be a major issue and necessitated additional experimental standard solution measurements to facilitate the determination of ϵ at lower temps. This allowed for the determination of ϵ as function of temperature over the range in which assays of the 5 actual samples were measured. Additional discrepancies between analysis techniques could be attributed to assay sample volume and no quantification of sample evaporation over time. Nevertheless, of the one sample that was analyzed by three different operators in triplicate using three different UV-vis cells (27 separate analyses), all apart from one yielded measured uranium concentrations within 0.6 % of the value measured by Davies-Gray. This data indicated that the technique now has high precision and ruggedness, with the comparative analysis revealing that additional temperature control (and quantification of sample evaporation) should lead to greater accuracy.

Any contaminant that absorbs light between 419-600 nm can potentially interfere with the developed UV-vis technique. Previously we have screened a suite of potential corrosion and fission product chemical contaminants and observed that Ru had the most intense transitions in the visible region, and that peak intensities could change as a function of time. Using 129 mg/g uranium sulfate solutions doped with 0, 2.1, 21 & 210 ppm Ru, the impact of the contaminant on the ratios of the 409 nm 430 nm & 500 nm absorbance values vs. the 419.5 nm λ_{max} peak were probed. It was observed that the impact of Ru contamination could be observed through variation in these ratios at lower Ru concentrations than could be observed in measureable change in 'observed' uranium concentration. Thus, at the current level of technique accuracy, it should be possible to develop a screening protocol that would reject data in which impactful levels of contaminants were present.

Details of the work are described in the following viewgraphs.

L.M. Colletti, K. Garduno, E.J. Lujan, A.
Mechler-Hickson, I. May & S.D. Reilly
(*R. Copping & D. Rios*)
Chemistry Division



Uranium Detection

Technique Validation

Meeting at SHINE Medical Technologies

3rd March 2016


UNCLASSIFIED

Slide 1


Previous Studies

A Visible Spectroscopy Technique for Uranium Analysis

- Uranyl absorption spectroscopy – can be applied to uranium concentration measurement in solution
- $A = \epsilon cl$
 - A= absorbance
 - ϵ = molar absorptivity ($M^{-1} cm^{-1}$)
 - c= concentration (M)
 - l= path length, cm
- λ_{max} (peak max, nm) and ϵ (molar absorptivity) vary with chemical composition
- Previous studies - a small aliquot of sample (e.g. 50 μ L) diluted into excess of 1 M HNO_3 or H_2SO_4 (2000 μ L)



FY14 Work - Molar Absorptivity of Uranium(VI) Determined from Matrix of Standard Uranium Solutions




- The molar absorptivity of uranium(VI) in 1.0 ± 0.1 M H_2SO_4 at 19.5 ± 1.7 °C is 13.736 ± 0.026 $\text{cm}^{-1} \text{M}^{-1}$ at 419.5 nm.

UNCLASSIFIED

FY14 Work - Assay Method Accuracy Testing (151.2 gU/L, 0.6353 mol/L)



- The white lines represent 1 standard deviation of the known molar uranium concentration based on gravimetric data from solution preparation.
- The square and circle points are the uranium concentrations measured by the spectroscopy assay method using 90 (red) or 50 (green) μL uranium aliquots, respectively.
- The error bars are the standard deviations in these measurements.
- Difference between known and measured values all $< 0.7\%$. For the 90 μL assays.



UNCLASSIFIED

FY 14 Work

Density Measurements on pH1 Uranium Sulfate Solutions





Note: This is one of the last slides where [U] conc. is routinely discussed in terms of volume of solution. From now on all [U] discussed in terms of mass of solution.

UNCLASSIFIED

FY14 Work

Analysis of Contaminants and Impurities





Raman Spectroscopy

UNCLASSIFIED

FY14 Work

Additional Analysis of Contaminants (1 M H_2SO_4)



UNCLASSIFIED

FY 14 Work

Guide to Uncertainty Measurements (GUM)

- End of FY14 technique relative uncertainty – “0.5” %
 - 62 % of uncertainty – molar absorptivity value
 - 14 % of uncertainty – cell path length
 - 8 % of uncertainty - spectrometer signal measurement

UNCLASSIFIED



Operated by Los Alamos National Security, LLC for the U.S. Department of Energy's NNSA

Slide 8



Technique Development FY15-16

A Summary of Progress

- **Accuracy** - *reducing technique uncertainties*
 - *Improved experimental procedures implemented*
 - *A new issue revealed, temperature, and a mitigation strategy to be implemented*
- **Specificity** - *impact of sample contaminants*
 - *Issue addressed, a mitigation strategy proposed*
- **Linearity & Range** – *concentration and temperature*
 - *Addressed, a temperature effect mitigation strategy to be implemented*
- **Precision & Ruggedness** – *reproducibility*
 - *Addressed*
- **Comparative Analysis** – *Davies-Gray measurements*
 - *Experimental measurements undertaken, and will continue*

UNCLASSIFIED



Operated by Los Alamos National Security, LLC for the U.S. Department of Energy's NNSA

Slide 9



Spectrometer Signal Move from 'Scans' to 'Reads' Method

- Increase signal to noise ratio @ 419.5 nm to increase measurement accuracy
- Cary instrument software allows for rapid scan of full wave length range + enhanced signal to noise averaging @ 419.5 nm
- Example previous test measurement (FY14)
 - Normal collection parameters, baseline corrected absorbance & multiple scans (at least 20 min) – abs. 0.42136 ± 0.00030
 - Modified scan with increased signal to noise averaging @ 419.5 nm (7 min) – abs. 0.41311 ± 0.00015

UNCLASSIFIED

Reads Method of Data Collection *(for one sample, one measurement, one cell)*

Examples Background Data from Spectrophotometer

Background scan, Cell 1 , 23.6 °C
7/9/2015 1:44:46 PM
SAT (s): 5
Total Read time: 98.469 s

| Wavelength (nm) | Abs Read 1 | Abs Read 2 | Abs Read 3 |
|-----------------|------------|------------|------------|
| 600.000 | 0.006578 | 0.006551 | 0.006557 |
| 550.000 | -0.004275 | -0.004263 | -0.004217 |
| 500.000 | 0.007631 | 0.007665 | 0.007691 |
| 430.000 | 0.017881 | 0.017896 | 0.017893 |
| 419.500 | 0.016612 | 0.016608 | 0.016613 |
| 409.000 | 0.013280 | 0.013300 | 0.013316 |


UNCLASSIFIED

Reads Method of Data Collection

Examples Sample Data from Spectrophotometer

Solution 1100 , Cell 1 , 23.9 °C
7/9/2015 2:42:02 PM
SAT (s): 5
Total Read time: 98.625 s

| Wavelength (nm) | Abs Read 1 | Abs Read 2 | Abs Read 3 |
|-----------------|------------|------------|------------|
| 600.000 | 0.012529 | 0.012534 | 0.012531 |
| 550.000 | 0.001569 | 0.001576 | 0.001564 |
| 500.000 | 0.014262 | 0.014278 | 0.014281 |
| 430.000 | 0.208575 | 0.208577 | 0.208578 |
| 419.500 | 0.246289 | 0.246289 | 0.246282 |
| 409.000 | 0.215309 | 0.215311 | 0.215312 |



UNCLASSIFIED

Reads Method of Data Collecting

Data Transferred to Spreadsheet

Accuracy

| BACKGROUND READS (H_2SO_4 Bkgd reads) | | | |
|--|-------------|-------------|-------------|
| Room Temp (°C): | 23.6 | 23.6 | 23.6 |
| Wavelength | Bkgd_C1_rd1 | Bkgd_C1_rd2 | Bkgd_C1_rd3 |
| 600.0 | 0.006578 | 0.006551 | 0.006557 |
| 550.0 | -0.004275 | -0.004263 | -0.004217 |
| 500.0 | 0.007631 | 0.007665 | 0.007691 |
| 430.0 | 0.017881 | 0.017896 | 0.017893 |
| 419.5 | 0.016612 | 0.016608 | 0.016613 |
| 409.0 | 0.013280 | 0.013300 | 0.013316 |

| SAMPLE READS | | | |
|-----------------|----------------|----------------|----------------|
| Room Temp (°C): | 23.9 | 23.9 | 23.9 |
| Wavelength | Soln11A_c1_rd1 | Soln11A_c1_rd2 | Soln11A_c1_rd3 |
| 600.0 | 0.012529 | 0.012534 | 0.012531 |
| 550.0 | 0.001569 | 0.001576 | 0.001564 |
| 500.0 | 0.014262 | 0.014278 | 0.014281 |
| 430.0 | 0.208575 | 0.208577 | 0.208578 |
| 419.5 | 0.246289 | 0.246289 | 0.246282 |
| 409.0 | 0.215309 | 0.215311 | 0.215312 |

UNCLASSIFIED

Reads Method of Data Collection *Initial Averaging of Triplicate Measurements*

| Wavelength | AVERAGE BACKGROUND READS FOR EACH CELL (H ₂ SO ₄ Bkgd reads) | |
|------------|---|------------------|
| | Bkgd_C1_rd_Ave | Bkgd_C1_rd_AveSD |
| 600.0 | 0.006562 | 0.000014 |
| 550.0 | -0.004252 | 0.000031 |
| 500.0 | 0.007662 | 0.000030 |
| 430.0 | 0.017890 | 0.000008 |
| 419.5 | 0.016611 | 0.000003 |
| 409.0 | 0.013299 | 0.000018 |

| Wavelength | AVERAGE SAMPLE READS FOR EACH CELL | |
|------------|------------------------------------|------------------|
| | Soln11A_c1_Ave | Soln11A_c1_AveSD |
| 600.0 | 0.012531 | 0.000003 |
| 550.0 | 0.001570 | 0.000006 |
| 500.0 | 0.014274 | 0.000010 |
| 430.0 | 0.208577 | 0.000002 |
| 419.5 | 0.246287 | 0.000004 |
| 409.0 | 0.215311 | 0.000002 |

UNCLASSIFIED

Reads Method of Data Collection

Background Corrected Sample Reads

Accuracy

| BACKGROUND-CORRECTED SAMPLE READS FOR EACH CELL (Subtract H_2SO_4 background) | | | |
|--|--------------------|--------------------------|----------|
| Wavelength | Soln11A_c1_Ave_BGc | Soln11A_c1_Ave_BGc_Sigma | w_i |
| 600.0 | 0.005969 | 0.000014 | 4.82E+09 |
| 550.0 | 0.005821 | 0.000031 | 1.03E+09 |
| 500.0 | 0.006611 | 0.000032 | |
| 430.0 | 0.190687 | 0.000008 | |
| 419.5 | 0.229676 | 0.000005 | |
| 409.0 | 0.202012 | 0.000018 | |
| | | Wt'd Ave | Wt'd SD |
| Ave Abs (550–600 nm): | 0.0059434 | 0.0000796 | |

UNCLASSIFIED



Operated by Los Alamos National Security, LLC for the U.S. Department of Energy's NNSA

Slide 15



Reads Method of Data Collection

Final baseline correction generates abs. @ 419.5 nm

| BASELINE-CORRECTED SAMPLE READS FOR EACH CELL (subtract Ave Abs 550–600 nm) | | |
|--|---------------------------|---------------------------------|
| Wavelength | Soln11A_c1_Ave_ BGcBLc | Soln11A_c1_Ave_ BGcBLc_Sigma |
| 600.0 | -0.00005 | 0.00008 |
| 550.0 | 0.00007 | 0.00009 |
| 500.0 | 0.00174 | 0.00009 |
| 430.0 | 0.34962 | 0.00008 |
| 419.5 | 0.42315 | 0.00008 (0.02 %) |
| 409.0 | 0.37116 | 0.00009 |

UNCLASSIFIED

Reads Method of Data Collection *Comments and Evolution of Methodology*

- Weighted average of baseline values at 550 and 600 nm introduced
- ‘New’ Reads method processes each cell separately
 - Initial ‘old’ reads method averaged the data from each of the three UV-vis cells used
- New baseline reads not always recorded for each cell every day - will implement this procedure for all future measurements

UNCLASSIFIED

“1.0” M H_2SO_4
Prepared from Fisher cert. 4.00 N (2 M, 17.542 wt. %) H_2SO_4

| SOLUTION PREPARATIONS | | | |
|--|--|----------------------------------|---|
| H_2SO_4 Bottle No. | 17.542 wt % H_2SO_4 Stock Soln. Mass (g) | H_2O Mass (g) | Calc. H_2SO_4 Soln. Conc. (wt % H_2SO_4) |
| 1 | 251.690 | 224.997 | 9.262 |
| 2 | 251.770 | 225.017 | 9.263 |
| 3 | 251.820 | 225.050 | 9.263 |
| 4 | 258.285 | 231.003 | 9.260 |

- “0.9” M (8.384 wt %) and “1.1” M (10.133 wt %) H_2SO_4 solutions prepared previously (FY 14 work).
- % Difference density ((lit.-meas.)/meas). all < 0.06 %

| DENSITY (g/mL) | | |
|------------------|-------------------------------|------------------------------|
| Temp. (°C) | Meas. H_2SO_4 | Lit. H_2SO_4 |
| Bottle #1 | | |
| 20.01 | 1.06049 | 1.06093 |
| 27.5 | 1.05734 | 1.05779 |
| 35.0 | 1.05396 | 1.05437 |
| Bottle #2 | | |
| 20.01 | 1.06036 | 1.06094 |
| 27.5 | 1.05719 | 1.0578 |
| 35.0 | 1.05379 | 1.05438 |
| Bottle #3 | | |
| 20.01 | 1.06067 | 1.06094 |
| 27.49 | 1.05754 | 1.05781 |
| 35.0 | 1.05409 | 1.05438 |
| Bottle #4 | | |
| 20.01 | 1.06032 | 1.06092 |
| 27.5 | 1.05715 | 1.05778 |
| 35.0 | 1.05375 | 1.05435 |

UNCLASSIFIED



Operated by Los Alamos National Security, LLC for the U.S. Department of Energy's NNSA

Slide 18

Uranium Standard Material *Lower Associated Uncertainty*

- CRM (Certified Reference Material) 112-A uranium metal standard
- Material repackaged by NBL – previously known as SRM (Standard Reference Material) 960
- LANL analytical chemistry team dissolved two different pieces of standard metal in nitric acid to produce two solution cuts
- GUM workbench standard uncertainty values provided
 - ca. 95 % of the uncertainty associated with the purity of the metal
- Cut A
 - 12.58152 g metal dissolved in nitric acid to yield 125.0549 g soln.
 - 100.58282 ± 0.0062 mgU/gA
- Cut B
 - 12.55217 g metal dissolved in nitric acid to yield 125.0207 g soln.
 - 100.37555 ± 0.0062 mgU/gB

UNCLASSIFIED

Uranium Standard Material CRM 112-A Certificate of Analysis

 **New Brunswick Laboratory**
U.S. Department of Energy

Certificate of Analysis
CRM 112-A

Uranium (normal) Metal Assay and Isotopic Standard

| | |
|---|---------------------|
| Uranium Assay: | 0.99975 g U/g metal |
| Uranium Assay Uncertainty: | 0.00006 g U/g metal |
| $^{234}\text{U}/^{238}\text{U}$ $^{235}\text{U}/^{238}\text{U}$ | |
| Atom Ratio: | 0.000052841 |
| Atom Ratio Uncertainty: | 0.000000082 |
| ^{234}U ^{235}U ^{238}U | |
| Atom Percent: | 0.0052458 |
| Atom Percent Uncertainty: | 0.0000081 |
| Weight Percent: | |
| Weight Percent: | 0.0051579 |
| Weight Percent Uncertainty: | 0.0000080 |
| Relative Atomic Weight: | |
| Relative Atomic Weight: | 238.028918 |
| Relative Atomic Weight Uncertainty: | |
| Relative Atomic Weight Uncertainty: | 0.000012 |

http://science.energy.gov/~/media/nbl/pdf/price-lists/certificates/CRM_112A_Uranium_Metal_Sept_2010.pdf

UNCLASSIFIED

Preparation of Uranium Sulfate Standard Solutions (1-8)

- Required for re-determination of molar absorptivity values
- Eight known mass aliquots of cut A solution heated to dryness in 100 mL Pyrex volumetric flask.
- Resultant solids suspended with H₂O and again heated to dryness to drive off residual HNO₃/NO₃⁻
- Final solids, presumed to be UO₃·xH₂O, reweighed to estimate x.
- Dissolution in H₂SO₄ ("0.9", "1.0" or "1.1" M), quantitative transfer and final mass measurements (all ca. 106 g)
- ***Solution number & sample number - interchangeable***



Picture shows UO₃·xH₂O formed during preparation of solns. 11 & 12

UNCLASSIFIED

Uranium Standard Solutions (1-8)

| Soln. No. | Initial Cut A aliquot mass (g) | Solid mass after heat (mg) | H ₂ SO ₄ Conc. (wt %) | Final Solution [U] (Wt %) | x in UO ₃ ·xH ₂ O(s) |
|-----------|--------------------------------|----------------------------|---|---------------------------|--|
| 1 | 1.18753 ± 0.00010 | 151.2 | 9.263 | 0.112672 ± 0.000011 | 0.84 |
| 2 | 2.37703 ± 0.00008 | 299.5 | 9.263 | 0.225686 ± 0.000016 | 0.67 |
| 3 | 3.56385 ± 0.00010 | 449.8 | 9.263 | 0.338174 ± 0.000023 | 0.70 |
| 4 | 4.74965 ± 0.00016 | 608.1 | 9.263 | 0.450727 ± 0.000032 | 0.94 |
| 5 | 7.12300 ± 0.00007 | 893.0 | 8.384 | 0.676456 ± 0.000042 | 0.59 |
| 6 | 7.11893 ± 0.00015 | 884.4 | 9.263 | 0.675093 ± 0.000044 | 0.44 |
| 7 | 7.11870 ± 0.00018 | 886.8 | 10.133 | 0.675200 ± 0.000045 | 0.49 |
| 8 | 14.24330 ± 0.00012 | 1775.0 | 9.263 | 1.351530 ± 0.000084 | 0.49 |

UNCLASSIFIED


Uranium Standard Solutions (1-8)

| Soln. No. | Temp (°C) | [U] @ spec. Temp (mM) | Final Solution H_2SO_4 Conc. Ignoring UO_3 Neutralization (M) ^a | Final Solution H_2SO_4 Conc. Including UO_3 Neutralization (M) ^b |
|-----------|-----------|-----------------------|---|--|
| 1 | 29.5 | 5.0060 ± 0.0007 | 0.997 | 0.992 |
| 2 | 29.8 | 10.0388 ± 0.0012 | 0.997 | 0.987 |
| 3 | 29.5 | 15.0623 ± 0.0018 | 0.997 | 0.982 |
| 4 | 28.5 | 20.1093 ± 0.0024 | 0.997 | 0.977 |
| 5 | 29.3 | 30.0961 ± 0.0034 | 0.898 | 0.868 |
| 6 | 28.7 | 30.1897 ± 0.0035 | 0.997 | 0.967 |
| 7 | 29.4 | 30.3840 ± 0.0035 | 1.097 | 1.067 |
| 8 | 27.1 | 60.9267 ± 0.0068 | 0.996 | 0.936 |

Rare discussion of molarity.

UNCLASSIFIED



Raman Spectra of Solutions 1-8 confirming nitrate removal



- Spectra all very similar to that of 1 M H_2SO_4
- Less nitrate to remove during preparation of solutions 1-8 vs. solutions 11-12 (see later discussion)

UNCLASSIFIED

Density Measurements of Solutions 1-8 Mettler Toledo DM50 (20.0, 27.5 & 35.0 °C)



- Samples 1, 2, 3, 4, 6 & 8 all “1” M H_2SO_4
- Samples 5, 6 & 7 all the same [U]
- Sample 5 = “0.9” M H_2SO_4
- Sample 7 = “1.1” M H_2SO_4

UNCLASSIFIED

Slide 25

Accuracy, Linearity and Range Precision and Ruggedness

Cell Manufacturer and Path Length

Hellma Analytics
 $(10.00 \pm 0.01 \text{ mm})$

NSG Precision
 $(10.00 \pm 0.02 \text{ mm})$

Starna Cells
 $(10.00 \pm 0.01 \text{ mm})$

Starna Cells
 $(2.00 \pm 0.01 \text{ mm})$


Starna Cells
 $(100.00 \pm 0.02 \text{ mm})$

UNCLASSIFIED

Accuracy, Linearity & Range, Precision & Ruggedness

Molar Absorptivity Data


10 mm Starna Cells, “1 M” H_2SO_4 Solutions Only



UNCLASSIFIED

Molar Absorptivity Data - 10 mm Starna Cells

Includes solutions 5 & 7 - "0.9 M" & "1.1 M" H_2SO_4



UNCLASSIFIED

Molar Absorptivity 10 mm Starna Cell Analysis and Observations

- Temperature dependence on molar absorptivity value could not be ignored
- Discarded data for lowest [U] concentration standard solution, solution 1
- Cells 1-3 – very similar performance
- Technique still valid for “ 1.0 ± 0.1 M” H_2SO_4
- Lower temp measurements – lower molar absorptivity values
 - Observed in literature for related uranyl systems, see Rao, L. & Tian, G. J. *Chem. Thermodynamics*, **40** (2008) 1001-1006
- Higher temperature measurements recorded in June 2015, with lower temperature measurements recorded in November 2015
- Only room temperature recorded, not solution temperature.
- Likely a small concentration impact on observed molar absorptivity, but the temperature effect more significant.

UNCLASSIFIED


Molar Absorptivity 10 mm Starna Cell Analysis and Observations

- No measurement of sample evaporation between June and November
 - Davies-Gray results on solution 11 & 12 show significance of evaporation (later discussion)
 - Scatter in lower temperature measurements, increase molar absorptivity uncertainties and may point to different extents of evaporation in different solutions
 - Evaporation would also lead to higher [U], raising absorbance value @ 419.5 nm and thus increasing the 'observed' molar absorptivity values
 - Any increase in observed molar absorptivity will lead to lower observed [U] conc. measurements by developed UV-vis technique

UNCLASSIFIED

Molar Absorptivity Data


Comparison - 10 mm Starna and 10 mm Hellma Analytics Cells



UNCLASSIFIED

Molar Absorptivity Data

Comparison - 10 mm Starna and 10 mm NSG Cells



UNCLASSIFIED

Molar Absorptivity


Comparison of 10 mm Starna, Hellma and NSG Cells

- Starna Cells
 - General consensus among the three operators that these cells were the most “user friendly”
- Hellma cell data recorded at highest temperature range
 - Providing further evidence of significant temperature impact on molar absorptivity
- Lower precision in cell path length for the NSG cells reflected in increased measurement uncertainties
- Appear to be a trend of small decrease in molar absorptivity going through cells 1, 2 & 3 for each of the cell manufacturers
 - No current rationale for this observation

UNCLASSIFIED

Molar Absorptivity Data


Comparison 2 mm Starna cells vs. 10 mm Starna Cells



UNCLASSIFIED

Molar Absorptivity Data


Comparison 100 mm Starna cells vs. 10 mm Starna Cells



UNCLASSIFIED

Molar Absorptivity 2 mm & 100 mm Starna Cells Observations

- 2 mm cells
 - Significant increase in path length uncertainties
 - Can reduce solution volume without dropping from 10 mm to 2 mm path length cell
- 10 mm cells
 - Should provide the lowest path length uncertainties, but difficult to insert into spectrophotometer cell holder (position reproducibility issues)
 - Solution [U] concentrations. too high for detailed study
 - absorbance values @ 419.5 nm for solutions 3 & 4 > 2
 - Could merit further investigation if issues are addressed



www.starnacells.com

UNCLASSIFIED

Test Solutions 11-15

- Solutions 11 & 12
 - Prepared as described for solutions 1-8 using weighed aliquots of 'Cut B' CRM 112-A in nitric acid
 - Differences in sample preparation 11 & 12 vs. 1-8
 - A much greater mass of uranium required (these samples would be assayed)
 - Adjusted to a final H_2SO_4 conc. of '0.1 M' (pH 1) vs. 1.0 M
- Solutions 13 & 15 – repurposed depleted uranium pH 1 stocks already 'in hand'
- Solutions 14
 - Prepared using commercially available uranyl sulfate tri-hydrate ($\text{UO}_2\text{SO}_4 \cdot 3\text{H}_2\text{O}$)
 - International Bio-Analytical Industries (IBI)
 - Dissolution in H_2O revealed presence of some insoluble white solids, discarded through centrifugation
 - Adjusted to pH 1 using H_2SO_4



UNCLASSIFIED

Test Solutions 11-15

| Solution Number | U Relative Atomic Weight (g/mol) | [U] (Wt %) | Final Soln Mass (g) | Final Soln pH |
|-----------------|----------------------------------|---------------------|---------------------|---------------|
| 11 | 238.028918 ± 0.000012 | 7.4144 ± 0.0005 | 67.4623 | 1.04 |
| 12 | 238.028918 ± 0.000012 | 11.5183 ± 0.0007 | 65.4698 | 0.90 |
| 13 | 238.04±0.01 dep U | Unknown | 106.6656 | 0.92 |
| 14 | 238.04±0.01 dep U | 8.3879 ± 0.0004 | 64.6229 | 1.04 |
| 15 | 238.04±0.01 dep U | Unknown | 133.4595 | 0.94 |

UNCLASSIFIED

Density Measurements of Solutions 11-15 and 11-D to 15-D (20.0, 27.5 & 35.0 °C)




D solutions – assay samples, already diluted in “1 M” H₂SO₄

UNCLASSIFIED

Slide 39


Raman Spectra of Solutions 11-15



- As would be expected, clearly more UO_2^{2+} (uranium) in Solution 11-15 vs. solutions 1-8
- SO_4^{2-} region of spectra more complex vs. solns 1-8, UO_2^{2+} complex bands more dominant
- Solution 14 – peak in HSO_4^- region looks different vs. Solutions 11, 12, 13 & 15

UNCLASSIFIED

Raman Spectra of Solutions 11-15



- Solutions 11-13 & 15 – no evidence of presence of NO₃⁻
- Solution 14 – peak in HSO₄⁻ region appears to be sharper and higher in energy, evidence of NO₃⁻?
- Solution 14 – prepared from UO₂SO₄·3H₂O purchased from IBI
 - Evidence that this chemical was prepared from a uranium nitrate starting material?

UNCLASSIFIED

Samples 11-15 UV-vis Analysis

- A, B & C assay analysis
 - 100 μ L of sample + 2000 μ L of “1 M” H_2SO_4 (mass measurements)
- D assay analysis
 - 1.4 mL of sample + 28 mL of “1 M” H_2SO_4 (mass measurements)
- Each assay sample analyzed in all three 10 mm Starna cells
 - Triplicate “1 M” H_2SO_4 background reads recorded
 - Triplicate sample reads recorded


UNCLASSIFIED

Samples 11-15 UV-vis Analysis

- Input parameters to spreadsheet used to calc. [U] (g/g)
 - Absorbance value @ 419.5 nm
 - Average triplicate sample reads background corrected, and then baseline corrected
 - The room temperature at the time of the measurement
 - Molar absorptivity at measurement temperature
 - For A, B, C & D samples
 - Density of both 1 M H_2SO_4 and the specific assayed sample at the UV-vis measurement temperature
 - For D' analysis of the D samples
 - Density of the D samples at the UV-vis measurement temperature

UNCLASSIFIED

Comparison of Methods for Analyzing Solutions 11-15 vs. Davies-Gray Results



All future discussions will relate to the updated method of UV-vis data analysis

UNCLASSIFIED

Slide 44

Accuracy, Linearity & Range Precision & Ruggedness Comparative Analysis

Uranium Concentration Analysis - Solutions 11-14 A, B & C assay measurements

| Soln No. | Cell # | Gravimetric | | Soln A Assay | | Soln B Assay | | Soln C Assay | | Davies-Gray Msmt | |
|-----------|--------|------------------------|----------|------------------------|----------|------------------------|----------|------------------------|----------|------------------------|----------|
| | | Value (mg U/g soln) | σ |
| 11 | 1 | 74.144 | 0.005 | 73.50 | 0.29 | 73.30 | 0.28 | 73.35 | 0.28 | 74.34 | 0.07 |
| | 2 | | | 73.60 | 0.29 | 73.27 | 0.27 | 73.31 | 0.28 | | |
| | 3 | | | 73.37 | 0.30 | 73.09 | 0.28 | 73.10 | 0.29 | | |
| 11 repeat | 1 | 74.144 | 0.005 | 73.51 | 0.28 | 73.85 | 0.29 | 73.71 | 0.30 | 74.34 | 0.07 |
| | 2 | | | 73.31 | 0.34 | 73.78 | 0.29 | 73.82 | 0.29 | | |
| | 3 | | | 73.38 | 0.29 | 73.52 | 0.30 | 73.58 | 0.30 | | |
| 12 | 1 | 115.183 | 0.007 | 114.50 | 0.44 | 114.52 | 0.45 | 114.52 | 0.44 | 115.24 | 0.08 |
| | 2 | | | 114.32 | 0.43 | 114.35 | 0.44 | 114.42 | 0.43 | | |
| | 3 | | | 114.16 | 0.44 | 114.23 | 0.45 | 114.28 | 0.44 | | |
| 13 | 1 | Unknown | | 121.74 | 0.49 | 122.11 | 0.48 | 121.98 | 0.48 | 123.59 | 0.09 |
| | 2 | | | 121.73 | 0.48 | 122.03 | 0.47 | 122.10 | 0.47 | | |
| | 3 | | | 121.63 | 0.50 | 121.89 | 0.49 | 121.94 | 0.49 | | |
| 14 | 1 | 83.879 | 0.004 | 76.43 | 0.32 | 76.40 | 0.31 | 76.22 | 0.31 | 77.23 | 0.07 |
| | 2 | | | 76.27 | 0.31 | 76.38 | 0.31 | 76.09 | 0.31 | | |
| | 3 | | | 76.25 | 0.33 | 76.28 | 0.31 | 75.99 | 0.31 | | |

UNCLASSIFIED

Accuracy, Linearity & Range Precision & Ruggedness Comparative Analysis

Uranium Concentration Analysis Solution 15 A, B & C Assay Measurements

| Soln No. | Cell # | Gravimetric | | Soln A Assay | | Soln B Assay | | Soln C Assay | | Davies-Gray Msmt | |
|----------|--------|---------------------------|----------|------------------------|----------|------------------------|----------|------------------------|----------|------------------------|-----------|
| | | Value (mg U/g soln) | σ | Value (mg U/g soln) | 2σ |
| 15 (AMH) | 1 | Unknown | | 126.98 | 0.51 | 126.88 | 0.50 | 126.95 | 0.50 | 127.58 | 0.09 |
| | 2 | | | 126.90 | 0.49 | 126.77 | 0.49 | 126.83 | 0.48 | | |
| | 3 | | | 126.82 | 0.51 | 126.69 | 0.50 | 125.53 | 0.49 | | |
| 15 (IM) | 1 | Unknown | | 127.19 | 0.49 | 127.17 | 0.49 | 127.36 | 0.49 | 127.58 | 0.09 |
| | 2 | | | 127.03 | 0.48 | 127.17 | 0.47 | 127.10 | 0.48 | | |
| | 3 | | | 126.99 | 0.50 | 127.52 | 0.50 | 127.17 | 0.50 | | |
| 15 (SR) | 1 | Unknown | | 127.50 | 0.49 | 127.44 | 0.49 | 127.44 | 0.49 | 127.58 | 0.09 |
| | 2 | | | 127.32 | 0.47 | 127.22 | 0.47 | 127.22 | 0.47 | | |
| | 3 | | | 127.50 | 0.49 | 127.34 | 0.49 | 127.28 | 0.49 | | |

Chemist

AMH Alex Meckler-Hickson
IM Iain May
SR Sean Reilly

UNCLASSIFIED



Operated by Los Alamos National Security, LLC for the U.S. Department of Energy's NNSA

Slide 46



Uranium Concentration Analysis

Solutions 11-15

D assays

Accuracy
Linearity & Range
Precision & Ruggedness
Comparative Analysis

| Solution Number | Cell # | Gravimetric | | Soln D Assay | | Soln D' Assay | | Davies-Gray Msmt | |
|-----------------|--------|---------------------|----------|---------------------|----------|---------------------|----------|---------------------|-----------|
| | | Value (mg U/g soln) | σ | Value (mg U/g soln) | σ | Value (mg U/g soln) | σ | Value (mg U/g soln) | 2σ |
| 11 | 1 | 74.144 | 0.005 | 74.24 | 0.29 | 74.23 | 0.29 | 74.34 | 0.07 |
| | 2 | | | 74.16 | 0.28 | 74.15 | 0.29 | | |
| | 3 | | | 74.12 | 0.28 | 74.11 | 0.28 | | |
| 11 repeat | 1 | 74.144 | 0.005 | | | | | 74.34 | 0.07 |
| | 2 | | | | | | | | |
| | 3 | | | | | | | | |
| 12 | 1 | 115.183 | 0.007 | 115.12 | 0.44 | 115.11 | 0.44 | 115.24 | 0.08 |
| | 2 | | | 114.90 | 0.43 | 114.90 | 0.43 | | |
| | 3 | | | 114.95 | 0.44 | 114.95 | 0.44 | | |
| 13 | 1 | Unknown | | 122.67 | 0.47 | 122.66 | 0.47 | 123.59 | 0.09 |
| | 2 | | | 122.47 | 0.46 | 122.46 | 0.46 | | |
| | 3 | | | 122.57 | 0.47 | 122.56 | 0.47 | | |
| 14 | 1 | 83.879 | 0.004 | 76.86 | 0.30 | 76.81 | 0.30 | 77.23 | 0.07 |
| | 2 | | | 76.77 | 0.29 | 76.70 | 0.29 | | |
| | 3 | | | 76.85 | 0.30 | 76.78 | 0.30 | | |
| 15 | 1 | Unknown | | 126.87 | 0.48 | 126.90 | 0.48 | 127.58 | 0.09 |
| | 2 | | | 126.66 | 0.47 | 126.69 | 0.47 | | |
| | 3 | | | 126.75 | 0.48 | 126.78 | 0.48 | | |

UNCLASSIFIED




Operated by Los Alamos National Security, LLC for the U.S. Department of Energy's NNSA

Slide 47



Accuracy, Precision & Ruggedness, Comparative Analysis

Solution 11




- Green dotted lines = \pm the expanded uncertainty of the Davies-Gray value for [U] (74.34 mg/g).
- Black dotted lines = $\pm 2\sigma$ of the gravimetrically determined [U] (74.144 mg/g)
- All error bars are $\pm 1\sigma$ of UV-vis determined [U]

UNCLASSIFIED

Accuracy, Precision & Ruggedness, Comparative Analysis

Solution 12




- Green dotted lines = \pm the expanded uncertainty of the Davies-Gray value for $[U]$ (115.24 mg/g).
- Black dotted lines = $\pm 2\sigma$ of the gravimetrically determined $[U]$ (115.183 mg/g)
- All error bars are $\pm 1\sigma$ of UV-vis determined $[U]$

UNCLASSIFIED

Accuracy, Precision & Ruggedness, Comparative Analysis

Solution 13




- Green dotted lines = \pm the expanded uncertainty of the Davies-Gray value for $[U]$ (123.59 mg/g).
- All error bars are $\pm 1\sigma$ of UV-vis determined $[U]$

UNCLASSIFIED

Accuracy, Precision & Ruggedness, Comparative Analysis

Solution 14




- Green dotted lines = \pm the expanded uncertainty of the Davies-Gray value for [U] (77.23 mg/g).
- Gravimetrically determined [U] not shown ("83.879 \pm 0.004 mg/g")
- All error bars are $\pm 1\sigma$ of UV-vis determined [U]

UNCLASSIFIED

Accuracy, Precision & Ruggedness, Comparative Analysis

Solution 15




- Green dotted lines = \pm the expanded uncertainty of the Davies-Gray value for [U] (127.58 mg/g).
- All error bars are $\pm 1\sigma$ of UV-vis determined [U]

UNCLASSIFIED

Accuracy, Precision & Ruggedness, Comparative Analysis

Solution 15 Different Operators



UNCLASSIFIED

Accuracy, Linearity & Range Precision & Ruggedness, Comparative Analysis

UV-Vis Analysis of Solutions 11-15 - Observations

- The UV-vis measurements trend to lower measured [U] values than both gravimetric and Davies-Gray data
 - Impact of solution evaporation not taken into consideration
 - Observed for gravimetric vs. Davies-Gray
 - Impact on Molar Absorptivity values previously discussed
 - Measurement of air temperature versus solution temperature
 - For most solutions, better agreement with D and D' samples vs. A, B & C samples
 - A, B & C sample volumes only just enough for analysis, lower absorbance if light beam not completely passing through sample
- Good agreement in results obtained by different operators (Solution 15)
- Density correction due to assay 'mixing' is not significant (D vs. D' analysis)

UNCLASSIFIED



Operated by Los Alamos National Security, LLC for the U.S. Department of Energy's NNSA

Slide 54



Impact of Ruthenium

- Only impacts uranium target (fuel) solution after Mo-99 production
- Even a few ppm Ru in solution can potentially increase the 'observed U' abs. @ 419.5 nm by > 0.1 %.
 - Requires a good understanding of both Ru and U process chemistry concentration levels
- A better understanding of ruthenium spectroscopy in both process type solutions and in 1 M H_2SO_4 assay solution would be useful
 - But likely not a showstopper
- Proposed technical solutions: - either define maximum Ru contaminant tolerance (e.g. based on U(VI) peak ratios) and/or subtract out the Ru spectroscopic component during data analysis
 - FY 15 results indicate that both are viable, with a focus on the former

UNCLASSIFIED

Preparation of Ruthenium Solution Samples

- Prepared using $10,000 \pm 100$ ppm ($\mu\text{g/mL}$) ruthenium ICP standard solution
 - NIST traceable, in 20 % HCl
- Dilution of aliquot of this standard solution into Bottle 1 “1.0 M” H_2SO_4 to yield a 100 ppm stock solution
- Subsequent serial dilutions to prepare 10, 1 & 0.1 ppm Ru working solutions in “1.0 M” H_2SO_4

UNCLASSIFIED

Preparation of Solution Samples 21-24

- Measurements made by mass to allow for accurate [U] analysis by UV-vis
- Used ca. 150 gU/L pH 1 dU sulfate stock solution
 - Previously used for FY14 Uranium Detection Task
 - FY14 Sample 15, 2/25/14
- Samples 21-24 prepared by adding 0.5 mL of pH 1 dU sulfate solution to 10 mL of Ru spiked “1.0” M H_2SO_4

| Sample No. | [Ru] (ppm) in “1 M” H_2SO_4 | Effective [Ru] (ppm) <i>if added</i> through original dU sample |
|------------|--|--|
| 21 | 0 | 0 |
| 22 | 0.1 | 2.1 |
| 23 | 1.0 | 21 |
| 24 | 10.0 | 210 |




UNCLASSIFIED

Ruthenium Sample Work 'Caveats'

- Ru standard solution in $\mu\text{g/mL}$, and pipetting used in dilution (i.e. deviation from accurate mass measurements)
 - However, orders of magnitude difference between Ru concs.
- Ru standard in HCl – trace HCl not removed
 - Impact on speciation cannot be ruled out
- For [U] conc. measurements
 - No individual temperature measurements for 1st June 2015 analysis, thus used end of day value of 26.6 °C
 - For FY14 Sample 15, 2/25/14 uranium
 - Density recorded only to 25.0 °C and extrapolated to 26.6 °C (see previous comment, but no impact observed)
 - Density likely increased with uranium concentration due to sample evaporation (negligible &/or uniform impact)
 - Assume density of Bottle 1 sulfuric acid (ignore 'negligible' impact of up to 210 ppm Ru)

UNCLASSIFIED


Ruthenium Scans Spectra average all three 10 mm Starna cells Spectra



- As previously observed, Ru spectroscopic features change as a function of time
- No observed significant concentration effects at >400 nm

UNCLASSIFIED


Ruthenium Scans/Reads Spectra Sample 21



Pink dash lines mark the reads analysis wavelengths @ 600.0, 550.0, 500.0, 430.0, 419.5 & 409.0 nm

UNCLASSIFIED

Ruthenium Scans Spectra Sample 22




UNCLASSIFIED

Ruthenium Scans Spectra

Sample 23


Specificity



UNCLASSIFIED

Ruthenium Scans Spectra Sample 24

Specificity




UNCLASSIFIED

Ruthenium Scans Spectra

Sample 24 – U(VI) Subtracted Out

Specificity



Impact on 430.0, 419.5 & 409.0 nm peak ratios changes with time

UNCLASSIFIED

Accuracy and Specificity

Impact of Ruthenium – “Observed” [U] reads method analysis

| Cell 1 Data | | 1st June '15 | | 1st July '15 | | 21st July '15 | |
|-------------|----------|--------------|----------|--------------|----------|---------------|----------|
| Ru (ppm) | Soln. No | [U] | σ | [U] | σ | [U] | σ |
| 0 | 21 | 129.43 | 0.51 | 129.46 | 0.49 | 129.37 | 0.49 |
| 2.1 | 22 | 129.67 | 0.51 | 129.81 | 0.50 | 129.69 | 0.49 |
| 21 | 23 | 132.50 | 0.54 | 129.46 | 0.50 | 129.31 | 0.51 |
| 210 | 24 | 184.58 | 0.73 | 138.32 | 0.54 | 135.39 | 0.57 |

| Cell 2 Data | | 1st June '15 | | 1st July '15 | | 21st July '15 | |
|-------------|----------|--------------|----------|--------------|----------|---------------|----------|
| Ru (ppm) | Soln. No | [U] | σ | [U] | σ | [U] | σ |
| 0 | 21 | 129.12 | 0.50 | 129.03 | 0.47 | 129.18 | 0.47 |
| 2.1 | 22 | 129.59 | 0.50 | 129.62 | 0.49 | 129.46 | 0.48 |
| 21 | 23 | 132.24 | 0.52 | 129.20 | 0.52 | 129.22 | 0.49 |
| 210 | 24 | 185.83 | 2.42 | 138.20 | 0.59 | 135.29 | 0.52 |


| Cell 3 Data | | 1st June '15 | | 1st July '15 | | 21st July '15 | |
|-------------|----------|--------------|----------|--------------|----------|---------------|----------|
| Ru (ppm) | Soln. No | [U] | σ | [U] | σ | [U] | σ |
| 1 | 21 | 129.44 | 0.51 | 128.99 | 0.49 | 129.22 | 0.49 |
| 2.1 | 22 | 129.71 | 0.52 | 129.69 | 0.51 | 129.60 | 0.49 |
| 21 | 23 | 131.85 | 0.53 | 129.46 | 0.51 | 129.35 | 0.51 |
| 210 | 24 | 187.17 | 1.61 | 138.51 | 0.58 | 135.24 | 0.55 |

UNCLASSIFIED

Slide 65

Impact of Ruthenium

Monitoring 409/419.5 nm Abs. Ratios - 1st June 2015 Data




- High uncertainties for Solution 24 absorbance ratios related to the applied reads analysis baseline methodology (a recurring theme)
- Impact of 210 ppm Ru (Solution 24) clearly observed through both peak ratio and '[U]'

UNCLASSIFIED

Impact of Ruthenium

Monitoring 409/419.5 nm Abs. Ratios - 1st June 2015 Data




- Impact of 21 ppm Ru (Solution 23) clearly observed through both peak ratio and '[U]'

UNCLASSIFIED

Impact of Ruthenium

Monitoring 409/419.5 nm Abs. Ratios - 1st June 2015 Data




- Impact of 2.1 ppm Ru (Solution 22) might be observed through peak ratio uncertainties, but not observed in '[U]'.

UNCLASSIFIED

Impact of Ruthenium

Monitoring 430/419.5 nm Abs. Ratios - 1st June 2015 Data




- Impact of 210 ppm Ru (Solution 24) clearly observed through both peak ratio and '[U]'

UNCLASSIFIED

Impact of Ruthenium

Monitoring 430/419.5 nm Abs. Ratios - 1st June 2015 Data




- Impact of 21 ppm Ru (Soln 23) might be observed through peak ratio/peak ratio uncertainties, and clearly observed in '[U]'.

UNCLASSIFIED

Impact of Ruthenium

Monitoring 430/419.5 nm Abs. Ratios - 1st June 2015 Data




- Impact of 2.1 ppm Ru (Soln 22) might be observed through peak ratio uncertainties, but not observed in '[U]'.

UNCLASSIFIED

Impact of Ruthenium

Monitoring 500/419.5 nm Abs. Ratios - 1st June 2015 Data




- Impact of 210 ppm Ru (Soln 24) clearly observed through both peak ratio and '[U]'.

UNCLASSIFIED

Impact of Ruthenium

Monitoring 500/419.5 nm Abs. Ratios - 1st June 2015 Data




- Impact of 21 ppm Ru (Solution 23) clearly observed through both peak ratio and '[U]'.

UNCLASSIFIED

Impact of Ruthenium

Monitoring 500/419.5 nm Abs. Ratios - 1st June 2015 Data




- Impact of 2.1 ppm Ru (Soln 22) might be observed through peak ratio uncertainties, but not by '[U]'.

UNCLASSIFIED

Impact of Ruthenium

Monitoring 409/419.5 nm Abs. Ratios - 21st July 2015 Data




- Impact of 210 ppm Ru (Solution 24) clearly observed through both peak ratio and '[U]'

UNCLASSIFIED

Impact of Ruthenium

Monitoring 409/419.5 nm Abs. Ratios - 21st July 2015 Data




- Impact of 21 ppm Ru (Solution 23) probably observed through peak ratio, but not by '[U]'

UNCLASSIFIED

Impact of Ruthenium

Monitoring 430/419.5 nm Abs. Ratios - 21st July 2015 Data




- Impact of 210 ppm Ru (Solution 24) could be observed through both peak ratio/peak ratio uncertainties, and definitely by '[U]'

UNCLASSIFIED

Impact of Ruthenium

Monitoring 500/419.5 nm Abs. Ratios - 21st July 2015 Data



- Impact of 210 ppm Ru (Solution 24) clearly observed through both peak ratio and '[U]'.

UNCLASSIFIED

Accuracy and Specificity

- Intense absorbance with spectral features that change over time
 - Ideal test for almost any potential interfering contaminant
- Monitoring peaks ratios using reads data can be used to indicate Ru interference
 - At least to the level of accuracy of our current [U] UV-vis technique
- Experimental protocol can be developed to screen data where peaks ratios/peak ratio uncertainties are outside a certain tolerance ratio

UNCLASSIFIED

FY 16 Tasks

- Temperature control block
 - All measurements at one temperature
- Simplifies density measurements – now all at one temperature
- Refine our Molar Absorptivity value
 - Use a fresh batch of CRM 112-A
- Repeat measurements of samples 11-15, and additional unknown samples
- Multiple Davies-Gray measurements
 - Correct for any sample evaporation
- Evaluate NIST U standard solution in nitric acid (3164) – 9.994 mg/g
 \pm 0.016 mg/g

UNCLASSIFIED

Example Temperature Control Unit



Peltier-Controlled Cuvette and Custom Instrumentation
Quantum Northwest
www.qnw.co,/versa20cary60/

UNCLASSIFIED