

Abstract for presentation to the Rio Grande Chapter of the Health Physics Society 2017 Spring Technical Meeting

Title: Neutron Dose in Whole Body Dosimeter Algorithm at Sandia National Laboratories

Presented by: David Farrar

Abstract:

The whole body dosimeter used at Sandia National Laboratories is used to measure occupational dose to personnel from external radiation. A discovery was made regarding the determination of dose measured by the dosimeter. There are instances where dose from neutrons could be missed. This presentation discusses the dosimeter; provides the overall dosimeter-to-dose process; and describes the discovery, its impact, and its remedy.

*Exceptional service in the national interest*



# Neutron Dose in Whole Body Dosimeter Algorithm at Sandia National Laboratories

David R. Farrar, CHP  
Radiation Protection Department  
Radiation Protection Dosimetry Program



Sandia National Laboratories is a multi-program laboratory managed and operated by Sandia Corporation, a wholly owned subsidiary of Lockheed Martin Corporation, for the U.S. Department of Energy's National Nuclear Security Administration under contract DE-AC04-94AL85000.

# Outline

- SNL whole body dosimeter and algorithm for calculating dose
- Algorithm for dose from neutrons
- Discovery
- Impact
- Remedy

# SNL Dosimeters

- Two types of personnel dosimeters used at SNL to measure occupational dose
  - Extremity dosimeter
    - Worn on finger
    - Not sensitive to neutrons
  - Whole body dosimeter
    - Worn on chest
    - Sensitive to neutrons
    - Focus of presentation, “dosimeter” hereafter



# Dosimeter


- Consists of 4 thermoluminescent dosimeter (TLD) chips
- Chips read in reader to get data about radiation exposure
- Each chip sensitive to radiation
  - Chips 1-3: non-neutrons only
  - Chip 4: non-neutrons & neutrons




# Dosimeter → Dose (Algorithm)

- Dosimeter data used in an algorithm to calculate occupational doses
  - Algorithm calculates deep dose from photons, deep dose from neutrons, eye dose, and shallow dose
  - Algorithm contains decision point on potential neutron exposure of dosimeter
    - If potential neutron exposure found, assessment of dose from neutrons made
  - Prior to the decision point the algorithm has another decision point due to a data test


# Dose algorithm



# Dose algorithm



# Dose algorithm



# Data Test

- Algorithm tests ratios of chip responses
  - Ratios are checked against boundary conditions
  - No tests consider response from neutron-sensitive chip (L4)
- If data “questionable”, algorithm assumes Cs-137 exposure and does not consider neutron exposure as a possibility

# Discovery

- In late 2014, examination of dosimeter data found a couple slightly elevated chip 4 (neutron sensitive) readings for individuals that work around devices capable of generating neutrons
- Due to previous work experience supporting devices, knew that neutron exposure was possible and could result in non-trivial doses
- Examination of algorithm results found the algorithm deemed these dosimeter results to be “questionable”
  - No assessment of neutron exposure performed by algorithm

# Work Around

- Work around: Added step in dose approval process to manually check for missed neutron doses
  - Prior to dose approval, dosimeter data checked with spreadsheet
  - Check performs neutron exposure evaluation regardless if dosimeter data “questionable”
    - Same evaluation used in algorithm
  - If check determines neutron exposure with possible dose consequence may have been missed by algorithm, issue dose investigation

# Impact

- So far:
  - Found ~15 dosimeters assigned to personnel where algorithm missed neutron dose
    - Each missed neutron dose ~40 mrem, maximum 50 mrem
  - Found ~10 dosimeters used for area monitoring where algorithm missed neutron dose
    - All from areas that are evacuated prior to neutron exposure

Avg Dose (rem)/Shot			
Deep	Shallow	Neutron	Neut Sig (gU) Location
0.029	0.029	7.957	429.28 1) - +90 deg - 1 - morning
0.028	0.028	6.930	374.99 2) - +90 deg - 2 - morning
0.043	0.043	6.727	377.30 3) - +90 deg - 3 - morning
0.027	0.027	7.578	409.71 4) - +90 deg - 4 - morning
0.029	0.029	7.105	383.61 5) - 0 deg - 1 - morning
0.031	0.042	<LLD	393.20 6) - 0 deg - 2 - morning
0.031	0.041	<LLD	376.12 7) - 0 deg - 3 - morning
0.030	0.030	7.801	420.44 8) - 0 deg - 4 - morning

Area dosimeters exposed to device capable of generating neutrons. “0 deg” dosimeters exposed to same neutron field.

- Fix: Working to modify SNL dose algorithm
  - Reverse steps of data test and determination of neutron exposure in algorithm
    - No longer need manual check
      - Save time and reduce human errors
  - Tested modified algorithm with data from Department of Energy Laboratory Accreditation Program (DOELAP) performance tests
    - Performance tests based on exposing dosimeters to isotopes (including deuterium-moderated Cf-252) and/or x-rays
    - Results from modified algorithm identical to results from current algorithm
  - Got approval from DOE Laboratory Accreditation Program to use algorithm
    - Updating documentation prior to 2017 implementation

# Future Algorithm

