

LA-UR-18-31101 (Accepted Manuscript)

Methane Gas Sensing Technologies in Combustion: Comprehensive Review

Brosha, Eric Lanich
Shuk, Pavel
McGuire, Chad

Provided by the author(s) and the Los Alamos National Laboratory (2019-03-22).

To be published in: Sensors & Transducers

DOI to publisher's version:

Permalink to record: <http://permalink.lanl.gov/object/view?what=info:lanl-repo/lareport/LA-UR-18-31101>

Disclaimer:

Approved for public release. Los Alamos National Laboratory, an affirmative action/equal opportunity employer, is operated by the Los Alamos National Security, LLC for the National Nuclear Security Administration of the U.S. Department of Energy under contract DE-AC52-06NA25396. Los Alamos National Laboratory strongly supports academic freedom and a researcher's right to publish; as an institution, however, the Laboratory does not endorse the viewpoint of a publication or guarantee its technical correctness.

Methane Gas Sensing Technologies in Combustion: Comprehensive Review

¹ Pavel SHUK, ¹ Chad MCGUIRE and ² Eric BROSHA

¹ Rosemount Measurement and Analytical, Emerson Automation Solutions, 6021 Innovation Blvd, Shakopee, MN 55379-4676, USA

² Los Alamos National Laboratory, P.O. Box 1663, Los Alamos, NM 87545, USA

¹ Tel.: +1 (952) 204 6030, fax: +1 (952) 204 2054

¹ E-mail: Pavel.Shuk@Emerson.com

Received: 30 November 2018 /Accepted: 31 December 2018 /Published: 31 January 2019

Abstract: Major methane (CH_4) gas sensing technologies for the application in a combustion environment are reviewed with many theoretical and practical aspects, as well as basic installation and operation details. A comprehensive CH_4 gas sensing technologies review is supported with the latest development trends. Performance and application options for methane measurements in the process using calorimetric, mixed potential electrochemical CH_4 sensor, semiconductor detector and tunable diode (TD) laser and quantum cascade (QC) laser spectroscopy are discussed for the possible application in power generation, chemical production, heating, process control, safety, and quality. Special attention is given to the technology application limits and analyzer's system requirements.

Keywords: Combustion analysis, Flue gas, Methane gas sensing, Calorimetric catalytic CH_4 sensor, Mixed potential CH_4 sensor, Semiconductor CH_4 detector, TDLS, QCLS.

1. Introduction

In the regular combustion process, hydrocarbon fuel (C_xH_y) would react with oxygen (O_2) producing primarily carbon dioxide (CO_2) and water (H_2O) with traces of other gases, e.g., sulfur dioxide (SO_2), and nitrogen oxides (NO_x) coming from nitrogen and fuel impurities oxidation. The remaining oxygen in the flue gas is measured for reliable and safe combustion control and optimized combustion efficiency. Zirconia oxygen potentiometric technology was applied successfully in the combustion control of power generation boilers, hot stoves for steelmaking, heating and combustion exhaust gas control of coke ovens for steelmaking, lime and cement kilns combustion control, incinerator combustion control, combustion control of heating furnaces for the oil refinery and petrochemical industry, and many other applications

[1-4]. Combustion control is mostly accomplished with O_2 measurement alone (Fig. 1) and an improved combustion efficiency and stability can be achieved with the concurrent measurement of carbon monoxide at $\sim 100\ldots 200$ ppm CO levels and $\sim 1\ldots 6\%$ of oxygen depending on the fuel type (Figs. 1-2). Monitoring of CO in the combustion process offers a new unique opportunity for safe flame operation with CO serving as a marker indicating incomplete or fuel-rich malfunction [5].

Three comprehensive carbon monoxide gas sensing technologies implemented so far in combustion analyzers on the market – calorimetric catalytic, mixed potential electrochemical, and Tunable Diode Laser Spectroscopy (TDLS) or Quantum Cascade Laser Spectroscopy (QCLS) – were recently reviewed [6].

At combustion startup or in the event of a burner malfunction or flame blowout, the combustor can be filled with an explosive mixture faster than the response time of normal flame sensors installed in a combustor. An extra methane (CH₄) measurement in a

combustion analyzer [7] would provide an extra safety feature for combustion control and diagnostics (see combustion flue gas diagram, Fig. 1). It will also increase combustion performance efficiency (Fig. 3).




Fig. 1. Combustion flue gas diagram with methane control needed for safety.




Fig. 2. Oxygen and carbon monoxide optimization in combustion control at coal fired Power Plant.





Fig. 3. Combustion efficiency dependence on temperature and oxygen excess.

2. Mixed Potential Electrochemical CH₄ Sensor

Mixed-potential gas sensing technology with an oxygen ion conducting zirconia solid electrolyte and two different activity electrodes invented by Sandler [8] was proposed for the detection of combustible gases. The presence of CH₄ or any other combustible species will affect the electrochemical oxygen sensor signal with competing reactions of oxygen reduction (1) and CH₄ or other combustible species oxidation (2) establishing mixed potential highly dependent on the CH₄ concentration (3) (see diagram, Fig. 4):



$$E_{\text{mix}} = E_{\text{eq}} - \text{Const} * \ln c (CH_4) \quad (3)$$

Mixed potential developed by the CH₄ sensor would depend mostly on kinetic factors and would be

a strong function of the process electrode electrochemical and catalytic activities and morphology. The sensor signal will be affected by changes in the gas/electrolyte/electrode triple phase boundary (TPB) area, limited methane or oxygen diffusion in the gas phase, electrodes and TPB, limited adsorption, and other combustion gas interference. The fact that the response is controlled by reaction kinetics places a limit on a maximum operating temperature for the device in order to achieve a usable signal-to-noise in the application. Various metals and metal oxide materials were investigated for sensitive electrodes in the mixed potential sensors in the past [9] and a special signal conditioning pulsed discharge technique was proposed recently to increase sensor sensitivity and selectivity [10-11].

The mixed potential CH₄ sensor is very sensitive and quite reproducible (Fig. 5) with a detection limit of <25 ppm CH₄. Unfortunately, mixed potential sensors reported to date have not been especially selective with strong interference to other hydrocarbons (CxHy), nitrogen oxides (NOx), carbon dioxide (CO₂) and carbon monoxide (CO) (Figs. 6-7) being shown.

Electrode	Catalytically active equilibrium electrode (Pt)	Catalytically inactive mixed potential electrode (Au alloy, oxide or composite)
Models		
Reactions	$\frac{1}{2} O_2 + V_O^{2+} + 2e^- \leftrightarrow O \text{ (YSZ)}$	$CH_4 + O \text{ (YSZ)} \leftrightarrow CO_2 + H_2O + V_O^{2+} + 2e^-$
Equations	$E_{\text{eq}} = \frac{RT}{4F} \ln \frac{p(O_2)_{\text{meas}}}{p(O_2)_{\text{ref}}}$	$E_{\text{noneq}} = \text{Const} * \ln c (CH_4)$

Fig. 4. Mixed potential CH₄ sensor electrochemical reactions diagram.




Fig. 5. Sensitivity of mixed potential CH₄ sensor with tin-doped indium oxide sensitive electrode (Adapted from [12], Copyright 2016, with permission from Elsevier).




Fig. 6. Selectivity of mixed potential CH₄ sensor with tin-doped indium oxide sensitive electrode (Adapted from [12], Copyright 2016, with permission from Elsevier).




Fig. 7. Selectivity of mixed potential sensors with Rh, Ir, Ru and Au sensitive electrodes (Adapted from [13], Copyright 2011, with permission from Elsevier).

Sensitivity of a mixed potential sensor can be highly improved (Fig. 8) using a special signal conditioning pulsed discharge technique [10-11]. However, despite considerable work over the years by many groups [9] to achieve a high degree of selectivity by manipulation of such factors as working electrode composition, and application of current bias to alter electrode reaction rates, achieving absolute selectivity is still a limitation of the technology.

Cross-sensitivity effects – particularly to hydrocarbons – at present prohibit application in a real combustion process with a wide temperature and pressure variation and highly challenging flue gas chemical composition. Recent efforts to improve mixed potential technology have focused on improving electrochemical interface morphology in order to permit fabrication of stable and reproducible devices.

It has been recently shown [14-15] that once drift and aging effects are eliminated, when arrays (consisting of 3-4 electrode elements) of stable mixed potential sensors are used together with the appropriate response model [14-16], the intrinsic cross-interference behavior becomes a means to isolate the target species in a more complex background gas

mixture to reasonable accuracy (e.g. 1-2 % error). Although this has not yet been demonstrated for CH₄ sensing, recent advancements in automotive mixed potential design and use may prove beneficial to stationary applications.




Fig. 8. Mixed potential CH₄ sensor sensitivity using conditioning pulsed discharge technique (Adapted from [12], Copyright 2016, with permission from Elsevier).

3. Calorimetric Catalytic CH₄ Sensor

For extractive combustion analyzers, advanced calorimetric catalytic CO sensors were developed by several analytical corporations utilizing combustible reaction heat on the special selective catalyst film over the resistance temperature detector (RTD) implemented in the sensor package for temperature measurements [17-18].


Calorimetric catalytic CO sensor packaged with protective and thermo-conductive layers over RTD film (Fig. 9) is highly reproducible (Fig. 10) in the flow and temperature controlled extractive measurements environment interference to all other combustion species, including methane (Table 1). The catalyst film was merged with RTD using highly thermally conductive packaging oxide materials to utilize more efficiently the heat transfer of the combustion reaction on catalyst film to RTD.

Table 1. Calorimetric catalytic CO sensor typical cross-sensitivity [6].


Gas of interest	CO sensor cross-sensitivity
H ₂	2 : 1
O ₂	1 : >10,000
SO ₂	1 : >1000
NO _x	1 : >1000
CH ₄	1 : >20,000
H ₂ O	1 : >1000

Conventional catalysts for methane combustion are based on noble metals, i.e., palladium (Pd) or platinum (Pt) deposited on oxides such as cerium oxide (CeO₂)

or metals. Methane's complete combustion reaction is highly exothermic (-890 kJ/mol):



releasing ~25 times more heat compared to partial methane oxidation (conversion) to syngas:



So, oxygen sufficiency or even excess would be critical for the calorimetric catalytic CH_4 sensor's reliable performance. Optimized catalytic film would selectively oxidize methane in the presence of other flue gas species. Like the calorimetric catalytic CO sensor implemented on combustion market, the CH_4 sensor would measure temperature and the associated sensor's RTD resistance change in the flue gas with oxygen excess, depending on the methane concentration (Fig. 11).




Fig. 9. Scanning electron micrograph of calorimetric sensor catalyst film/metallic substrate interface [6].




Fig. 10. Calorimetric catalytic CO sensor response and stability [6].




Fig. 11. Calorimetric sensor RTD resistance change dependence on methane concentration.

Catalyst microstructure and sensor temperature optimization would permit quite a reproducible response (Figs. 12-13) and good linearity of the calorimetric CH₄ sensor (Fig. 14).




Fig. 12. Calorimetric CH₄ sensor response and reproducibility in 0...5 % CH₄ mixture with 10 % O₂.




Fig. 13. Calorimetric catalytic CH₄ sensor response to 5 % CH₄ (mixture with 10 % O₂).




Fig. 14. Calorimetric catalytic CH₄ sensor linearity.

A calorimetric catalytic CH₄ sensor has a response of ~30 s for 90 % signal (Fig. 15) and the flue gas flow control extractive environment would contribute to CH₄ detection accuracy, reproducibility, and reliability in the very challenging temperature, pressure, flow and chemical flue gas combustion environment.




Fig. 15. Calorimetric catalytic CH₄ sensor response.

The only calorimetric catalytic CH₄ sensor available on the combustion market is quite selective and accurate with $\sim \pm 5$ % of full scale or ± 0.25 % CH₄ error in extractive methane measurements in flow and temperature-controlled environment.

Typical response of this CH₄ sensor is ~40 s for t₉₀ and an oxygen concentration variation between 3 and 10 % might bring up to a 0.4 % CH₄ error. Some chemicals in combustion like silicon compounds, sulfur oxides, chlorine, and heavy metals might poison catalyst material and will reduce sensor sensitivity and life.

4. Semiconductor CH₄ Detector

Metal oxide semiconductor (MOS) based gas detectors, invented in the 1960 s by Taguchi [19], were intensively investigated to detect different reducing gases, including CH₄ [20-24], and are used nowadays worldwide for the indoor detection of environmental combustible gases [25].

MOS-type gas sensors change resistance (R) as a result of a change in adsorbed oxygen concentration reducing Fermi level and resistance (Fig. 16). The adsorbed oxygen formed in a clean environment (detector's base line) will be consumed by the reaction with methane or carbon monoxide, resulting in surface resistance reduction related to the concentration of methane, carbon monoxide or any other reducing gas. An MOS detector (Fig. 17) is made of sintered n-type metallic oxide from iron, zinc and thin families with imbedded heater holding the sensor temperature at

~250-400 °C. Control of metal oxide physical-chemical properties, i.e., their stoichiometry, porosity, grain size and shape, as well as the proper addition of dopants or surface catalysts, are all effective tools to tune the sensitivity as well as selectivity [26].




Fig. 16. Semiconductors detector operation diagram [25].



Fig. 17. Semiconductors CH_4 detectors [25].

Unfortunately, the implementation of these semiconducting sensors in the combustion process have been limited due to insufficient selectivity, significant interference to moisture and even oxygen, and long-term stability difficulties. Materials that are more stable at high temperatures based on gallium oxide (Ga_2O_3) have been applied in semiconductor sensors [27], and different physical, catalytic gas conversion filters have been recommended for selective indoor CH_4 , H_2 and automotive exhaust gas nitrogen oxides detection [28].

5. Tunable Diode Laser (TDL) and Quantum Cascade Laser (QCL) Spectroscopies

Tunable diode laser (TDL) and quantum cascade laser (QCL) spectroscopies are innovative optical measurement techniques utilizing different lasers (Fig. 18), e.g., tunable diode and quantum cascade, to

detect a variety of combustion gases including methane [29-31].




Fig. 18. Tunable diode (A) and quantum cascade (B) lasers.

TDL and QCL spectroscopies are highly distinguished from the conventional process photometry with the laser's ability to be scanned across the narrow CH_4 absorption peaks many times per second by trimming the current through the laser.

With a typical scan in the range of 0.2 to 0.3 nm, the laser would provide much higher selectivity in applications. The IR absorption spectrum of methane (CH_4) is like a fingerprint, providing CH_4 identification at 1.65 μ (TDL) or 7.9 μ (QCL) and measurements in a wide concentration range down to ~0.6 ppm CH_4 (TDL) or even ~10 ppb CH_4 (QCL) as a detection limit. The detection wavelengths and limits of some combustion gases are summarized in Table 2.

The TDL or QCL analyzer would provide extractive measurement across the duct or pipe (Fig. 19) and would serve as a probe (Fig. 20) for single point near real-time measurement.

A series of process adaption application tools were developed for in-situ, extractive or in cross-pipe installations. The probe design would not require a special optical path alignment, but would deliver single point measurement.

Table 2. Combustion gases absorption lines and detection limits.

Target gas	Wavelength (μ)	Detection limit (ppm)
Methane (CH_4)	1.65 (TDL) 7.9 (QCL)	0.15 0.003
Carbon monoxide (CO)	1.57 2.33 4.80	30 0.5 0.01
Oxygen	0.76	8.956
Carbon dioxide (CO_2)	1.96	3
Nitric oxide (NO)	1.8 2.65	60 1
Nitrogen dioxide (NO_2)	0.68	0.3
Water (H_2O)	1.39	0.06



Fig. 19. A typical installation of TDL analyzer across the gas pipe or combustion duct.




Fig. 20. An in-situ TDL probe with opening for the clean (top) and with diffuser for dusty (bottom) application.

In the probe design the laser source and detector are located in the probe housing with the laser beam being reflected at the probe end, back to the detector by quartz prism, with max operation temperature of <250 °C or by gold mirror with max operation temperature of <430 °C. With a unique folded-path design, the probes can be installed in almost all pipes and stacks with no alignment needed.

Temperature and pressure variation in the process has to be compensated and might cause an additional error in the measurements. IR light reflection at high temperatures, combined with wide background radiation from the fire box and process windows fouling, might bring additional challenges in many applications. There are quite accurate TDL analyzers with $\sim\pm 4$ % or ± 0.02 % CH_4 error and sub-ppm detection limits in some applications.

However, TDL analyzers are quite expensive, but considering the combustion process's highly challenging flue gas temperature, pressure and flow variation, an in-situ multipoint temperature measurement [32-33] and an in-situ CH_4 validation would be highly recommended, so these analyzers provide reliability and accuracy. Some of the available TDL/QCL Analyzers on the market are summarized in Table 3.

Table 3. TDL and QCL Analyzers.

Product and Technology	Detection limit, accuracy	Application
TDLS8000 TDLS	N/A ± 4 % or ± 0.02 % CH_4	Across duct
LaserGas III SP CO (CH_4 optional) TDLS	$0.05\% \pm 1\%$	Across duct or extractive
RLGD-100 TDLS	N/A ± 10 %	Portable CH_4 detector
SERVOTOUGH MiniLaser 3Plus TDLS	0.2 ppm ± 1 %	Across duct or extractive
GM901 TDLS	N/A	Across duct In-situ <430 °C
TDLS GPro 500 TDLS	1 ppm ± 2 % of reading or ± 1 ppm CH_4	In-situ probe <250 °C <600 °C with additional thermal barrier
CT5800 QCLS	0.5 ppm ± 1 %	Extractive

6. Conclusions

From four available major methane gas sensing technologies, only two – catalytic calorimetric and TDLS/QCLS – have so far found practical application in instrumentation in the combustion market, as they produce quite reliable and accurate ($\sim 2\ldots 5$ % error) extractive or across the stack in-situ methane measurements (Table 4).

Table 4. Performance and application summary of major methane gas sensing technologies.

Technology	Temperature (T) and Pressure (P) effect	Major applications
Calorimetric catalytic, Extractive	Fixed temperature, Minor P effect, Flow sensitive	Power generation, Middle to large boilers, Petrochemical industry
TDLS/QCLS Extractive, across the duct, in-situ	T & P algorithms, T & P fluctuation effect?	Petrochemical industry, High temperature furnaces

Acknowledgements

Authors would like to thank Professor Ronald Hanson (Stanford University) for sharing his publications and reports on TDLS/QCLS application in combustion process.

References

- [1]. N. Docquier, S. Candel, Combustion control and sensors: a review, *Progress in Energy and Combustion Science*, Vol. 28, Issue 2, 2002, pp. 107-150.
- [2]. P. Shuk, Process Zirconia oxygen analyzer - state of art, *Technisches Messen*, Vol. 77, Issue 1, 2010, pp. 19-23.
- [3]. P. Shuk, Oxygen gas sensing in combustion process: from basics to application, in *Processing of the 2nd International Conference on Sensors and Electronic Instrumental Advances (SEIA'16)*, Barcelona, Spain, 21-23 September 2016, pp. 8-10.
- [4]. P. Shuk, Oxygen gas sensing technologies application: a comprehensive review, in Sensors for Everyday Life, Smart Sensors, Measurement and Instrumentation, S. C. Mukhopadhyay, *et al.* (Eds.), *Springer, Heidelberg - New York*, 2017, pp. 81-107.
- [5]. R. Hanson, Applications of quantitative laser sensors to kinetics, propulsion and practical energy systems, *Proceedings of the Combustion Institute*, Vol. 33, Issue 1, 2011, pp. 1-40.
- [6]. P. Shuk, C. McGuire, Carbon monoxide gas sensing technologies in combustion process, *Sensors & Transducers*, Vol. 217, Issue 11, 2017, pp. 1-13.
- [7]. P. Shuk, C. McGuire, E. L. Brosha, Methane gas sensing technologies for combustion process, in *Processing of the 4th International Conference on Sensors and Electronic Instrumentation Advances (SEIA'18)*, Amsterdam, The Netherlands, 18-21 September 2018, pp. 22-27.
- [8]. Y. Sandler, Electrochemical sensor for reactive gas mixtures, US patent # 3,981,785, 1976.
- [9]. N. Miura, T. Sato, S. Anggraini, H. Ikeda, S. Zhuiykov, A review of mixed-potential type zirconia-based gas sensors, *Ionics*, Vol. 20, Issue 7, 2014, pp. 901-925.
- [10]. B. Farber, Method of sensor conditioning for improving signal output stability for mixed gas measurements, US patent # 7,585,402, 2008.
- [11]. B. Farber, Method of sensor conditioning for improving signal output stability for mixed gas measurements, US patent # 9,304,101, 2016.
- [12]. P. K. Sekhar, J. Kysar, E. Brosha, C. R. Kreller, Development and testing of an electrochemical methane sensor, *Sensors and Actuators B: Chemical*, Vol. 228, Issue 6, June 2016, pp. 162-167.
- [13]. V. V. Plashnitsa, S. A. Anggraini, N. Miura, CO sensing characteristics of YSZ-based planar sensor using Rh-sensing electrode composed of tetrahedral sub-micron particles, *Electrochemistry Communications*, Vol. 13, Issue 5, May 2011, pp. 444-446.
- [14]. J. Tsitron, C. R. Kreller, P. K. Sekhar, R. Mukundan, F. H. Garzon, E. L. Brosha, A. V. Morozov, Bayesian decoding of the ammonia response of a zirconia-based mixed-potential sensor in the presence of hydrocarbon interference, *Sensors and Actuators B: Chemical*, Vol. 192, Issue 1, March 2014, pp. 283-293.
- [15]. U. Javed, K. P. Ramaiyan, C. R. Kreller, E. L. Brosha, R. Mukundan, A. V. Morozov, Using sensor arrays to decode NOx/NH₃/C₃H₈ gas mixtures for automotive exhaust monitoring, *Sensors and Actuators B: Chemical*, Vol. 264, Issue 1, July 2018, pp. 110-118.
- [16]. J. Tsitron, A. D. Ault, J. R. Broach, A. V. Morozov, Decoding complex chemical mixtures with a physical model of a sensor array, *PLoS Comput. Biol.* Vol. 7, Issue 10, October 2011, e1002224, pp. 1-13.
- [17]. R. Kocache, D. Holman, J. Swan, Sensor for combustible gases, US patent # 5,549,871, 1996.
- [18]. P. Shuk, P. Murphy, New advanced CO-sensor for the process application, in *Processing of the 1st International Conference on Sensing Technology*, Palmerston North, New Zealand, 21-23 November 2005, pp. 226-227.
- [19]. N. Taguchi, A metal oxide gas sensor, Japanese patent # 45-38200, 1962.
- [20]. T. Seiyama, A. Kato, K. Fujiishi, M. Nagatani, A new detector for gaseous components using semiconductive thin films, *Analytical Chemistry*, Vol. 34, Issue 11, 1962, pp. 1502-1503.
- [21]. N. Taguchi, A gas-detecting device, US patent # 3,631,436, 1971.
- [22]. Y. Shimizu, M. Egashira, Basic aspects and challenges of semiconductor gas sensors, *MRS Bulletin*, Vol. 24, Issue 6, 1999, pp. 18-24.
- [23]. D. E. Williams, Semiconducting oxides as gas-sensitive resistors, *Sensors and Actuators B: Chemical*, Vol. 57, Issues 1-3, 1999, pp. 1-16.
- [24]. N. Barsan, D. Koziej, U. Weimar, Metal oxide-based gas sensor research: How to?, *Sensors and Actuators B: Chemical*, Vol. 121, Issue 1, 2007, pp. 18-35.
- [25]. <http://www.figaro.co.jp/en/product/feature/combustible-gas-sensors.html>.
- [26]. A. Ponzoni, C. Baratto, N. Cattabiani, M. Falasconi, V. Galstyan, E. Nunez-Carmona, F. Rigoni, V. Sberveglieri, G. Zambotti, D. Zappa, Metal oxide gas sensors, a survey of selectivity issues addressed at the SENSOR Lab, Brescia (Italy), *Sensors*, Vol. 17, Issue 4, March 2017, pp. 1-27.
- [27]. M. Fleischer, H. Meixner, Selectivity in high-temperature operated semiconductor gas-sensors, *Sensors and Actuators B: Chemical*, Vol. 52, Issue 1, 1998, pp. 179-187.
- [28]. M. Fleischer, S. Kornely, T. Weh, J. Frank, H. Meixner, Selective gas detection with temperature-operated metal oxides using catalytic filters, *Sensors and Actuators B: Chemical*, Vol. 69, Issue 1, 2000, pp. 205-210.

[29]. E. D. Hinkley, P. L. Kelley, Detection of air pollutants with Tunable Diode Lasers, *Science*, Vol. 171, Issue 3972, 1971, pp. 635–639.

[30]. R. K. Hanson, Applications of quantitative sensor laser to kinetics, propulsion and practical energy systems, *Proceedings of the Combustion Institute*, Vol. 33, 2011, pp. 1-40.

[31]. C. S. Goldenstein, R. M. Spearrin, J. B. Jeffries, R. K. Hanson, Infrared laser-absorption sensing for combustion gases, *Progress in Energy and Combustion Science*, Vol. 60, May 2017, pp. 132-176.

[32]. R. K. Hanson, K. J. Whitty, Tunable diode laser sensors to monitor temperature and gas composition in high-temperature coal gasifiers, *DOE report DE- FE0001180, OSTI.GOV*, December 2014, pp. 1-39.

[33]. C. A. Almodovar, R. M. Spearrin, R. K. Hanson, Two-color laser absorption near 5 μm for temperature and nitric oxide sensing in high-temperature gases, *Journal of Quantitative Spectroscopy & Radiative Transfer*, Vol. 203, HITRAN2016 Special Issue, December 2017, pp. 572–581.




Published by International Frequency Sensor Association (IFSA) Publishing, S. L., 2019
(<http://www.sensorsportal.com>).



International Frequency Sensor Association (IFSA) Publishing

Jacob Y. Wong, Roy L. Anderson

Non-Dispersive Infrared Gas Measurement



Formats: printable pdf (Acrobat) and print (hardcover), 120 pages

ISBN: 978-84-615-9732-1,
e-ISBN: 978-84-615-9512-9

Written by experts in the field, the *Non-Dispersive Infrared Gas Measurement* begins with a brief survey of various gas measurement techniques and continues with fundamental aspects and cutting-edge progress in NDIR gas sensors in their historical development.

- It addresses various fields, including:
- Interactive and non-interactive gas sensors
- Non-dispersive infrared gas sensors' components
- Single- and Double beam designs
- Historical background and today's of NDIR gas measurements

Providing sufficient background information and details, the book *Non-Dispersive Infrared Gas Measurement* is an excellent resource for advanced level undergraduate and graduate students as well as researchers, instrumentation engineers, applied physicists, chemists, material scientists in gas, chemical, biological, and medical sensors to have a comprehensive understanding of the development of non-dispersive infrared gas sensors and the trends for the future investigation.

http://sensorsportal.com/HTML/BOOKSTORE/NDIR_Gas_Measurement.htm



Universal Frequency-to-Digital Converter (UFDC-1)

- 16 measuring modes: frequency, period, its difference and ratio, duty-cycle, duty-off factor, time interval, pulse width and space, phase shift, events counting, rotation speed
- 2 channels
- Programmable accuracy up to 0.001 %
- Wide frequency range: 0.05 Hz ... 7.5 MHz (120 MHz with prescaling)
- Non-redundant conversion time
- RS-232, SPI and I²C interfaces
- Operating temperature range -40 °C... +85 °C