

This paper describes objective technical results and analysis. Any subjective views or opinions that might be expressed in the paper do not necessarily represent the views of the U.S. Department of Energy or the United States Government.

*Exceptional service in the national interest*



# Calibration of Torque Instruments

Andrew Mackrory

Sandia National Laboratories



Sandia National Laboratories is a multimission laboratory managed and operated by National Technology & Engineering Solutions of Sandia, LLC, a wholly owned subsidiary of Honeywell International, Inc., for the U.S. Department of Energy's National Nuclear Security Administration under contract DE-NA0003525.

# Disclaimer

- Certain commercial equipment, instruments, or materials are identified in this training in order to adequately describe the procedures. Such identification does not imply recommendation or endorsement by the author(s) or Sandia National Laboratories, nor does it imply that the materials or equipment identified are the only or best available for the purpose.
- This training course and the course materials do not constitute an official interpretation of any standard.
- Any views or opinions expressed by the presenter(s) are their own and do not reflect the views of Sandia National Laboratories.

# Acknowledgements

- Sandia National Laboratories supported the preparation and delivery of this training. Special thanks to Matt McLane, Carlos Mascarenas, Jeff Domain, and Rick Mertes
- Pat Pierce of AKO, Inc. arranged for the loan of the TSD 650 P Portable Torque Wrench Test System

# Introductions


**Hello**  
my name is

# Learning Objectives

- Vocabulary, definitions
- National and international standards
- Development of procedures
  - Hands-on calibration of torque wrenches
- Measurement uncertainty
- Other types of torque instruments
  - Calibration of torque testers

**SAFETY:  
MORE THAN JUST YOUR OWN**


# Safety During Calibration



Buddy System



# Safety During Calibration



Buddy System



# Safety Due to Calibration




# Safety Due to Calibration



# **TORQUE FUNDAMENTALS: VOCABULARY AND DEFINITIONS**

# Forces



# Force is not Mass



**1 kg**  
**2.2 lbm**  
**9.8 N**  
**2.2 lbf**



**Mass**  
**(amount of matter)**  
**Weight**  
**(a force)**




**1 kg**  
**2.2 lbm**  
**1.6 N**  
**0.36 lbf**

# Force Conversion Factors

To convert from	to	Multiply by
kilogram-force (kgf)	newton (N)	<b>9.80665</b>
ounce(avoirdupois)-force (ozf)	newton (N)	0.2780139
pound-force (lbf)	newton (N)	4.448222
newton (N)	kilogram-force (kgf)	0.101972
newton (N)	ounce(avoirdupois)-force (ozf)	3.596942
newton (N)	pound-force (lbf)	0.2248089

# Pure Torque


Tap Wrench: Designed for **pure torque**



# Forces on a Torque Wrench

**Torque of  
fastener on  
wrench**

**Force of hand on wrench**



**Force of fastener on wrench**

# Forces on a Torque Wrench

If left hand was not used:

Force of hand on wrench




Undesired torque on wrench




Force of fastener on wrench

# Extensions and Attachments




# Effect on Calibration 1



$$T_i = \frac{T_a \times L_W}{L_E}$$

$$T_a = \frac{T_i \times L_E}{L_W}$$


# Effect on Calibration 2



$$T_i = \frac{T_a \times L_W}{L_E}$$

$$T_a = \frac{T_i \times L_E}{L_W}$$

# Extensions and Attachments



# Handle Extensions (“cheater bars”)




# Torque Conversion Factors

To convert from	to	Multiply by*
kilogram-force meter (kgf·m)	newton-meter (N·m)	<b>9.80665</b>
ounce (avoirdupois)-force inch (ozf·in)	newton-meter (N·m)	0.007061552
ounce (avoirdupois)-force inch (ozf·in)	millinewton-meter (mN·m)	7.061552
pound-force foot (lbf·ft)	newton-meter (N·m)	1.355818
pound-force inch (lbf·in)	newton-meter (N·m)	0.1129848

\* To convert the other way, divide by the number shown

# CALIBRATION FUNDAMENTALS

# What is Calibration?





# Traceability

**NIST**  
National Institute of  
Standards and Technology  
U.S. Department of Commerce


 Sandia  
National  
Laboratories

**NOTICE**  
**TRAINED**  
**PERSONNEL**  
**ONLY**



Each arrow represents a separate calibration certificate

# Measurement Uncertainty



## Generally Adequate:

**Test Uncertainty Ratio (T.U.R.) of 4:1 (or higher)**

**1.5:1 < T.U.R. < 4:1**

**→ Apply guard banding**

**T.U.R. < 1.5:1**

**→ Calibration system is inadequate**





# **NATIONAL AND INTERNATIONAL STANDARDS**

United States Department of Commerce  
National Institute of Standards and Technology



## Certificate of Accreditation to ISO/IEC 17025:2005

NVLAP LAB CODE: 105002-0

**Sandia National Laboratories**  
Albuquerque, NM


*is accredited by the National Voluntary Laboratory Accreditation Program for specific services,  
listed on the Scope of Accreditation, for:*

### **Calibration Laboratories**

*This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality  
management system (refer to joint ISO-ILAC-IAF Communiqué dated January 2009).*

2017-03-01 through 2018-03-31

*Effective Dates*



For the National Voluntary Laboratory Accreditation Program

# Certificate Requirements

Mechanical Calibration Laboratory  
 Sandia National Laboratories, Albuquerque, New Mexico 87185-0663  
 Accredited by the National Voluntary Laboratory Accreditation Program for the scope of accreditation under Lab Code 105002-0



②  
 ③  
 ① **Calibration Certificate**  
 Document #: SNL-9484\_11667296

### Item Identification

Asset Number SNL-9484  
 Description Wrench, Torque  
 Model 6062C  
 Serial # DPG34594  
 Manufacturer Proto Tools  
 Customer Asset Id  
 Purchase Order  
 Customer N/A



Custodian  
 Location  
 Date of Receipt April 26, 2016  
 Dates Tested (Start - End) May 03, 2016 - May 03, 2016  
 Date Approved May 03, 2016  
 Calibration Expiration Date November 03, 2017

**Calibration Description**  
 Calibration Lab MECH  
 Calibration Procedure, rev. MCP-1059 VERSION 3, Torque Calibration  
 Temperature 68 deg F  
 Humidity 10-80 %RH  
 Barometric Pressure N/A mmHg  
 As Found Condition PASS  
 As Left Condition PASS  
 Software Used None  
 Tamper Seal None

④ Page 1 of 3  
 SNL-9484\_11667296

Mechanical Calibration Laboratory



Sandia National Laboratories, Albuquerque, New Mexico 87185-0663  
 Accredited by the National Voluntary Laboratory Accreditation Program for the scope of accreditation under Lab Code 105002-0

⑬

### Calibration Results

The as found/as left test results are within the tolerance(s) listed below

SNL-9484	Test point	Setting	Reading CW (lb-in)	Tolerance CW (+/- 4% of setting)	Reading CCW (lb-in)	Tolerance CCW (+/- 6% of setting)
Range	200			41.60	41.60	42.40
Units	lb-in	20%	40	38.40		37.60
CW Tolerance	4%	60%	118.07	124.80	120.25	127.20
CCW Tolerance	6%			112.00		112.80
Standard SNL-1351	100%	200	203.33	208.00	203.53	212.00
Minimum Torque (20 lb-in)				192.00		188.00

Notes: Torque wrench was calibrated in CW and CCW directions.

⑯

### Uncertainty Statement

The expanded uncertainty for the calibration of the item identified above was determined from a combined standard uncertainty with a coverage factor of k=2 providing a level of confidence of 95%. The expanded uncertainty is estimated to be +/- 0.40% of the indicated value of the standard.

When 4:1 test accuracy ratio is not satisfied, guard banding will be applied to determine acceptance.

The uncertainty value(s) reported pertain to the calibration test(s) performed and should not be confused with the accuracy specifications.

### Equipment (Standard) Used

Asset #	Description	Model	Expires
SNL-1351	Calibrator, Torque	TSD111/050	November 12, 2016



Page 1 of 3  
 SNL-9484\_11667296

Mechanical Calibration Laboratory



Sandia National Laboratories, Albuquerque, New Mexico 87185-0663  
 Accredited by the National Voluntary Laboratory Accreditation Program for the scope of accreditation under Lab Code 105002-0

⑫

### Traceability

Values and the associated uncertainties supplied by the Primary Standards Lab (PSL) are traceable to the SI through one or more of the following:

1. Reference standards whose values are disseminated by the National Institute of Standards and Technology (United States of America) or, where appropriate, to the national metrological institute of another nation participating in the CIPM MRA;
2. Reference standards whose values are disseminated by a laboratory that has demonstrated competence, measurement capability, and traceability for those values;
3. The accepted value(s) of fundamental physical phenomena (intrinsic standards);
4. Ratio(s) or other non-maintained standards established by either a self-calibration and/or a direct calibration technique;
5. Standards maintained and disseminated by the PSL in special cases and where warranted, such as consensus standards where no national or international standards exist;

Note 1: This certificate or report shall not be reproduced except in full, without the advance written approval of the Primary Standards Lab at Sandia National Laboratories.

Note 2: For National Voluntary Laboratory Accreditation Program (NVLAP) accredited capabilities, the PSL at Sandia National Laboratories is accredited by NVLAP for the specific scope of accreditation under Laboratory Code 105002-0. This certificate or report shall not be used by the customer to claim product endorsement by NVLAP, the Primary Standards Laboratory, Sandia National Laboratories or any agency of the U. S. Government.

Note 3: The as received condition of the standard, set of standards, or measurement equipment described herein was as expected, unless otherwise noted in the body of the certificate or report.

Note 4: The presence of names and titles under "Authorization" are properly authenticated electronic signatures conforming to the equivalent identification signature requirements of ISO 17025:2003 5.10.2.1.

### Authorization

Calibrated By:

Leyva, David M.  
 Metrologist

Approved By:

Bryce, Edwin A.  
 Technical Project Leader

⑭

End-of-Document

Page 3 of 3  
 SNL-9484\_11667296

# Certificate Requirements

Mechanical Calibration Laboratory  
 Sandia National Laboratories, Albuquerque, New Mexico 87185-0663  
 Accredited by the National Voluntary Laboratory Accreditation Program for the scope of accreditation under Lab Code 105002-0

②  
 ③  
 ① **Calibration Certificate**  
 Document #: SNL-9484\_11667296



### Item Identification

Asset Number SNL-9484  
 Description Wrench, Torque  
 Model 6062C  
 Serial # DPG34594  
 Manufacturer Proto Tools  
 Customer Asset Id  
 Purchase Order  
 Customer



Custodian  
 Location  
 Date of Receipt April 26, 2016  
 Dates Tested (Start - End) May 03, 2016 - May 03, 2016  
 Date Approved May 03, 2016  
 Calibration Expiration Date November 03, 2017

⑩  
 ⑪

### Calibration Description

Calibration Lab MECH  
 Calibration Procedure, rev. MCP-1059 VERSION 3, Torque Calibration  
 Temperature 68 deg F  
 Humidity 10-80 %RH  
 Barometric Pressure N/A mmHg  
 As Found Condition PASS  
 As Left Condition PASS  
 Software Used None  
 Tamper Seal None

⑬  
 ⑬  
 ⑬

④ Page 1 of 3  
 SNL-9484\_11667296

Mechanical Calibration Laboratory  
 Sandia National Laboratories, Albuquerque, New Mexico 87185-0663  
 Accredited by the National Voluntary Laboratory Accreditation Program for the scope of accreditation under Lab Code 105002-0



⑬

### Calibration Results

The as found/as left test results are within the tolerance(s) listed below

SNL-9484	Test point	Setting	Reading CW (lb-in)	Tolerance CW (% of setting)	Reading CCW (lb-in)	Tolerance CCW (% of setting)
Range	200		40.25	±1.60	41.60	±2.40
Units	lb-in	20%	40	±8.00	41.60	±12.00
CW Tolerance	4%	60%	118.07	±4.72	120.25	±4.81
CCW Tolerance	6%			±6.48		±7.22
Standard SNL-1351	100%	200	203.33	±8.13	203.53	±12.20
Minimum Torque (20 lb-in)				±192.00		±188.00

⑬

### Uncertainty Statement

The expanded uncertainty for the calibration of the item identified above was determined from a combined standard uncertainty with a coverage factor of k=2 providing a level of confidence of 95%. The expanded uncertainty is estimated to be ±0.40% of the indicated value of the standard.

When 4:1 test accuracy ratio is not satisfied, guard banding will be applied to determine acceptance.

The uncertainty value(s) reported pertain to the calibration test(s) performed and should not be confused with the accuracy specifications.

### Equipment (Standard) Used

Asset #	Description	Model	Expires
SNL-1351	Calibrator, Torque	TSD111/050	November 12, 2016

⑬

Page 1 of 3  
 SNL-9484\_11667296

Mechanical Calibration Laboratory  
 Sandia National Laboratories, Albuquerque, New Mexico 87185-0663  
 Accredited by the National Voluntary Laboratory Accreditation Program for the scope of accreditation under Lab Code 105002-0



⑬

### Traceability

Values and the associated uncertainties supplied by the Primary Standards Lab (PSL) are traceable to the SI through one or more of the following:

- Reference standards whose values are disseminated by the National Institute of Standards and Technology (United States of America) or, where appropriate, to the national metrological institute of another nation participating in the CIPM MRA;
- Reference standards whose values are disseminated by a laboratory that has demonstrated competence, measurement capability, and traceability for those values;
- The accepted value(s) of fundamental physical phenomena (intrinsic standards);
- Ratio(s) or other non-maintained standards established by either a self-calibration and/or a direct calibration technique;
- Standards maintained and disseminated by the PSL in special cases and where warranted, such as consensus standards where no national or international standards exist;

Note 1: This certificate or report shall not be reproduced except in full, without the advance written approval of the Primary Standards Lab at Sandia National Laboratories.

Note 2: For National Voluntary Laboratory Accreditation Program (NVLAP) accredited capabilities, the PSL at Sandia National Laboratories is accredited by NVLAP for the specific scope of accreditation under Laboratory Code 105002-0. This certificate or report shall not be used by the customer to claim product endorsement by NVLAP, the Primary Standards Laboratory, Sandia National Laboratories or any agency of the U. S. Government.

Note 3: The as received condition of the standard, set of standards, or measurement equipment described herein was as expected, unless otherwise noted in the body of the certificate or report.

Note 4: The presence of names and titles under "Authorization" are properly authenticated electronic signatures conforming to the equivalent identification signature requirements of ISO 17025:2003 5.10.2.1.

### Authorization

Calibrated By:  
 Leyva, David M.  
 Metrologist

⑬

Approved By:  
 Bryce, Edwin A.  
 Technical Project Leader

End-of-Document

Page 3 of 3  
 SNL-9484\_11667296

# ASME B107.300-2010 (R2016)


- Torque Instruments
  - Hand Torque Tools (mechanical)
  - Electronic Torque Instruments
  - Electronic Tester, Hand Torque Tools

- Available from:

[https://www.asme.org/products/codes-standards/b107300-2010-torque-instruments-\(1\)](https://www.asme.org/products/codes-standards/b107300-2010-torque-instruments-(1))

- \$43
- The purpose of the standard is to:
  - “define essential performance and safety characteristics”
  - “specify test methods to evaluate performance” (for example, calibration to evaluate accuracy)

# Torque Instrument Classification



# Torque Instrument Classification



①



②



③






④


# Other Torque Wrench Features




# ASME B107.300 – Other Content



$\pm 4\%$   
 $\pm 6\%$





# ISO 6789 Parts 1 and 2 (2017)



## COMPARISON OF SCOPE

# EURAMET cg-14



# DEVELOPING CALIBRATION PROCEDURES

# Step 1: Calibration Specifications



## TORQUE SYSTEM ACCURACY:

When TSD Transducer (s) are connected to and calibrated with approved Torque Transducer Calibration Systems and procedures, the minimum expected accuracy is +/- 1 division or count from 0 to 10% of the Transducer range and 0.1% IV (of indicated value) from 10% thru 100% of the Transducer Range. Each TSD Transducer has a minimum of 10,000 divisions or counts with full 12,000 and 16,000 count conversions.

**A.K.O. Inc.**

# Calibration Data Sheet

## Torque Instrument Calibration Data Sheet

Example for NCSLI Technical Exchange, February 2018

### Unit Under Test

Manufacturer: \_\_\_\_\_  
 Model: \_\_\_\_\_  
 Serial Number: \_\_\_\_\_  
 Other Identifiers: \_\_\_\_\_  
 Condition, as received: \_\_\_\_\_  
 ASME Tool Classification: \_\_\_\_\_  
 Maximum Graduated Value: \_\_\_\_\_  
 Units: \_\_\_\_\_  
Accuracy over 20 % to 100% of Rated Capacity  
 CW: \_\_\_\_\_ % IV  
 CCW: \_\_\_\_\_ % IV

### Calibration System

Manufacturer: \_\_\_\_\_  
 Model: \_\_\_\_\_  
 Serial Number: \_\_\_\_\_  
 ASME Classification: \_\_\_\_\_

### Transducers Used, Low to High Torque

Manufacturer:	_____	_____
Model:	_____	_____
Serial Number:	_____	_____
Calibration Date:	_____	_____
Upper Range of Use:	_____	_____
Units:	_____	_____
Accuracy over Range of Use (% IV):	_____	_____
Worst T.U.R. (CW and CCW):	_____	_____

### Test Conditions

Date: \_\_\_\_\_  
 Procedure Used: ASME B107.300-2010 (B107.14)  
 Lab Temperature: \_\_\_\_\_ °F  
 Lab Humidity: \_\_\_\_\_ %RH

CW = clockwise  
 CCW = counterclockwise  
 IV = indicated value  
 MIC = maximum instrument capacity

Tol. = tolerance  
 T.U.R. = test uncertainty ratio (> 4:1 required)  
 %RH = percent relative humidity

This data sheet was prepared by Andrew Mackrory (Sandia National Laboratories) for instructional use only at the NCSLI Technical Exchange, February 2018. The author, Sandia National Laboratories, and NCSLI accept no liability for the use of this data sheet. It is the responsibility of class participants to check the data sheet for accuracy and conformance to applicable standards.

### Calibration Process and Data - As Found

Tool set to Maximum Graduated Value and operated 3 times in calibration direction (CW or CCW).

Tool set to Minimum Graduated Value and then set to first test point.

All setting adjustments are made by approaching target test value from below.

Tool operated 3 times at each test point before recording data.

Torque immediately removed from tool after taking a reading.

Torque Setting - CW	Tol. Limits			Readings,			Max Deviation	RESULT
% of MIC	0	0	0	4	5	6	% of Tolerance	
20%	0	0	0					
60%	0	0	0					
100%	0	0	0					

Torque Setting - CCW	Tol. Limits			Readings,			Max Deviation	RESULT
% of MIC	0	0	0	4	5	6	% of Tolerance	
20%	0	0	0					
60%	0	0	0					
100%	0	0	0					

Adjustments Made:

### Calibration Data - As Left

Torque Setting - CW	Tol. Limits			Readings,			Max Deviation	RESULT
% of MIC	0	0	0	4	5	6	% of Tolerance	
20%	0	0	0					
60%	0	0	0					
100%	0	0	0					

Torque Setting - CCW	Tol. Limits			Readings,			Max Deviation	RESULT
% of MIC	0	0	0	4	5	6	% of Tolerance	
20%	0	0	0					
60%	0	0	0					
100%	0	0	0					

Condition, as left: \_\_\_\_\_

Notes: \_\_\_\_\_  
(exclusions/additions/deviations)

Metrologist: \_\_\_\_\_

Unit Under Test (Make/Model/Serial): / /

# Calibration Data Sheet

## Torque Instrument Calibration Data Sheet

Example for NCSLI Technical Exchange, February 2018

### Unit Under Test

Manufacturer: Craftsman  
 Model: unknown  
 Serial Number: 4051182296  
 Other Identifiers: N/A

Condition, as received: Set to 25 ft lb, graduation markings on handle nearly rubbed off, handle graduations do not line up with scale.

ASME Tool Classification: Type II (Setting), Class A (graduated), Style 2 (ratcheting)  
 Maximum Graduated Value: 75  
 Units: ft-lb

### Accuracy over 20 % to 100% of Rated Capacity

CW: 4% % IV  
 CCW: 6% % IV

### Calibration System

Manufacturer: AKO, Inc.  
 Model: TSD 650-P  
 Serial Number: 1138  
 ASME Classification: Type II (Hand actuated), Class A (0.25% accuracy)

### Transducers Used, Low to High Torque

	AKO, Inc.	AKO, Inc.
Manufacturer:	AKO, Inc.	AKO, Inc.
Model:	TSD 821	TSD 821
Serial Number:	6272	6272
Calibration Date:	22-Jan-2018	22-Jan-2018
Upper Range of Use:	30	300
Units:	ft-lb	ft-lb
Accuracy over Range of Use (% IV):	0.13%	0.10%
Worst T.U.R. (CW and CCW):	30.8:1	40:1

### Test Conditions

Date: 9-Feb-2018  
 Procedure Used: ASME B107.300-2010 (B107.14)  
 Lab Temperature: 72 °F  
 Lab Humidity: 40 %RH

CW = clockwise  
 CCW = counterclockwise  
 IV = indicated value  
 MIC = maximum instrument capacity

Tol. = tolerance  
 T.U.R. = test uncertainty ratio (> 4:1 required)  
 %RH = percent relative humidity

This data sheet was prepared by Andrew Mackrory (Sandia National Laboratories) for instructional use only at the NCSLI Technical Exchange, February 2018. The author, Sandia National Laboratories, and NCSLI accept no liability for the use of this data sheet. It is the responsibility of class participants to check the data sheet for accuracy and conformance to applicable standards.

### Calibration Process and Data - As Found

Tool set to Maximum Graduated Value and operated 3 times in calibration direction (CW or CCW).

Tool set to Minimum Graduated Value and then set to first test point.

All setting adjustments are made by approaching target test value from below.

Tool operated 3 times at each test point before recording data.

Torque immediately removed from tool after taking a reading.

Torque Setting - CW		Tol. Limits		Readings, ft-lb			Max Deviation	RESULT
% of MIC	ft-lb	ft-lb	ft-lb	4	5	6	% of Tolerance	
20%	15	14.4	15.6	14.88	14.76	14.74	43%	PASS
60%	45	43.2	46.8	44.88	44.78	44.78	12%	PASS
100%	75	72	78	75.74	76.04	75.78	35%	PASS

Torque Setting - CCW		Tol. Limits		Readings, ft-lb			Max Deviation	RESULT
% of MIC	ft-lb	ft-lb	ft-lb	4	5	6	% of Tolerance	
20%	15	14.1	15.9	15.82	15.64	15.74	91%	PASS
60%	45	42.3	47.7	46.66	46.4	46.46	61%	PASS
100%	75	70.5	79.5	77.9	77.8	77.84	64%	PASS

Adjustments Made: No

### Calibration Data - As Left

Torque Setting - CW		Tol. Limits		Readings, ft-lb			Max Deviation	RESULT
% of MIC	ft-lb	ft-lb	ft-lb	4	5	6	% of Tolerance	
20%	15	14.4	15.6	N/A	N/A	N/A		
60%	45	43.2	46.8	N/A	N/A	N/A		
100%	75	72	78	N/A	N/A	N/A		

Torque Setting - CCW		Tol. Limits		Readings, ft-lb			Max Deviation	RESULT
% of MIC	ft-lb	ft-lb	ft-lb	4	5	6	% of Tolerance	
20%	15	14.1	15.9	N/A	N/A	N/A		
60%	45	42.3	47.7	N/A	N/A	N/A		
100%	75	70.5	79.5	N/A	N/A	N/A		

Condition, as left: Set to 10 ft-lb (minimum graduated value)

Notes: MIC taken from foot-pounds scale, which has slightly lower range than Nm scale. (exclusions/additions/deviations) Wrench not adjusted per customer request.

Metrologist: Andrew Mackrory 9-Feb-2018


Unit Under Test (Make/Model/Serial): Craftsman / unknown / 4051182296

# HANDS-ON CALIBRATION




# Why Exercise a Wrench?

CW Torque Values compared to Tolerance  
(50 successive loadings)




# Middle of the Handle

Effect of the Position of the Reaction Post




# Temperature

## Effect of Wrench Temperature




# Speed of Torque Application

Effect of Speed on Screwdriver Grip Tool  
Limiting Torque (Setting = 3 Nm)



# ADJUSTMENT

# Don't chase your tail...



# DATA ANALYSIS: PASS OR FAIL?


# Pass or Fail?

**For a Type II, Class A, Style 1 Wrench, determine the result of calibration:**

<b>Wrench Setting (lbf-ft)</b>	<b>Measured Torque (lbf-ft)</b>	<b>Determine PASS/FAIL</b>
<b>100</b>	<b>103.6</b>	
<b>100</b>	<b>95.7</b>	
<b>60</b>	<b>62.8</b>	

# CALIBRATION OF TORQUE TESTERS







# MEASUREMENT UNCERTAINTY FOR TORQUE WRENCH CALIBRATION

# Sources of Uncertainty




# What Does the Customer Need?



# Gage R&R

# 3x



# Uncertainty Budget

Quantity	evaluation of standard uncertainty	standard uncertainty in N·m	relative standard uncertainty in %
Repeatability in unchanged mounting position $b'$	type A	$u_{b'} = \frac{b'}{S\sqrt{2}}$	$w_{b'} = \frac{b'}{\sqrt{2}} \cdot \frac{100}{\bar{X}}$
Reproducibility in different mounting positions $b$	type A	$u_b = \frac{b}{S\sqrt{n}}$	$w_b = \frac{b}{\sqrt{n}} \cdot \frac{100}{\bar{X}}$
Deviation resulting from fitting curve $f_a$	type B with triangular distribution	$u_{f_a} = \frac{ f_a }{S\sqrt{6}}$	$w_{f_a} = \frac{ f_a }{\sqrt{6}} \cdot \frac{100}{X_a}$
Resolution $r$	type B with rectangular distribution	$u_r = \frac{r}{\sqrt{12}}$	$w_r = \frac{r}{\sqrt{12}} \cdot \frac{100}{M_k}$
Reference torque	type B	$u_{tcm}$	$w_{tcm}$

# Uncertainty Statement

The single biggest problem in communication is the illusion that it has taken place

- George Bernard Shaw

# SETTING A CALIBRATION INTERVAL


# Calibration Intervals



# THE ANSWERS IN THE BACK

# Pure Torque

Tap Wrench: Designed for **pure torque**



Calculating the torque on the tap:

1. Select the center of tap as the axis
2.  $1.5 \text{ lbf} \times 2 \text{ in} = +3 \text{ lbf-in}$
3.  $1.5 \text{ lbf} \times 2 \text{ in} = \underline{+3 \text{ lbf-in}}$
4. Total:  $+6 \text{ lbf-in}$

# Pass or Fail?

For a Type II, Class A, Style 1 Wrench, determine the result of calibration:

Wrench Setting (lbf-ft)	Measured Torque (lbf-ft)	Determine PASS/FAIL
100	103.6	PASS
100	95.7	FAIL
60	62.8	FAIL

# GUARD BANDING

# T.U.R. and Guard Banding

- You need to calibrate something to  $\pm 1\%$  of reading, your calibration system uncertainty is  $\pm 0.4\%$  of reading
- Calculate the T.U.R.

$$T.U.R. = \frac{U_{certification}}{U_{cal.process}}$$

$$T.U.R. = \frac{1}{0.4} = 2.5$$

1.5:1 < T.U.R. < 4:1

→ apply guard  
banding

# T.U.R. and Guard Banding

- The guard banding equation to give new limits for acceptance:

$$\textit{Adjusted Limit} = \sqrt{1 - \frac{1}{T.U.R.^2}} \times \textit{Limit}$$

- Calculate the limits for acceptance for the example on the last slide:

$$\textit{Adjusted Limit} = \sqrt{1 - \frac{1}{2.5^2}} \times \underline{1\% \text{ of reading}}$$

$$\textit{Adjusted Limit} = \underline{0.91} \% \text{ of reading}$$

# T.U.R. and Guard Banding

*Adjusted Limit* = 0.91 % of reading

Item Setpoint (units)	Measured Value (units)	Allowable Deviation (units)	Determine PASS/FAIL
100	101	0.91	FAIL
100	99.2	0.91	PASS
50	50.5	0.455	FAIL
50	50.4	0.455	PASS