



Primary Standards Laboratory

Metrology Program

Fact Sheet

Mass and Force

The Primary Standards Laboratory (PSL) maintains a variety of primary mass and force standards to assure accurate and traceable measurements for its customers. All primary mass and force standards are directly traceable to the National Institute of Standards and Technology (NIST). Mass certifications are available from 1 milligram to 64 kilograms.

When the highest accuracy is needed, groups of weights are intercompared using specified sequences that have high redundancy. A certification is issued giving the mass and associated uncertainty of each weight in the customer set.

Force is measured either against dead weight-certified load cells or directly by dead weight. Force can be measured in both compression and tension.

•MASS – BEST UNCERTAINTIES

Meet or exceed OIML Class E1 and ANSI/ASTM Classes 1, 1.1 tolerances up to 64 kg

•FORCE – BEST UNCERTAINTIES

Proving rings and load cells to 1000 lbf \pm 20 ppm
Proving rings and load cells 1000-100,000 lbf \pm 200 ppm

Major Resources

- State-of-the-art laboratory environmental controls
- 10,000 lbf, 30,000 lbf, and 100,000 lbf Morehouse load frames
- 1,000 lbf Morehouse dead weight tester
- Mass comparators (partial list):
 - Mettler AX64004 - 100 μ g resolution, 64 kg capacity
 - Mettler KA30-3/P - 2 mg resolution, 30 kg capacity
 - Sartorius CC10000U-L - 10 μ g resolution, 10 kg capacity
 - Mettler AX1006AH - 1 μ g resolution, 1 kg capacity
 - Mettler AT106 - 1 μ g resolution, 100 g capacity
 - Mettler UMT6 - 0.1 μ g resolution, 6 g capacity
- Electronic top-loading balances to 60 kg capacity

Selected Accomplishments



SAND2008-???? Unlimited Release. Sandia is a multiprogram laboratory operated by Sandia Corporation, a Lockheed Martin Company, for the United States Department of Energy's National Nuclear Security Administration under contract DE-AC04-94AL85000.

- NIST/NVLAP (National Voluntary Laboratory Accreditation Program), accredited under NVLAP lab code 105002-0. For full details see <http://ts.nist.gov/standards/scopes/1050020.pdf>
- Successful participation in mass and force round robins administered by the NCSL, NASA, and Lockheed Martin



Automatic Weighing of Kilogram Reference Standards

Contacts

Hy D. Tran, PhD, PE
Sandia National Laboratories
P. O. Box 5800; M/S 0665
Albuquerque, NM 87185
Phone: (505) 844-5417
FAX: (505) 844-4372
Email: hdtran@sandia.gov

James E. Pacheco
Sandia National Laboratories
P. O. Box 5800; M/S 0665
Albuquerque, NM 87185
Phone: (505) 844-9175
FAX: (505) 844-4372
Email: jepache@sandia.gov

