



Primary Standards Laboratory Metrology Program

Fact Sheet

Acceleration and Shock

The Primary Standards Laboratory (PSL) maintains a wide variety of primary acceleration and shock standards to assure accurate and traceable measurements for customers. Capabilities include acceleration measurements to 50 kHz for a wide variety of accelerometers and shock pulses to 10,000 g.

While the automated vibration calibration system is designed to operate over a range of 1 Hz to 50 kHz, traceable calibrations are performed from 2 Hz to 10 kHz. Test parameters and data presentation instructions can be programmed for each calibration.



UD Automated Vibration Measurement System

The system will measure capacitance for piezoelectric accelerometers, output bias voltage for internal electronic accelerometers and resistance, zero measurand output, excitation voltage, damping, natural or resonant frequency, shunt, and simulated electrical calibration for piezoresistive accelerometers.

Traceability is established through direct comparison with a NIST-calibrated accelerometer.

Capabilities

Below is a representative sample of our uncertainties. We are NVLAP accredited under Lab Code 105002-0 by the National Institute of Standards and Technology/National Voluntary Laboratory Accreditation Program (NIST/NVLAP) in most of our capabilities. For full details see <http://ts.nist.gov/standards/scopes/1050020.pdf>

Acceleration

1 Hz to 10 kHz $\pm 2.5\%$

Shock

100 to 10,000 g $\pm 4\% \text{ to } 8\%$

Major Resources

- Automated vibration measurement system
- Measurement of acceleration over temperature range -40°C to $+66^\circ \text{C}$
- Measurement of linear acceleration from 10 to 50,000 m/sec^2

Contacts

David A. Sanchez

Sandia National Laboratories
P. O. Box 5800;M/S 0665
Albuquerque, NM 87185
Phone: (505) 844-4439
FAX: (505) 844-4372
Email: dasanch@sandia.gov

James E. Pacheco

Sandia National Laboratories
P. O. Box 5800;M/S 0665
Albuquerque, NM 87185
Phone: (505) 844-9175
FAX: (505) 844-4372
Email: jepache@sandia.gov