

Dosimetry User's Perspective on Covariance Needs

Presentation to: Workshop on Neutron Cross Section Covariances

At: Port Jefferson, NY

On: June 24-29, 2008

Session 1: User's Perspective

P. J. Griffin

June 24, 2008

Sandia is a multiprogram laboratory operated by Sandia Corporation, a Lockheed Martin Company,
for the United States Department of Energy under contract DE-AC04-94AL



Sandia National Laboratories





Outline

- **Scope of dosimetry applications**
- **Major dosimetry concerns**
- **Details of dosimetry needs**
- **Balance of dosimetry needs**
- **Path forward**



Sandia National Laboratories



Application Priorities

- **Important Dosimetry Applications**
 - ◆ Neutron spectrum adjustment
 - ◆ Fluence monitors
 - ◆ Material identification
 - ◆ Secondary gamma environments



Sandia National Laboratories



Application 1: Spectrum Adjustment


- **Methods**
 - Iterative, e.g. SAND-II, GRAVEL
 - Least squares, e.g. LSL-M2, STAY'SL
 - Maximum entropy, e.g. MAXED
 - Bacchus-Gilbert, e.g. UFO
- **How important are cross section covariance matrices?**
 - Answer varies with specific application
 - Previous REAL-80/84 series of comparison indicates this is very important
- **How important is trial spectrum**
 - Recent work suggests it can be very important
 - So, what about covariance for trial spectrum?



Sandia National Laboratories



Why has ENDF/B-VII abandon us? Walled us out?



18/04/2007-227 © INKCINCT Cartoons www.inkcinct.com.au



Sandia National Laboratories



CSEWG ENDF/B-VII Issues


- You took away almost ALL of our covariance data
 - Dosimetry reactions of interest to us were generally eliminated
 - Even the standard reactions that we use have a covariance only over a restricted energy range
- We need to tell users why we do not use/recommend ENDF/B-VII data
 - Telling them we recommend use of the ENDF/B-VI dosimetry cross section, which is the same as the ENDF/B-VII values, starts a very confusing dialog





CSEWG Rating of Older ENDF/B-VI Covariance Data

- CSEWG followed a rigorous process for evaluating older data
- Problem is the results of the rating:
 - We clearly do not want wrong/bad data
 - But we do not want NO data either
- Maybe the bar was set too high
 - Suggest accepting yellow and green ratings and only eliminating red rated reactions
 - It is understood that an over-estimate of the uncertainty can be almost as bad as an under-estimate



Sandia National Laboratories



What is our real request?

- We are not, in general, requesting smaller uncertainties – just an estimate of the current uncertainty.
 - All covariance data does not have to be of “standard” quality. There is a role for “reference” quality data.
 - The covariance must be consistent with the actual cross section – and not an independent quantity. “controlled” quality data is not useful.
- Why even give us the cross section data if you really have no estimate of the uncertainty?
 - If you have an “expert judgment” we will even go with that – but we can not use data with unstated uncertainty.
 - Dosimetry users can not be trusted to provide this “expert judgment”. The evaluator or evaluation community must provide this.



Sandia National Laboratories



Covariance for ^{235}U Fission Spectrum

- Motivation

- Good covariance data has only been published for ^{252}Cf . ^{235}U covariance will drive the *a priori* spectrum covariance
- Two parameter Watt fission representation inadequate. Replacement will also be used in updated radiation transport calculations

- Deficiency

- For 2-parameter shape (highly correlated) a single good threshold sensor can result in an unrealistically small *posteriori* uncertainty for high energy spectrum tail.

- Request

- Covariance for ^{235}U fast fission neutron spectrum of similar detail (not necessarily accuracy) as in ^{252}Cf spontaneous fission.



Sandia National Laboratories



natCd Request

- **Motivation**

- Cd is used as a cover for activation foils
- It moves the dosimeter response above thermal energy region

- **Deficiency**

- Systematic offsets in spectrum adjustments are seen in epi-Cadmium region
- Is this an issue with the Cd cross section here? the *a priori* spectrum/uncertainty? a biased adjustment process for covers?

- **Request**

- Uncertainty/covariance data so that we can evaluate where the cause of the adjustment problem



Sandia National Laboratories



New Cover Materials

- **^{nat}Gd is now being used as a foil cover**
 - ◆ Not as good a cover for shifting energy response, but it avoids ES&H issues associated with ^{nat}Cd
 - ◆ New ENDF/B-VII cross sections are now available, and has high/fast energy ($> 1 \text{ keV}$) covariance data for total, elastic, and (n,γ) reactions.
 - ◆ Analysis of cover correction with uncertainty is pending.
 - » What do we do for covariance/uncertainty for energies $< 10 \text{ keV}$.



Sandia National Laboratories



Treatment of Activation Measurements

- **How important is this covariance?**
 - Still under investigation
 - New applications are developing a refined approach that will be used to test importance
 - **Uncertainty contributions:**
 - » HPGe efficiency calibration curve
 - » Energy of gamma line read
 - » Multiple reactions from same foil
 - » Multiple readings on different counters
 - » Sampling uncertainty
 - » Summing corrections
 - » Detector design – e.g. dead oxide layer for low energy photons, sensitive detector volume
 - **Correlation between bare and covered cross sections is important**
 - » Treat as two/three uncorrelated reactions; B4C covered; Cd-B4C covered; bare-Cd-B4C covered



Sandia National Laboratories



PKA Recoil Uncertainty for Si, GaAs, and Fe

- **Motivation**

- Standards require uncertainty for exposure metrics
 - » Fe dpa for PWR/LWR material embrittlement
 - » Si displacement kerma for electronic device gain degradation
 - » GaAs for displacement kerma and damage deviation due to FP recombination in clusters

- **Deficiency**

- Cross section uncertainty for all reactions in Si, Fe, Ga, and As
- Recoil spectrum definition – File 5/6 – and uncertainty

- **Request**

- Cross section covariance data for these materials
- Reaction-dependent recoil particle energy uncertainty – maybe covariance – File 35
- Sufficiency of File 35 format not clear – need data for these materials to look at impact



Sandia National Laboratories



Special Dosimetry User Needs

- Attach to ENDF File 1 data the relative abundance for isotope
 - ◆ Critically important for ^{58}Fe , an important dosimetry reaction, where the abundance has varied significantly over the years – and users have no idea how to combine the abundance they assume in their activation analysis with that used for the cross section evaluation
- Link to state-of-the-art decay data (gamma decay energies, photon yield, and branching ratios)
 - ◆ Independent of cross section, so it can be provided by dosimetry community in separate document



Sandia National Laboratories



Dosimetry Need – ~ 1-MeV Response Sensor

- Motivation

- ◆ Fast neutron (0.1 – 3 MeV) damage dominates many damage mechanisms. This energy does not have good, easy-to-read dosimetry reactions.

- Deficiency

- ◆ Current – inadequate – sensors:
 $^{237}\text{Np}(\text{n},\text{f})$, $^{93}\text{Nb}(\text{n},\text{n}')$, $^{103}\text{Rh}(\text{n},\text{n}')$, $^{115}\text{In}(\text{n},\text{n}')$, Si transistor gain degradation

- Request

- ◆ Cross sections for alternate candidate reactions



Sandia National Laboratories



Additional Need: $^{58}\text{Ni}(\text{n},\text{p})^{58\text{m}}\text{Co}$ BR Uncertainty

- Motivation

- $^{58}\text{Ni}(\text{n},\text{p})^{58\text{m}}\text{Co}$ reaction is an important monitor foil for research reactors. It has a high energy sensitivity (> 3-MeV) that makes it ideal.

- Deficiency

- The reading of this foil must be delayed due to decay of $^{58}\text{Ni}(\text{n},\text{p})^{58\text{m}}\text{Co}$ product

- Request

- Energy-dependent branching ratios for this reaction to the ground and metastable state – and uncertainty in the branching ratio

- » Hetrick personal communication currently used for this early-reading, but no uncertainty data is available.



Sandia National Laboratories



Dosimetry Need – High Energy Cross Sections

- **Motivation**

- Fusion community needs improved dosimetry for material damage. Simulators (IFMIF) have neutron components up to ~60 MeV

- **Deficiency**

- Need high fidelity cross sections and covariance data in this range
 - Highly correlated smooth cross sections provide little sensitivity for spectrum adjustments

- **Request**

- Candidate reactions have been identified. Covariance data required.



Sandia National Laboratories



Related Issues

- Diagonal covariance matrices have been accepted – but with great reluctance.
- Too large of a cross section uncertainty will permit the spectrum adjustment to be driven by the *a priori* spectrum.
 - ◆ *a priori* spectrum uncertainties and covariance are much more poorly known/defined
- Chi-squared per degree-of-freedom is an important metric in the spectrum adjustment. Poor covariance estimates defeat the value of this metric.



Sandia National Laboratories



Role for Low Fidelity Covariance Data

- Low fidelity effort focuses on use of parametric variation in nuclear physics models, e.g. EMPIRE
 - Calculation-only approach not sufficient for dosimetry purposes
 - Experimental data must play a role
 - Covariance we use must be related to the cross section used – not appended to a different evaluation that had different development roles – ASTM E1018
 - Uncertainty in physics models – not just parametric variation – needs to be incorporated
 - » E.g. Issue with Watt fission spectrum in LEPRICON methodology



Sandia National Laboratories



Need for Expanded Covariance Matrices

- **Types of expanded considerations:**
 - ◆ Cross reaction
 - ◆ Cross material
- **Need:**
 - ◆ Not clear, we need some sensitivity studies here
- **Status:**
 - ◆ Expanded covariance for reactions and recoil spectrum are believed to be more important, at this time, for most dosimetry applications.
 - ◆ A simultaneous fitting of important dosimetry reactions is desired, i.e. GLUCS-like



Sandia National Laboratories



Application 2: Material Identification, Secondary Gamma Environments

- **Uncertainties for prompt gamma emission data**
 - Energies and yield
 - Reference data now available and used in latest ENDF/B-VI Rel. 8 cross sections.
 - » recent IAEA PGNA work and their Frankle-Reedy and Budapest data



Sandia National Laboratories



Path Forward


- Address covariance for ENDF/B-VII cross sections
 - ?? Include dosimetry sub-library
 - ?? Add in covariance matrices that should be usable by the dosimetry community even if they do not have extensive parameter variations of the supporting nuclear model calculations
- Continue with current cross section evaluation community emphasis on covariance matrices for all quantities



Sandia National Laboratories



Maybe this dialog will act as a map for both communities



Sandia National Laboratories



Are we too greedy in our request?

Dosimetry users are making significant requests from cross section community. Do we have a balanced perspective?



- We may be demanding, but we do have a balanced approach



Sandia National Laboratories



Spectrum Adjustment Refinements in Current Use

- Detailed a *priori* spectrum with structure
- Iterative unfolds that:
 - ◆ preserve this a *priori* structure – smooth adjustment, not trial spectrum
 - ◆ MC perturbation of trial spectrum to obtain uncertainty metric
- Cover treatment, beyond attenuation
- Covariance for a *priori* calculated spectrum



Sandia National Laboratories




Cover Treatment

- We have moved past a simple exponential attenuation model
 - Use energy-dependent adjoint response
 - Captures self-shielding when a thick cover is used
- Issue with cover-to-cover thickness variation from dosimetry suppliers
 - Now use specific thickness measurements
- Perturbation of free field spectrum by cover is addressed
 - No foil stacking in B_4C covers
 - Single B_4C covers per irradiation with monitor foil
 - Even monitor foil perturbation is an issue




Cover Correction Factors


Effect of covers on $^{197}\text{Au}(\text{n},\gamma)$




Self-shielding on $^{197}\text{Au}(\text{n},\gamma)$



Effect of B₄C+Cd for reactions



Self-shielding on reactions




Sandia National Laboratories


Covariance in *a priori* Spectrum

Use fission, $1/E^\alpha$, and Maxwellian components coupled with transitional regions as determined by fit to *a priori* spectrum

Fast Burst Reactor



ACRR PbB Bucket




Sandia National Laboratories



Simultaneous Spectrum Adjustment: Neutron Spectrum

Simultaneous spectrum adjustment:

- fast burst reactor cavity
- pool-type reactor cavity
- pool-type reactor in PbB bucket




Sandia National Laboratories



Simultaneous Spectrum Adjustment: Activities

Simultaneous spectrum adjustment:

- fast burst reactor cavity
- pool-type reactor cavity
- pool-type reactor in PbB bucket




Sandia National Laboratories




Covariance in *a priori* Spectrum

Example of constructed *a priori* spectrum covariance matrices

Fast Burst Reactor



ACRR PbB Bucket



Sandia National Laboratories



Dosimetry Priorities

- **High quality cross sections with covariance for reactions used**
- **Fission spectrum covariance to assist a *priori* spectrum covariance**
- **Cd covariance matrices**
- **Recoil spectrum covariance matrices for Si, Ga, As, Fe**
- **High energy dosimetry reaction covariance matrices**



Sandia National Laboratories



Questions???



Sandia National Laboratories

Example Double Ratios of Activation Foils

