

DT# 48132 QA:NA
Cb 813106

High-Precision Isotope Analysis of Uranium and Thorium by TIMS

Leonid A. Neymark and James B. Paces

U.S. Geological Survey, Denver, Colorado

The U.S. Geological Survey (USGS) Yucca Mountain Project Branch laboratory in Denver, Colorado, conducts routine high-precision isotope analyses of uranium (U) and thorium (Th) using thermal ionization mass-spectrometry (TIMS). The measurements are conducted by a solid-source mass-spectrometer equipped with a Faraday multi-collector system and an energy filter in front of an active-film-type secondary electron multiplier (SEM). The abundance sensitivity of the instrument (signal at mass 237 over ^{238}U in natural U) with the energy filter is $\sim 15 \times 10^{-9}$ and peak tails are reduced by a factor of ~ 100 relative to the Faraday cup measurements. Since instrument installation in April 2004, more than 500 rock and water samples have been analyzed in support of isotope-geochemical studies for the U.S. Department of Energy's Yucca Mountain Project.

Isotope ratios of sub-nanogram to microgram U and Th samples are measured on graphite-coated single-filament and double-filament assemblies using zone-refined rhenium filaments. Ion beams less than 5 millivolt (mV) are measured with the SEM, which is corrected for non-linearity on the basis of measurements of National Institute of Standards and Technology (NIST) U-500 and 4321B standards with ion beams ranging from 0.01 to 8 mV. Inter-calibration between the SEM and the Faraday multi-collector is performed for every mass cycle using a ~ 5 mV beam switched between Faraday cup and SEM ("bridging" technique), because SEM-Faraday inter-calibrations prior to the

measurement failed to produce acceptable results. Either natural (^{235}U) or artificial (^{236}U , ^{229}Th) isotopes were used for the bridging. Separate runs are conducted for minor isotopes using SEM only. These techniques result in high within-run precisions of <0.1 to 0.2 percent for $^{234}\text{U}/^{238}\text{U}$ and 0.2 to 0.5 percent for $^{230}\text{Th}/^{238}\text{U}$.

The performance of the instrument is monitored using several U and Th isotope standards. The mean measured $^{234}\text{U}/^{238}\text{U}$ in NIST SRM 4321B is $(52.879 \pm 0.004) \times 10^{-6}$ (95 percent confidence, $n=134$). The mean fractionation factor in these analyses is 0.09 ± 0.06 percent per mass unit. The $^{236}\text{U}/^{238}\text{U}$ value of $\sim 1.1 \times 10^{-9}$ measured in this standard defines an upper limit for the ^{236}U background in the analyses. The mean atomic $^{234}\text{U}/^{238}\text{U}$ and activity $^{230}\text{Th}/^{238}\text{U}$ for 29 runs of an internal “secular equilibrium” uranium ore standard is $(54.82 \pm 0.02) \times 10^{-6}$ and 1.002 ± 0.002 , respectively. The Th isotopic composition in the IRMM-036 standard was obtained after addition of high-purity ^{229}Th tracer and measuring $^{232}\text{Th}/^{229}\text{Th}$ and $^{230}\text{Th}/^{229}\text{Th}$ ratios in separate lower- and higher-temperature runs, respectively. The resulting $^{230}\text{Th}/^{232}\text{Th}$ for 30 runs is $31.09 \pm 0.04 \times 10^{-7}$, in excellent agreement with the certified value of $31.13 \pm 0.78 \times 10^{-7}$.

The USGS laboratory measured four samples of depleted to low-enriched uranium in the recent Regular European Inter-laboratory Measurement Evaluation Program (REIMEP-18). The results of these analyses are within 0.1 to 2 percent of the certified values.